

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091625\
 Data File : VX047592.D
 Acq On : 16 Sep 2025 13:35
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX091625

Quant Time: Sep 17 06:47:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	108	-0.01
2 T	Dichlorodifluoromethane	0.518	0.462	10.8	99	0.00
3 P	Chloromethane	0.680	0.595	12.5	97	0.00
4 C	Vinyl Chloride	0.666	0.610	8.4#	99	0.00
5 T	Bromomethane	0.422	0.410	2.8	104	0.00
6 T	Chloroethane	0.445	0.415	6.7	101	0.00
7 T	Trichlorofluoromethane	0.989	0.931	5.9	101	0.00
8 T	Diethyl Ether	0.405	0.378	6.7	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.578	0.561	2.9	103	0.00
10 T	Methyl Iodide	0.902	0.834	7.5	98	0.00
11 T	Tert butyl alcohol	0.089	0.091	-2.2	110	0.00
12 CM	1,1-Dichloroethene	0.594	0.559	5.9#	100	0.00
13 T	Acrolein	0.082	0.083	-1.2	110	0.00
14 T	Allyl chloride	1.226	1.146	6.5	102	0.00
15 T	Acrylonitrile	0.328	0.333	-1.5	103	0.00
16 T	Acetone	0.346	0.339	2.0	115	0.00
17 T	Carbon Disulfide	1.626	1.463	10.0	100	0.00
18 T	Methyl Acetate	0.770	0.850	-10.4	104	0.00
19 T	Methyl tert-butyl Ether	2.169	2.105	3.0	101	0.00
20 T	Methylene Chloride	0.732	0.674	7.9	99	0.00
21 T	trans-1,2-Dichloroethene	0.640	0.597	6.7	99	0.00
22 T	Diisopropyl ether	2.376	2.289	3.7	99	-0.01
23 T	Vinyl Acetate	1.901	1.867	1.8	101	0.00
24 P	1,1-Dichloroethane	1.263	1.213	4.0	101	0.00
25 T	2-Butanone	0.432	0.438	-1.4	107	0.00
26 T	2,2-Dichloropropane	1.050	1.039	1.0	105	0.00
27 T	cis-1,2-Dichloroethene	0.783	0.765	2.3	102	0.00
28 T	Bromochloromethane	0.551	0.514	6.7	96	-0.01
29 T	Tetrahydrofuran	0.260	0.263	-1.2	103	0.00
30 C	Chloroform	1.276	1.244	2.5#	101	0.00
31 T	Cyclohexane	1.052	0.968	8.0	101	0.00
32 T	1,1,1-Trichloroethane	1.082	1.060	2.0	101	0.00
33 S	1,2-Dichloroethane-d4	0.796	0.725	8.9	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	108	0.00
35 S	Dibromofluoromethane	0.335	0.306	8.7	100	0.00
36 T	1,1-Dichloropropene	0.490	0.457	6.7	101	0.00
37 T	Ethyl Acetate	0.525	0.517	1.5	105	-0.01
38 T	Carbon Tetrachloride	0.532	0.500	6.0	100	-0.01
39 T	Methylcyclohexane	0.556	0.541	2.7	103	0.00
40 TM	Benzene	1.488	1.410	5.2	100	0.00
41 T	Methacrylonitrile	0.271	0.274	-1.1	103	0.00
42 TM	1,2-Dichloroethane	0.566	0.531	6.2	98	0.00
43 T	Isopropyl Acetate	0.860	0.846	1.6	102	0.00
44 TM	Trichloroethene	0.360	0.346	3.9	101	0.00
45 C	1,2-Dichloropropane	0.381	0.366	3.9#	100	0.00
46 T	Dibromomethane	0.277	0.266	4.0	100	0.00
47 T	Bromodichloromethane	0.572	0.558	2.4	99	0.00
48 T	Methyl methacrylate	0.429	0.427	0.5	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091625\
 Data File : VX047592.D
 Acq On : 16 Sep 2025 13:35
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX091625

Quant Time: Sep 17 06:47:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.004	0.005	-25.0	106	0.00
50 S	Toluene-d8	1.160	1.077	7.2	101	0.00
51 T	4-Methyl-2-Pentanone	0.477	0.488	-2.3	103	0.00
52 CM	Toluene	0.917	0.883	3.7#	101	0.00
53 T	t-1,3-Dichloropropene	0.584	0.578	1.0	101	0.00
54 T	cis-1,3-Dichloropropene	0.624	0.615	1.4	102	0.00
55 T	1,1,2-Trichloroethane	0.354	0.338	4.5	99	0.00
56 T	Ethyl methacrylate	0.555	0.572	-3.1	103	0.00
57 T	1,3-Dichloropropane	0.607	0.594	2.1	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.227	0.238	-4.8	94	0.00
59 T	2-Hexanone	0.343	0.347	-1.2	104	0.00
60 T	Dibromochloromethane	0.407	0.406	0.2	101	0.00
61 T	1,2-Dibromoethane	0.360	0.355	1.4	101	0.00
62 S	4-Bromofluorobenzene	0.440	0.410	6.8	104	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
64 T	Tetrachloroethene	0.326	0.315	3.4	103	0.00
65 PM	Chlorobenzene	1.136	1.103	2.9	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.392	0.383	2.3	99	0.00
67 C	Ethyl Benzene	1.960	1.953	0.4#	102	0.00
68 T	m/p-Xylenes	0.729	0.722	1.0	101	0.00
69 T	o-Xylene	0.706	0.708	-0.3	101	0.00
70 T	Styrene	1.236	1.225	0.9	101	0.00
71 P	Bromoform	0.293	0.300	-2.4	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
73 T	Isopropylbenzene	3.801	3.762	1.0	101	0.00
74 T	N-amyl acetate	1.733	1.746	-0.8	102	0.00
75 P	1,1,2,2-Tetrachloroethane	1.150	1.147	0.3	102	0.00
76 T	1,2,3-Trichloropropane	1.012	1.076	-6.3	104	0.00
77 T	Bromobenzene	0.936	0.902	3.6	100	0.00
78 T	n-propylbenzene	4.494	4.430	1.4	101	0.00
79 T	2-Chlorotoluene	2.752	2.704	1.7	101	0.00
80 T	1,3,5-Trimethylbenzene	3.123	3.127	-0.1	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.397	0.401	-1.0	106	0.00
82 T	4-Chlorotoluene	3.251	3.196	1.7	101	0.00
83 T	tert-Butylbenzene	3.197	3.177	0.6	103	0.00
84 T	1,2,4-Trimethylbenzene	3.166	3.161	0.2	103	0.00
85 T	sec-Butylbenzene	3.797	3.805	-0.2	103	0.00
86 T	p-Isopropyltoluene	3.215	3.206	0.3	102	0.00
87 T	1,3-Dichlorobenzene	1.739	1.705	2.0	102	0.00
88 T	1,4-Dichlorobenzene	1.785	1.749	2.0	104	0.00
89 T	n-Butylbenzene	2.975	3.020	-1.5	105	0.00
90 T	Hexachloroethane	0.564	0.573	-1.6	105	0.00
91 T	1,2-Dichlorobenzene	1.657	1.634	1.4	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.233	0.237	-1.7	102	0.00
93 T	1,2,4-Trichlorobenzene	1.077	1.101	-2.2	104	0.00
94 T	Hexachlorobutadiene	0.366	0.381	-4.1	108	0.00
95 T	Naphthalene	3.279	3.334	-1.7	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091625\
Data File : VX047592.D
Acq On : 16 Sep 2025 13:35
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX091625

Quant Time: Sep 17 06:47:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.002	1.023	-2.1	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6