

Surrogate Summary

SDG No.: **Q3109**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3109-01	MW1D	1,2-Dichloroethane-d4	50	49.9	100		74	125
		Dibromofluoromethane	50	51.8	104		75	124
		Toluene-d8	50	47.1	94		86	113
		4-Bromofluorobenzene	50	47.5	95		77	121
Q3109-02	MW1	1,2-Dichloroethane-d4	50	45.3	91		74	125
		Dibromofluoromethane	50	48.2	96		75	124
		Toluene-d8	50	47.0	94		86	113
		4-Bromofluorobenzene	50	48.9	98		77	121
Q3109-02DL	MW1DL	1,2-Dichloroethane-d4	50	46.5	93		74	125
		Dibromofluoromethane	50	44.6	89		75	124
		Toluene-d8	50	46.6	93		86	113
		4-Bromofluorobenzene	50	51.5	103		77	121
VN0916WBL01	VN0916WBL01	1,2-Dichloroethane-d4	50	49.1	98		74	125
		Dibromofluoromethane	50	50.7	101		75	124
		Toluene-d8	50	47.1	94		86	113
		4-Bromofluorobenzene	50	45.7	91		77	121
VN0916WBS01	VN0916WBS01	1,2-Dichloroethane-d4	50	44.1	88		74	125
		Dibromofluoromethane	50	47.5	95		75	124
		Toluene-d8	50	45.0	90		86	113
		4-Bromofluorobenzene	50	46.8	94		77	121
VN0916WBSD0	VN0916WBSD01	1,2-Dichloroethane-d4	50	43.5	87		74	125
		Dibromofluoromethane	50	48.6	97		75	124
		Toluene-d8	50	45.6	91		86	113
		4-Bromofluorobenzene	50	46.4	93		77	121
VX0917WBL01	VX0917WBL01	1,2-Dichloroethane-d4	50	50.2	100		74	125
		Dibromofluoromethane	50	47.0	94		75	124
		Toluene-d8	50	47.4	95		86	113
		4-Bromofluorobenzene	50	52.9	106		77	121
VX0917WBS01	VX0917WBS01	1,2-Dichloroethane-d4	50	50.5	101		74	125
		Dibromofluoromethane	50	50.9	102		75	124
		Toluene-d8	50	51.5	103		86	113
		4-Bromofluorobenzene	50	51.4	103		77	121