

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3109
 Lab File ID: VN087847.D Date Analyzed: 09/16/2025
 Instrument ID: MSVOA_N Time Analyzed: 13:55
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	239136	8.21	433645	9.08	409197	11.85
UPPER LIMIT	478272	8.706	867290	9.583	818394	12.347
LOWER LIMIT	119568	7.706	216823	8.583	204599	11.347
EPA SAMPLE NO.						
MW1D	186859	8.21	417903	9.08	408304	11.85
MW1	207001	8.21	471522	9.08	448961	11.84
VN0916WBL01	184438	8.21	413572	9.08	397360	11.85
VN0916WBS01	212441	8.21	416535	9.08	384629	11.85
VN0916WBSD01	229272	8.21	433869	9.08	392898	11.85

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.