

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
Lab Code: ACE SDG NO.: Q3109
Lab File ID: VX047601.D Date Analyzed: 09/17/2025
Instrument ID: MSVOA_X Time Analyzed: 08:51
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	203321	5.53	347822	6.74	296232	10.04
UPPER LIMIT	406642	6.032	695644	7.239	592464	10.537
LOWER LIMIT	101661	5.032	173911	6.239	148116	9.537
EPA SAMPLE NO.						
MW1DL	157331	5.54	316909	6.75	305419	10.04
VX0917WBL01	235094	5.54	478397	6.75	460978	10.04
VX0917WBS01	195658	5.54	351376	6.74	311445	10.04

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.