

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3109</u>
Lab File ID:	<u>VN087847.D</u>	Date Analyzed:	<u>09/16/2025</u>
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>13:55</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	226483	13.77				
UPPER LIMIT	452966	14.27				
LOWER LIMIT	113242	13.27				
EPA SAMPLE NO.						
MW1D	185626	13.77				
MW1	231412	13.77				
VN0916WBL01	180365	13.77				
VN0916WBS01	210239	13.77				
VN0916WBSD01	209579	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.