

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091725\
 Data File : VL042934.D
 Acq On : 17 Sep 2025 13:07
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/18/2025
 Supervised By :Mahesh Dadoda 09/18/2025

Quant Time: Sep 18 02:31:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Thu Sep 18 02:25:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	151993	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.959	114	373216	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.882	117	318402	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.377	95	236647	9.509	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.100%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.583	62	305	0.037	ppbv	93
32) 1,1,1-Trichloroethane	3.376	97	970m	0.037	ppbv	
35) Carbon Tetrachloride	3.771	117	993m	0.037	ppbv	
38) Trichloroethene	4.551	130	563m	0.038	ppbv	
52) Tetrachloroethene	8.221	164	366m	0.036	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.963	83	927m	0.037	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL091725\
Data File : VL042934.D
Acq On : 17 Sep 2025 13:07
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Sep 18 02:31:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL091725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Thu Sep 18 02:25:28 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 09/18/2025
Supervised By : Mahesh Dadoda 09/18/2025

