

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042952.D
 Acq On : 22 Sep 2025 10:48
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Quant Time: Sep 23 02:10:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Tue Sep 23 02:07:33 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 09/23/2025
 Supervised By : Mahesh Dadoda 09/24/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	162797	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	511767	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	469265	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.370	95	348878	9.960	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.600%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.499	85	23644	1.080	ppbv	97
3) Chlorodifluoromethane	1.476	51	29646	1.086	ppbv	99
4) Chloromethane	1.531	50	10120	1.021	ppbv	97
5) Vinyl Chloride	1.580	62	9228	1.030	ppbv	99
6) Bromomethane	1.667	94	3493	1.105	ppbv	96
7) Chloroethane	1.703	64	3872	1.113	ppbv #	81
8) Dichlorotetrafluoroethane	1.554	85	20172	1.079	ppbv	97
9) Propene	1.482	41	12808	1.079	ppbv	92
10) Heptane	4.972	43	33540m	1.053	ppbv	
11) Trichlorofluoromethane	1.874	101	22397	1.061	ppbv	98
12) 1,1,2-Trichlorotrifluo...	2.136	101	18180	1.063	ppbv	99
13) Ethanol	1.719	45	1463	1.286	ppbv #	33
14) Bromoethene	1.780	108	6783	1.049	ppbv #	87
15) Acetone	1.832	43	25273	1.192	ppbv	99
16) 1,3-Butadiene	1.609	39	10993	1.099	ppbv	88
17) tert-Butyl alcohol	2.039	59	25141	1.081	ppbv #	88
18) 1,1-Dichloroethene	2.033	96	8192	1.035	ppbv	96
19) Isopropyl Alcohol	1.884	45	13241	1.039	ppbv #	41
20) Methylene Chloride	2.059	84	10045	0.985	ppbv	97
21) Allyl Chloride	2.094	41	15801	1.070	ppbv	95
22) trans-1,2-Dichloroethene	2.337	96	9960	1.108	ppbv	96
23) Vinyl Acetate	2.457	43	30884	1.020	ppbv #	96
24) 1,1-Dichloroethane	2.405	63	19077	1.082	ppbv	99
25) Ethyl Acetate	2.832	43	56877	1.061	ppbv	97
26) Hexane	2.822	57	26276	1.076	ppbv	96
27) Carbon Disulfide	2.149	76	23373	1.067	ppbv #	32
28) Methyl tert-Butyl Ether	2.441	73	12580	0.991	ppbv #	86
29) Chloroform	2.842	83	33240	1.041	ppbv	99
30) Cyclohexane	3.855	84	21707	1.063	ppbv	98
31) cis-1,2-Dichloroethene	2.716	61	23828	1.041	ppbv	96
32) 1,1,1-Trichloroethane	3.360	97	35233	1.018	ppbv	97
34) 2-Butanone	2.554	43	37431	1.044	ppbv	97
35) Carbon Tetrachloride	3.758	117	34016	1.033	ppbv	96
36) Benzene	3.651	78	49652	1.033	ppbv	100
37) 1,2-Dichloroethane	3.214	62	26482	1.048	ppbv	98
38) Trichloroethene	4.538	130	23906	1.001	ppbv	98
39) 1,2-Dichloropropane	4.311	63	18493m	1.053	ppbv	
40) 1,4-Dioxane	4.583	88	8804m	1.080	ppbv	
41) Tetrahydrofuran	3.055	42	20845	1.011	ppbv	98
42) Bromodichloromethane	4.480	83	35768m	1.016	ppbv	
43) Methyl Methacrylate	4.871	69	20289	1.040	ppbv	98
44) 2,2,4-Trimethylpentane	4.635	57	84970	1.023	ppbv	96
45) t-1,3-Dichloropropene	6.402	75	21210m	1.011	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092225\
 Data File : VL042952.D
 Acq On : 22 Sep 2025 10:48
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/23/2025
 Supervised By :Mahesh Dadoda 09/24/2025

Quant Time: Sep 23 02:10:27 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL092225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Sep 23 02:07:33 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.587	75	27025m	1.011	ppbv	
47) 1,1,2-Trichloroethane	6.574	97	19311	1.041	ppbv	96
48) Dibromochloromethane	7.383	129	30015	1.012	ppbv	99
49) Bromoform	9.477	173	25929	1.005	ppbv	98
50) 4-Methyl-2-Pentanone	5.748	43	48546	1.007	ppbv	98
51) 2-Hexanone	7.431	43	39627	1.034	ppbv #	99
52) Tetrachloroethene	8.218	164	18188	0.978	ppbv	97
53) Toluene	6.926	91	58200	1.023	ppbv	100
54) 1,2-Dibromoethane	7.645	107	26608	1.020	ppbv	94
56) 1,1,1,2-Tetrachloroethane	8.920	131	22802	1.015	ppbv	94
57) Chlorobenzene	8.914	112	45873	1.071	ppbv	98
58) Ethyl Benzene	9.351	91	80473	1.044	ppbv	97
59) m/p-Xylene	9.529	91	128139m	2.108	ppbv	
60) o-Xylene	9.959	91	64703	1.066	ppbv	99
61) Styrene	9.869	104	29061	1.012	ppbv	98
62) Isopropylbenzene	10.535	105	96621	1.068	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.959	83	31695	0.993	ppbv	96
64) n-propylbenzene	10.985	120	24820	1.049	ppbv	98
65) tert-Butylbenzene	11.526	119	86303	1.072	ppbv	99
66) Benzyl Chloride	11.610	91	8140	0.964	ppbv	100
67) sec-Butylbenzene	11.762	105	120070	1.070	ppbv	100
69) p-Isopropyltoluene	11.921	119	101749	1.074	ppbv	99
70) n-Butylbenzene	12.277	91	98908	1.069	ppbv	99
71) 2-Chlorotoluene	10.895	91	70577	1.033	ppbv	98
72) 4-Ethyltoluene	11.121	105	79095	1.040	ppbv	99
73) 1,3,5-Trimethylbenzene	11.199	105	63174	1.024	ppbv	98
74) 1,2,4-Trimethylbenzene	11.532	105	71411	1.049	ppbv	95
75) 1,3-Dichlorobenzene	11.603	146	44769	1.060	ppbv	99
76) 1,4-Dichlorobenzene	11.668	146	42879	1.022	ppbv	97
77) 1,2-Dichlorobenzene	11.940	146	42911	1.068	ppbv	98
78) Hexachloro-1,3-Butadiene	13.879	225	34798	1.107	ppbv	99
79) Naphthalene	13.510	128	61925	1.028	ppbv	98
80) Naphthalene,2-methyl-	14.455	142	17116	0.920	ppbv	97
81) 1,2,4-Trichlorobenzene	13.435	180	32411	1.007	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

