

VOLATILE CONTINUING CALIBRATION CHECK

|                     |                   |                        |                              |
|---------------------|-------------------|------------------------|------------------------------|
| Lab Name:           | <u>Alliance</u>   | Contract:              | <u>TRIS02</u>                |
| Lab Code:           | <u>ACE</u>        | SDG No.:               | <u>Q3120</u>                 |
| Instrument ID:      | <u>MSVOA_N</u>    | Calibration Date/Time: | <u>09/17/2025 09:42</u>      |
| Lab File ID:        | <u>VN087869.D</u> | Init. Calib. Date(s):  | <u>09/04/2025 09/04/2025</u> |
| Heated Purge: (Y/N) | <u>N</u>          | Init. Calib. Time(s):  | <u>16:23 17:46</u>           |
| GC Column:          | <u>RXI-624</u>    | ID:                    | <u>0.25</u> (mm)             |

| COMPOUND              | RRF   | RRF020 | MIN<br>RRF | %D     | MAX%D |
|-----------------------|-------|--------|------------|--------|-------|
| Ethyl Acetate         | 0.686 | 0.591  |            | -13.85 |       |
| Isopropyl Acetate     | 1.076 | 0.922  |            | -14.31 |       |
| Acetone               | 0.375 | 0.315  |            | -16    |       |
| Methylene Chloride    | 2.052 | 1.906  |            | -7.11  |       |
| 1,2-Dichloroethane-d4 | 2.563 | 2.512  | 0.01       | -1.99  |       |
| Toluene-d8            | 1.395 | 1.343  | 0.01       | -3.73  |       |
| 4-Bromofluorobenzene  | 0.482 | 0.479  | 0.2        | -0.62  |       |
| n-amyl Acetate        | 0.727 | 0.552  |            | -24.07 |       |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.