

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3120</u>	Analytical Method:	<u>SW624.1</u>
Client:	<u>Tris Pharma, Inc.</u>	Datafile :	<u>VN087871.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0917WBSD01	Ethyl Acetate	20	19.5	ug/L	98	11		68	140	20
	Isopropyl Acetate	20	19.1	ug/L	96	6		70	130	20
	Acetone	100	89.7	ug/L	90	2		41	148	20
	Methylene Chloride	20	20.5	ug/L	103	4		60	140	20
	n-amyl Acetate	20	14.0	ug/L	70	1		70	130	20