

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TRIS02
 Lab Code: ACE SDG NO.: Q3120
 Lab File ID: VN087869.D Date Analyzed: 09/17/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:42
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	52147	7.79	281875	9.08	258937	11.84
UPPER LIMIT	104294	8.294	563750	9.583	517874	12.341
LOWER LIMIT	26073.5	7.294	140938	8.583	129469	11.341
EPA SAMPLE NO.						
OUTFALL-DSN-001	49388	7.80	249408	9.08	226027	11.84
OUTFALL-DSN-002	48910	7.79	249908	9.08	220815	11.85
VN0917WBL01	50672	7.80	260263	9.08	234698	11.85
VN0917WBS01	53476	7.79	296032	9.08	268776	11.85
VN0917WBSD01	52165	7.79	278561	9.08	250388	11.85

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.