

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: TRIS02

Lab Code: ACE

SDG NO.: Q3120

Lab File ID: VN087713.D

BFB Injection Date: 09/04/2025

Instrument ID: MSVOA_N

BFB Injection Time: 10:10

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.1
75	30.0 - 60.0% of mass 95	57.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.9 (1.3) 1
174	50.0 - 100.0% of mass 95	67
175	5.0 - 9.0% of mass 174	5.4 (8.1) 1
176	95.0 - 101.0% of mass 174	65.8 (98.3) 1
177	5.0 - 9.0% of mass 176	4.7 (7.1) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VN087725.D	09/04/2025	16:23
VSTDIC020	VSTDIC020	VN087726.D	09/04/2025	16:44
VSTDIC050	VSTDIC050	VN087727.D	09/04/2025	17:05
VSTDIC100	VSTDIC100	VN087728.D	09/04/2025	17:25
VSTDIC150	VSTDIC150	VN087729.D	09/04/2025	17:46