

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>PSEG04</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q3122</u>
Instrument ID: <u>MSVOA_N</u>	Calibration Date(s): <u>08/21/2025</u> <u>08/21/2025</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>12:09</u> <u>14:44</u>
GC Column: <u>RXI-624</u> ID: <u>0.25</u> (mm)	

LAB FILE ID:		RRF001 = VN087619.D	RRF005 = VN087620.D	RRF020 = VN087621.D				
		RRF050 = VN087622.D	RRF100 = VN087623.D	RRF150 = VN087624.D				
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Benzene	1.853	1.637	1.727	1.684	1.639	1.652	1.699	4.9
Toluene	1.148	0.987	1.062	1.050	1.037	1.033	1.053	5
Ethyl Benzene	2.232	1.993	2.224	2.144	2.152	2.178	2.154	4
m/p-Xylenes	0.759	0.735	0.811	0.812	0.808	0.814	0.790	4.3
o-Xylene	0.777	0.707	0.794	0.796	0.788	0.784	0.774	4.3
1,2-Dichloroethane-d4		1.025	1.016	0.959	1.035	1.089	1.024	4.5
Dibromofluoromethane		0.369	0.344	0.334	0.373	0.386	0.361	6
Toluene-d8		1.384	1.266	1.223	1.408	1.475	1.351	7.7
4-Bromofluorobenzene		0.446	0.454	0.450	0.514	0.540	0.481	9

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.