

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>PSEG04</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3122</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>09/18/2025 10:25</u>
Lab File ID:	<u>VN087885.D</u>	Init. Calib. Date(s):	<u>08/21/2025 08/21/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:09 14:44</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Benzene	1.699	1.885		10.95	20
Toluene	1.053	1.207		14.63	20
Ethyl Benzene	2.154	2.520		16.99	20
m/p-Xylenes	0.790	0.964		22.02	20
o-Xylene	0.774	0.917		18.48	20
1,2-Dichloroethane-d4	1.024	0.836		-18.36	20
Dibromofluoromethane	0.361	0.373		3.32	20
Toluene-d8	1.351	1.305		-3.4	20
4-Bromofluorobenzene	0.481	0.477		-0.83	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.