

VOLATILE CONTINUING CALIBRATION CHECK

|                     |                   |                        |                              |
|---------------------|-------------------|------------------------|------------------------------|
| Lab Name:           | <u>Alliance</u>   | Contract:              | <u>PSEG04</u>                |
| Lab Code:           | <u>ACE</u>        | SDG No.:               | <u>Q3122</u>                 |
| Instrument ID:      | <u>MSVOA_N</u>    | Calibration Date/Time: | <u>09/19/2025 10:10</u>      |
| Lab File ID:        | <u>VN087899.D</u> | Init. Calib. Date(s):  | <u>08/21/2025 08/21/2025</u> |
| Heated Purge: (Y/N) | <u>N</u>          | Init. Calib. Time(s):  | <u>12:09 14:44</u>           |
| GC Column:          | <u>RXI-624</u>    | ID:                    | <u>0.25</u> (mm)             |

| COMPOUND              | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|-----------------------|-------|--------|------------|--------|-------|
| Benzene               | 1.699 | 1.806  |            | 6.3    | 20    |
| Toluene               | 1.053 | 1.147  |            | 8.93   | 20    |
| Ethyl Benzene         | 2.154 | 2.365  |            | 9.8    | 20    |
| m/p-Xylenes           | 0.790 | 0.919  |            | 16.33  | 20    |
| o-Xylene              | 0.774 | 0.878  |            | 13.44  | 20    |
| 1,2-Dichloroethane-d4 | 1.024 | 0.875  |            | -14.55 | 20    |
| Dibromofluoromethane  | 0.361 | 0.390  |            | 8.03   | 20    |
| Toluene-d8            | 1.351 | 1.356  |            | 0.37   | 20    |
| 4-Bromofluorobenzene  | 0.481 | 0.503  |            | 4.57   | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.