

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN082125

Review By	John Carlone	Review On	8/22/2025 9:17:52 AM
Supervise By	Mahesh Dadoda	Supervise On	8/25/2025 8:11:38 AM
SubDirectory	VN082125	HP Acquire Method	HP Processing Method 82N082125W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP135214		
Initial Calibration Stds	VP135215,VP135216,VP135217,VP135218,VP135219,VP135220		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP135221		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN087614.D	21 Aug 2025 09:30	JC\MD	Ok
2	VSTDCCC020	VN087615.D	21 Aug 2025 10:04	JC\MD	Not Ok
3	VN0821WBS01	VN087616.D	21 Aug 2025 10:37	JC\MD	Not Ok
4	VN0821WBSD01	VN087617.D	21 Aug 2025 11:11	JC\MD	Not Ok
5	VN0821WBL01	VN087618.D	21 Aug 2025 11:36	JC\MD	Not Ok
6	VSTDICC001	VN087619.D	21 Aug 2025 12:09	JC\MD	Ok,M
7	VSTDICC005	VN087620.D	21 Aug 2025 13:08	JC\MD	Ok,M
8	VSTDICC020	VN087621.D	21 Aug 2025 13:41	JC\MD	Ok,M
9	VSTDICCC050	VN087622.D	21 Aug 2025 14:02	JC\MD	Ok,M
10	VSTDICC100	VN087623.D	21 Aug 2025 14:23	JC\MD	Ok,M
11	VSTDICC150	VN087624.D	21 Aug 2025 14:44	JC\MD	Ok,M
12	IBLK	VN087625.D	21 Aug 2025 15:05	JC\MD	Ok
13	VSTDICV050	VN087626.D	21 Aug 2025 15:47	JC\MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN091825

Review By	John Carlone	Review On	9/19/2025 8:14:56 AM
Supervise By		Supervise On	
SubDirectory	VN091825	HP Acquire Method	HP Processing Method 82N082125W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135534		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135535,VP135536		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN087884.D	18 Sep 2025 08:48	JC\MD	Ok
2	VSTDCCC050	VN087885.D	18 Sep 2025 10:25	JC\MD	Ok,M
3	VN0918MBL01	VN087886.D	18 Sep 2025 10:58	JC\MD	Ok
4	VN0918WBL01	VN087887.D	18 Sep 2025 11:19	JC\MD	Ok
5	VN0918WBS01	VN087888.D	18 Sep 2025 11:40	JC\MD	Ok,M
6	VN0918WBSD01	VN087889.D	18 Sep 2025 12:24	JC\MD	Ok,M
7	PB169730TB	VN087890.D	18 Sep 2025 12:45	JC\MD	Ok
8	Q3117-02	VN087891.D	18 Sep 2025 13:06	JC\MD	Ok
9	VN0918MBS01	VN087892.D	18 Sep 2025 13:28	JC\MD	Ok,M
10	VN0918MBSD01	VN087893.D	18 Sep 2025 14:06	JC\MD	Ok,M
11	Q3122-02	VN087894.D	18 Sep 2025 14:48	JC\MD	Dilution
12	IBLK	VN087895.D	18 Sep 2025 15:09	JC\MD	Ok
13	IBLK	VN087896.D	18 Sep 2025 15:30	JC\MD	Ok
14	VSTDCCC050	VN087897.D	18 Sep 2025 15:53	JC\MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN091925

Review By	Semsettin Yesilyurt	Review On	9/22/2025 7:02:17 AM
Supervise By	Maresh Dadoda	Supervise On	9/22/2025 7:03:58 AM
SubDirectory	VN091925	HP Acquire Method	HP Processing Method 82N082125W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135546		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135547,VP135548		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN087898.D	19 Sep 2025 09:36	JC\MD	Ok
2	VSTDCCC050	VN087899.D	19 Sep 2025 10:10	JC\MD	Ok,M
3	VN0919MBL01	VN087900.D	19 Sep 2025 10:44	JC\MD	Ok
4	VN0919WBL01	VN087901.D	19 Sep 2025 11:05	JC\MD	Ok
5	VN0919MBS01	VN087902.D	19 Sep 2025 11:26	JC\MD	Ok,M
6	VN0919MBSD01	VN087903.D	19 Sep 2025 11:59	JC\MD	Ok,M
7	Q3122-02DL	VN087904.D	19 Sep 2025 12:20	JC\MD	Not Ok
8	IBLK	VN087905.D	19 Sep 2025 12:41	JC\MD	Ok
9	IBLK	VN087906.D	19 Sep 2025 13:02	JC\MD	Ok
10	Q3122-02DL	VN087907.D	19 Sep 2025 13:24	JC\MD	Ok
11	IBLK	VN087908.D	19 Sep 2025 13:45	JC\MD	Ok
12	IBLK	VN087909.D	19 Sep 2025 14:06	JC\MD	Ok
13	VSTDCCC050	VN087910.D	19 Sep 2025 14:54	JC\MD	Ok,M

M : Manual Integration