

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3122</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Parkway Generation Operating LLC</u>	Datafile :	<u>VN087892.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VN0918MBS01	Benzene	2000	1900	ug/Kg	95			84	121	
	Toluene	2000	1900	ug/Kg	95			83	122	
	Ethyl Benzene	2000	1800	ug/Kg	90			82	124	
	m/p-Xylenes	4000	3800	ug/Kg	95			83	124	
	o-Xylene	2000	1800	ug/Kg	90			83	123	