

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: PSEG04
Lab Code: ACE SDG NO.: Q3122
Lab File ID: VN087899.D Date Analyzed: 09/19/2025
Instrument ID: MSVOA_N Time Analyzed: 10:10
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	248323	8.21	438441	9.08	407017	11.84
UPPER LIMIT	496646	8.706	876882	9.583	814034	12.341
LOWER LIMIT	124162	7.706	219221	8.583	203509	11.341
EPA SAMPLE NO.						
FOTP915DL	163737	8.21	353909	9.08	344231	11.85
VN0919MBL01	177027	8.21	384630	9.08	371538	11.85
VN0919MBS01	227431	8.21	427601	9.08	403690	11.85

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.