

Continuing Calibration Report for SequenceID : FD092225AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049737.D

Aromatic C12-C16	7916565.000	60.000	6.132	8.759	131942.750	135865.661	2.887
Aromatic C16-C21	10200765.000	80.000	8.760	13.050	127509.563	135770.533	6.085
Aromatic C10-C12	4502846.000	40.000	4.389	6.131	112571.150	116056.185	3.003
Aromatic C21-C36	19572572.000	180.000	13.051	18.469	108736.511	118321.901	8.101
Aromatic EPH	42192748.000	360.000	4.389	18.469	117202.078	124871.589	6.142