

Continuing Calibration Report for SequenceID : FD092225AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049746.D

Aromatic C12-C16	8102983.000	60.000	6.132	8.759	135049.717	135865.661	0.601
Aromatic C16-C21	10726454.000	80.000	8.760	13.050	134080.675	135770.533	1.245
Aromatic C10-C12	4630056.000	40.000	4.389	6.131	115751.400	116056.185	0.263
Aromatic C21-C36	20745012.000	180.000	13.051	18.469	115250.067	118321.901	2.596
Aromatic EPH	44204505.000	360.000	4.389	18.469	122790.292	124871.589	1.667