

WATER DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE

Lab Name: Alliance

Client: ATG-GREENVILLE ARR

Lab Code: ACE

SDG No: Q3124

Client Sample ID : PB169733BSD

Datafile: FG016618.D

COMPOUND	SPIKE ADDED ug/L	CONCENTRATION ug/L	LCS/LCSD CONCENTRATION ug/L	% REC	QC LIMITS (%)
DRO	200	0	157	78	78-117

LCS/LCSD % Recovery RPD : 16.4