

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3125

Project: Monthly 2025

Run No.: LB137232

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137232BS							
BOD5	mg/L	198	174		88	1	84.6-115.4	09/17/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3125

Project: Monthly 2025

Run No.: LB137250

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137250BS							
TSS	mg/L	550	532		97	1	90-110	09/19/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3125

Project: Monthly 2025

Run No.: LB137253

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137253BS							
TPH	mg/L	20.0	16.6		83	1	78-114	09/22/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3125

Project: Monthly 2025

Run No.: LB137253

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137253BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	09/22/2025