

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137278

Review By	mohan	Review On	9/24/2025 10:25:09 AM
Supervise By	jaswal	Supervise On	9/24/2025 10:55:19 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87304,MP87305,MP87306,MP87307,MP87308,MP87309,MP87310		
ICV Standard	MP87311		
CCV Standard	MP87314		
ICSA Standard			
CRI Standard	MP87315		
LCS Standard			
Chk Standard	MP87312,MP87313,MP87316,MP87301		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	09/23/25 13:51		mohan	OK
2	S0.05	S0.05	CAL2	09/23/25 13:59		mohan	OK
3	S0.2	S0.2	CAL3	09/23/25 14:01		mohan	OK
4	S2.5	S2.5	CAL4	09/23/25 14:04		mohan	OK
5	S5	S5	CAL5	09/23/25 14:11		mohan	OK
6	S7.5	S7.5	CAL6	09/23/25 14:20		mohan	OK
7	S10	S10	CAL7	09/23/25 14:39		mohan	OK
8	ICV42	ICV42	ICV	09/23/25 14:45		mohan	OK
9	ICB42	ICB42	ICB	09/23/25 14:47		mohan	OK
10	CCV28	CCV28	CCV	09/23/25 14:50		mohan	OK
11	CCB28	CCB28	CCB	09/23/25 14:52		mohan	OK
12	CRA	CRA	CRDL	09/23/25 14:54		mohan	OK
13	HighStd	HighStd	HIGH STD	09/23/25 14:56		mohan	OK
14	ChkStd	ChkStd	SAM	09/23/25 14:59		mohan	OK
15	PB169802BL	PB169802BL	MB	09/23/25 15:01		mohan	OK
16	PB169802BS	PB169802BS	LCS	09/23/25 15:03		mohan	OK
17	Q3125-02	Comp	SAM	09/23/25 15:06		mohan	OK
18	Q3125-02DUP	CompDUP	DUP	09/23/25 15:08		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137278

Review By	mohan	Review On	9/24/2025 10:25:09 AM
Supervise By	jaswal	Supervise On	9/24/2025 10:55:19 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87304,MP87305,MP87306,MP87307,MP87308,MP87309,MP87310		
ICV Standard	MP87311		
CCV Standard	MP87314		
ICSA Standard			
CRI Standard	MP87315		
LCS Standard			
Chk Standard	MP87312,MP87313,MP87316,MP87301		

19	Q3125-02MS	CompMS	MS	09/23/25 15:10		mohan	OK
20	Q3125-02MSD	CompMSD	MSD	09/23/25 15:16		mohan	OK
21	Q3158-01	EFFLUENT-DAY-1-M	SAM	09/23/25 15:18		mohan	OK
22	CCV29	CCV29	CCV	09/23/25 15:20		mohan	OK
23	CCB29	CCB29	CCB	09/23/25 15:23		mohan	OK
24	Q3125-02L	CompL	SD	09/23/25 15:25		mohan	OK
25	Q3125-02A	CompA	PS	09/23/25 15:27		mohan	OK
26	CCV30	CCV30	CCV	09/23/25 15:29		mohan	OK
27	CCB30	CCB30	CCB	09/23/25 15:32		mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137314

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	09/25/25 10:55		Jaswal	OK
2	S1	S1	CAL2	09/25/25 11:00		Jaswal	OK
3	S2	S2	CAL3	09/25/25 11:04		Jaswal	OK
4	S3	S3	CAL4	09/25/25 11:08		Jaswal	OK
5	S4	S4	CAL5	09/25/25 11:12		Jaswal	OK
6	S5	S5	CAL6	09/25/25 11:16		Jaswal	OK
7	ICV01	ICV01	ICV	09/25/25 11:20		Jaswal	OK
8	LLICV01	LLICV01	LLICV	09/25/25 11:29		Jaswal	OK
9	ICB01	ICB01	ICB	09/25/25 11:33		Jaswal	OK
10	CRI01	CRI01	CRDL	09/25/25 11:38		Jaswal	OK
11	ICSA01	ICSA01	ICSA	09/25/25 11:42		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	09/25/25 11:46		Jaswal	OK
13	ICSADL	ICSADL	ICSA	09/25/25 11:50		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	09/25/25 12:00		Jaswal	OK
15	CCV01	CCV01	CCV	09/25/25 12:12		Jaswal	OK
16	CCB01	CCB01	CCB	09/25/25 12:17		Jaswal	OK
17	LR-2	LR-2	HIGH STD	09/25/25 12:40		Jaswal	OK
18	LR-3	LR-3	HIGH STD	09/25/25 12:44		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137314

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

19	PB169800BL	PB169800BL	MB	09/25/25 12:54		Jaswal	OK
20	PB169800BS	PB169800BS	LCS	09/25/25 12:58		Jaswal	OK
21	PB169812BL	PB169812BL	MB	09/25/25 13:02		Jaswal	OK
22	PB169812BS	PB169812BS	LCS	09/25/25 13:07		Jaswal	OK
23	LR-1	LR-1	HIGH STD	09/25/25 13:11		Jaswal	OK
24	Q3187-01	228	SAM	09/25/25 13:15		Jaswal	OK
25	Q3188-02	RB22120	SAM	09/25/25 13:19		Jaswal	OK
26	CCV02	CCV02	CCV	09/25/25 13:23		Jaswal	OK
27	CCB02	CCB02	CCB	09/25/25 13:27		Jaswal	OK
28	PB169833BL	PB169833BL	MB	09/25/25 13:32		Jaswal	OK
29	PB169833BS	PB169833BS	LCS	09/25/25 13:36		Jaswal	OK
30	Q3125-02	Comp	SAM	09/25/25 13:40		Jaswal	OK
31	Q3125-02DUP	CompDUP	DUP	09/25/25 13:44		Jaswal	OK
32	Q3125-02L	CompL	SD	09/25/25 13:49		Jaswal	OK
33	Q3125-02MS	CompMS	MS	09/25/25 13:53		Jaswal	OK
34	Q3125-02MSD	CompMSD	MSD	09/25/25 13:57		Jaswal	OK
35	Q3125-02A	CompA	PS	09/25/25 14:08		Jaswal	OK
36	Q3168-06	TOTES COMP#2	SAM	09/25/25 14:13	Confirm wt/vol	Jaswal	OK
37	Q3172-01	FW7D	SAM	09/25/25 14:17		Jaswal	OK
38	CCV03	CCV03	CCV	09/25/25 14:21		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137314

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

39	CCB03	CCB03	CCB	09/25/25 14:30		Jaswal	OK
40	Q3172-02	FW6D	SAM	09/25/25 14:35		Jaswal	OK
41	Q3168-05	TOTES COMP#1	SAM	09/25/25 14:39		Jaswal	OK
42	Q3168-05DUP	TOTES COMP#1DUP	DUP	09/25/25 14:43		Jaswal	OK
43	Q3168-05L	TOTES COMP#1L	SD	09/25/25 14:47		Jaswal	OK
44	Q3168-05MS	TOTES COMP#1MS	MS	09/25/25 14:52		Jaswal	OK
45	Q3168-05MSD	TOTES COMP#1MSD	MSD	09/25/25 14:56		Jaswal	OK
46	Q3168-05A	TOTES COMP#1A	PS	09/25/25 15:00		Jaswal	OK
47	Q3189-01	72-11989	SAM	09/25/25 15:04		Jaswal	OK
48	Q3176-01	EME-TS12-01	SAM	09/25/25 15:08		Jaswal	OK
49	Q3176-02	EME-TS13-01	SAM	09/25/25 15:12		Jaswal	OK
50	CCV04	CCV04	CCV	09/25/25 15:16		Jaswal	OK
51	CCB04	CCB04	CCB	09/25/25 15:20		Jaswal	OK
52	Q3176-03	EME-TS14-02	SAM	09/25/25 15:24		Jaswal	OK
53	Q3178-01	EME-TS14-01	SAM	09/25/25 15:28		Jaswal	OK
54	Q3178-02	MS	MS	09/25/25 15:40		Jaswal	OK
55	Q3178-03	MSD	MSD	09/25/25 15:44		Jaswal	OK
56	Q3178-01A	EME-TS14-01A	PS	09/25/25 15:48		Jaswal	OK
57	Q3183-01	WC-1	SAM	09/25/25 15:52		Jaswal	OK
58	Q3183-04	WC-4	SAM	09/25/25 15:56		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137314

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

59	Q3183-07	WC-3	SAM	09/25/25 16:00		Jaswal	OK
60	Q3178-01DUP	EME-TS14-01DUP	DUP	09/25/25 16:07		Jaswal	OK
61	Q3178-01L	EME-TS14-01L	SD	09/25/25 16:11		Jaswal	OK
62	CCV05	CCV05	CCV	09/25/25 16:16		Jaswal	OK
63	CCB05	CCB05	CCB	09/25/25 16:20		Jaswal	OK
64	PB169832BL	PB169832BL	MB	09/25/25 16:29		Jaswal	OK
65	PB169832BS	PB169832BS	LCS	09/25/25 16:43		Jaswal	OK
66	PB169754TB	PB169754TB	MB	09/25/25 16:47		Jaswal	OK
67	PB169816TB	PB169816TB	MB	09/25/25 16:52		Jaswal	OK
68	Q3168-04	ARS20-0010-Oily Soil	SAM	09/25/25 16:56		Jaswal	OK
69	Q3168-04DUP	ARS20-0010-Oily Soil	DUP	09/25/25 17:01		Jaswal	OK
70	Q3168-04L	ARS20-0010-Oily Soil	SD	09/25/25 17:05		Jaswal	OK
71	Q3168-04MS	ARS20-0010-Oily Soil	MS	09/25/25 17:09		Jaswal	OK
72	Q3168-04MSD	ARS20-0010-Oily Soil	MSD	09/25/25 17:13		Jaswal	OK
73	Q3168-04A	ARS20-0010-Oily Soil	PS	09/25/25 17:23		Jaswal	OK
74	CCV06	CCV06	CCV	09/25/25 17:27		Jaswal	OK
75	CCB06	CCB06	CCB	09/25/25 17:31		Jaswal	OK
76	Q3170-02DL	MOD-25-0271DL	SAM	09/25/25 17:35	Confirm wt/vol & Straight 5x Dilution for all elements	Jaswal	OK
77	CCV07	CCV07	CCV	09/25/25 17:40		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137314

Review By		Review On					
Supervise By		Supervise On					
STD. NAME		STD REF.#					
ICAL Standard							
ICV Standard							
CCV Standard							
ICSA Standard							
CRI Standard							
LCS Standard							
Chk Standard							
78	CCB07	CCB07	CCB	09/25/25 17:44		Jaswal	OK