

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3125

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV42	Mercury	4.08	4.0	102	95 - 105	CV	09/23/2025	14:45	LB137278

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3125

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV28	Mercury	5.14	5.0	103	90 - 110	CV	09/23/2025	14:50	LB137278
CCV29	Mercury	5.11	5.0	102	90 - 110	CV	09/23/2025	15:20	LB137278
CCV30	Mercury	5.18	5.0	104	90 - 110	CV	09/23/2025	15:29	LB137278

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3125

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	1970	2000	98	95 - 105	P	09/25/2025	11:20	LB137314
	Copper	1010	1000	101	95 - 105	P	09/25/2025	11:20	LB137314
	Lead	3850	4000	96	95 - 105	P	09/25/2025	11:20	LB137314
	Nickel	1970	2000	99	95 - 105	P	09/25/2025	11:20	LB137314
	Zinc	2030	2000	101	95 - 105	P	09/25/2025	11:20	LB137314

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3125

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	5.96	6.0	99	80 - 120	P	09/25/2025	11:29	LB137314
	Copper	20.5	20.0	103	80 - 120	P	09/25/2025	11:29	LB137314
	Lead	10.2	12.0	85	80 - 120	P	09/25/2025	11:29	LB137314
	Nickel	40.1	40.0	100	80 - 120	P	09/25/2025	11:29	LB137314
	Zinc	42.4	40.0	106	80 - 120	P	09/25/2025	11:29	LB137314

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3125

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2420	2500	97	90 - 110	P	09/25/2025	12:12	LB137314
	Copper	1230	1250	98	90 - 110	P	09/25/2025	12:12	LB137314
	Lead	4800	5000	96	90 - 110	P	09/25/2025	12:12	LB137314
	Nickel	2410	2500	96	90 - 110	P	09/25/2025	12:12	LB137314
	Zinc	2490	2500	100	90 - 110	P	09/25/2025	12:12	LB137314
CCV02	Cadmium	2460	2500	99	90 - 110	P	09/25/2025	13:23	LB137314
	Copper	1250	1250	100	90 - 110	P	09/25/2025	13:23	LB137314
	Lead	4880	5000	98	90 - 110	P	09/25/2025	13:23	LB137314
	Nickel	2460	2500	99	90 - 110	P	09/25/2025	13:23	LB137314
	Zinc	2630	2500	105	90 - 110	P	09/25/2025	13:23	LB137314
CCV03	Cadmium	2440	2500	98	90 - 110	P	09/25/2025	14:21	LB137314
	Copper	1240	1250	99	90 - 110	P	09/25/2025	14:21	LB137314
	Lead	4840	5000	97	90 - 110	P	09/25/2025	14:21	LB137314
	Nickel	2440	2500	98	90 - 110	P	09/25/2025	14:21	LB137314
	Zinc	2470	2500	99	90 - 110	P	09/25/2025	14:21	LB137314
CCV04	Cadmium	2440	2500	98	90 - 110	P	09/25/2025	15:16	LB137314
	Copper	1240	1250	99	90 - 110	P	09/25/2025	15:16	LB137314
	Lead	4830	5000	97	90 - 110	P	09/25/2025	15:16	LB137314
	Nickel	2440	2500	98	90 - 110	P	09/25/2025	15:16	LB137314
	Zinc	2510	2500	100	90 - 110	P	09/25/2025	15:16	LB137314
CCV05	Cadmium	2520	2500	101	90 - 110	P	09/25/2025	16:16	LB137314
	Copper	1290	1250	103	90 - 110	P	09/25/2025	16:16	LB137314
	Lead	4990	5000	100	90 - 110	P	09/25/2025	16:16	LB137314
	Nickel	2520	2500	101	90 - 110	P	09/25/2025	16:16	LB137314
	Zinc	2640	2500	106	90 - 110	P	09/25/2025	16:16	LB137314
CCV06	Cadmium	2470	2500	99	90 - 110	P	09/25/2025	17:27	LB137314
	Copper	1260	1250	101	90 - 110	P	09/25/2025	17:27	LB137314
	Lead	4880	5000	98	90 - 110	P	09/25/2025	17:27	LB137314
	Nickel	2460	2500	99	90 - 110	P	09/25/2025	17:27	LB137314
	Zinc	2520	2500	101	90 - 110	P	09/25/2025	17:27	LB137314
CCV07	Cadmium	2520	2500	101	90 - 110	P	09/25/2025	17:40	LB137314
	Copper	1280	1250	103	90 - 110	P	09/25/2025	17:40	LB137314
	Lead	4980	5000	100	90 - 110	P	09/25/2025	17:40	LB137314
	Nickel	2520	2500	101	90 - 110	P	09/25/2025	17:40	LB137314

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3125

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Zinc	2590	2500	104	90 - 110	P	09/25/2025	17:40	LB137314
