

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Aramark Uniforms</u>	SDG No.: <u>Q3125</u>
Contract: <u>ARAM01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	1.36	1.0	136	-5	7	09/25/2025	11:42	LB137314
	Copper	13.1	2.0	655	-18	22	09/25/2025	11:42	LB137314
	Lead	-7.23			-12	12	09/25/2025	11:42	LB137314
	Nickel	3.09	2.0	154	-38	42	09/25/2025	11:42	LB137314
	Zinc	1.46			-40	40	09/25/2025	11:42	LB137314
ICSAB01	Cadmium	1030	972	106	826	1120	09/25/2025	11:46	LB137314
	Copper	511	511	100	434	588	09/25/2025	11:46	LB137314
	Lead	41.9	49.0	86	37	61	09/25/2025	11:46	LB137314
	Nickel	1030	954	108	810	1100	09/25/2025	11:46	LB137314
	Zinc	1090	952	114	809	1095	09/25/2025	11:46	LB137314
ICSA	Cadmium	-1.82	1.0	182	-5	7	09/25/2025	11:50	LB137314
	Copper	14.0	2.0	700	-18	22	09/25/2025	11:50	LB137314
	Lead	-20.1			-12	12	09/25/2025	11:50	LB137314
	Nickel	8.41	2.0	420	-38	42	09/25/2025	11:50	LB137314
	Zinc	147			-40	40	09/25/2025	11:50	LB137314
ICSAB	Cadmium	852	972	88	826	1120	09/25/2025	12:00	LB137314
	Copper	457	511	89	434	588	09/25/2025	12:00	LB137314
	Lead	21.8	49.0	44	37	61	09/25/2025	12:00	LB137314
	Nickel	858	954	90	810	1100	09/25/2025	12:00	LB137314
	Zinc	954	952	100	809	1095	09/25/2025	12:00	LB137314