

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>Q3125</u>
contract: <u>ARAM01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3125-02</u>	client id: <u>CompMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3125-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	96.8		1.00	J	100	96		P
Copper	ug/L	75 - 125	225		88.4		150	91		P
Lead	ug/L	75 - 125	476		20.0		500	91		P
Mercury	ug/L	75 - 125	3.82		0.10	J	4.0	93		CV
Nickel	ug/L	75 - 125	254		13.6	J	250	96		P
Zinc	ug/L	75 - 125	568		500		100	68		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>Q3125</u>
contract: <u>ARAM01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3125-02</u>	client id: <u>CompMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3125-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	97.0		1.00	J	100	96		P
Copper	ug/L	75 - 125	225		88.4		150	91		P
Lead	ug/L	75 - 125	479		20.0		500	92		P
Mercury	ug/L	75 - 125	3.88		0.10	J	4.0	94		CV
Nickel	ug/L	75 - 125	255		13.6	J	250	97		P
Zinc	ug/L	75 - 125	563		500		100	64		P