

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3127

Client ID : SSTDCCC0.4

Date Analyzed: 09/19/2025

Lab File ID: BN037795.D

Time Analyzed: 02:47

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	7724	7.833	19361	10.64	8621	14.48
UPPER LIMIT	15448	8.333	38722	11.137	17242	14.984
LOWER LIMIT	3862	7.333	9680.5	10.137	4310.5	13.984
EPA SAMPLE NO.						
01 TT119S1-20250916	5936	7.83	15845	10.64	6915	14.48
02 TT119S1-20250916MS	5852	7.83	15625	10.64	7051	14.48
03 TT119S1-20250916MSD	5675	7.83	15407	10.64	6857	14.48

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.