

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3127

Client ID : SSTDCCC0.4

Date Analyzed: 09/22/2025

Lab File ID: BN037849.D

Time Analyzed: 10:19

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	6007	7.833	15935	10.64	7024	14.48
UPPER LIMIT	12014	8.333	31870	11.137	14048	14.984
LOWER LIMIT	3003.5	7.333	7967.5	10.137	3512	13.984
EPA SAMPLE NO.						
01 PB169743BL	6064	7.83	15212	10.63	6032	14.48
02 RE131D1-20250915DL	5210	7.83	13414	10.63	5405	14.48
03 RE131D2-20250915DL	5543	7.83	14287	10.63	5844	14.48
04 DUP07-20250915DL	5382	7.83	14120	10.63	5769	14.48
05 RE132D6-20250916DL	5456	7.83	14189	10.63	5837	14.48
06 TT101D2-20250915DL	5466	7.83	14302	10.63	5898	14.48

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.