

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
 Data File : VX047795.D
 Acq On : 25 Sep 2025 08:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 26 01:20:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	78	-0.01
2 T	Dichlorodifluoromethane	0.518	0.627	-21.0	98	0.00
3 P	Chloromethane	0.680	0.789	-16.0	94	0.00
4 C	Vinyl Chloride	0.666	0.779	-17.0#	92	0.00
5 T	Bromomethane	0.422	0.471	-11.6	87	0.00
6 T	Chloroethane	0.445	0.495	-11.2	88	0.00
7 T	Trichlorofluoromethane	0.989	1.083	-9.5	85	0.00
8 T	Diethyl Ether	0.405	0.451	-11.4	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.578	0.618	-6.9	83	0.00
10 T	Methyl Iodide	0.902	0.964	-6.9	83	0.00
11 T	Tert butyl alcohol	0.089	0.100	-12.4	88	0.00
12 CM	1,1-Dichloroethene	0.594	0.663	-11.6#	86	0.00
13 T	Acrolein	0.082	0.095	-15.9	92	0.00
14 T	Allyl chloride	1.226	1.201	2.0	78	0.00
15 T	Acrylonitrile	0.328	0.414	-26.2#	94	0.00
16 T	Acetone	0.346	0.336	2.9	83	0.00
17 T	Carbon Disulfide	1.626	1.765	-8.5	89	0.00
18 T	Methyl Acetate	0.770	0.985	-27.9#	88	0.00
19 T	Methyl tert-butyl Ether	2.169	2.414	-11.3	84	0.00
20 T	Methylene Chloride	0.732	0.817	-11.6	88	0.00
21 T	trans-1,2-Dichloroethene	0.640	0.703	-9.8	85	0.00
22 T	Diisopropyl ether	2.376	2.620	-10.3	83	-0.01
23 T	Vinyl Acetate	1.901	2.117	-11.4	83	0.00
24 P	1,1-Dichloroethane	1.263	1.370	-8.5	83	0.00
25 T	2-Butanone	0.432	0.496	-14.8	89	0.00
26 T	2,2-Dichloropropane	1.050	0.538	48.8#	40#	0.00
27 T	cis-1,2-Dichloroethene	0.783	0.854	-9.1	83	-0.01
28 T	Bromochloromethane	0.551	0.621	-12.7	84	-0.01
29 T	Tetrahydrofuran	0.260	0.319	-22.7	91	0.00
30 C	Chloroform	1.276	1.421	-11.4#	84	0.00
31 T	Cyclohexane	1.052	1.056	-0.4	80	0.00
32 T	1,1,1-Trichloroethane	1.082	1.142	-5.5	80	0.00
33 S	1,2-Dichloroethane-d4	0.796	0.817	-2.6	83	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.335	0.351	-4.8	85	0.00
36 T	1,1-Dichloropropene	0.490	0.485	1.0	80	-0.01
37 T	Ethyl Acetate	0.525	0.587	-11.8	88	-0.01
38 T	Carbon Tetrachloride	0.532	0.538	-1.1	80	-0.01
39 T	Methylcyclohexane	0.556	0.530	4.7	75	0.00
40 TM	Benzene	1.488	1.598	-7.4	84	0.00
41 T	Methacrylonitrile	0.271	0.316	-16.6	89	0.00
42 TM	1,2-Dichloroethane	0.566	0.594	-4.9	82	0.00
43 T	Isopropyl Acetate	0.860	0.943	-9.7	84	0.00
44 TM	Trichloroethene	0.360	0.380	-5.6	82	0.00
45 C	1,2-Dichloropropane	0.381	0.418	-9.7#	84	0.00
46 T	Dibromomethane	0.277	0.297	-7.2	83	0.00
47 T	Bromodichloromethane	0.572	0.613	-7.2	81	0.00
48 T	Methyl methacrylate	0.429	0.477	-11.2	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
 Data File : VX047795.D
 Acq On : 25 Sep 2025 08:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 26 01:20:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.004	0.006	-50.0#	96	0.00
50 S	Toluene-d8	1.160	1.182	-1.9	82	0.00
51 T	4-Methyl-2-Pentanone	0.477	0.566	-18.7	89	0.00
52 CM	Toluene	0.917	0.962	-4.9#	82	0.00
53 T	t-1,3-Dichloropropene	0.584	0.544	6.8	70	0.00
54 T	cis-1,3-Dichloropropene	0.624	0.601	3.7	74	0.00
55 T	1,1,2-Trichloroethane	0.354	0.390	-10.2	85	0.00
56 T	Ethyl methacrylate	0.555	0.624	-12.4	83	0.00
57 T	1,3-Dichloropropane	0.607	0.678	-11.7	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.227	0.323	-42.3#	95	0.00
59 T	2-Hexanone	0.343	0.392	-14.3	87	0.00
60 T	Dibromochloromethane	0.407	0.449	-10.3	83	0.00
61 T	1,2-Dibromoethane	0.360	0.409	-13.6	86	0.00
62 S	4-Bromofluorobenzene	0.440	0.443	-0.7	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.326	0.353	-8.3	88	0.00
65 PM	Chlorobenzene	1.136	1.192	-4.9	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.392	0.413	-5.4	81	0.00
67 C	Ethyl Benzene	1.960	2.012	-2.7#	80	0.00
68 T	m/p-Xylenes	0.729	0.757	-3.8	81	0.00
69 T	o-Xylene	0.706	0.737	-4.4	80	0.00
70 T	Styrene	1.236	1.303	-5.4	82	0.00
71 P	Bromoform	0.293	0.318	-8.5	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
73 T	Isopropylbenzene	3.801	3.828	-0.7	80	0.00
74 T	N-amyl acetate	1.733	1.835	-5.9	83	0.00
75 P	1,1,2,2-Tetrachloroethane	1.150	1.273	-10.7	87	0.00
76 T	1,2,3-Trichloropropane	1.012	1.006	0.6	75	0.00
77 T	Bromobenzene	0.936	0.969	-3.5	83	0.00
78 T	n-propylbenzene	4.494	4.460	0.8	78	0.00
79 T	2-Chlorotoluene	2.752	2.749	0.1	79	0.00
80 T	1,3,5-Trimethylbenzene	3.123	3.119	0.1	79	0.00
81 T	trans-1,4-Dichloro-2-butene	0.397	0.321	19.1	65	0.00
82 T	4-Chlorotoluene	3.251	3.219	1.0	79	0.00
83 T	tert-Butylbenzene	3.197	3.051	4.6	76	0.00
84 T	1,2,4-Trimethylbenzene	3.166	3.187	-0.7	80	0.00
85 T	sec-Butylbenzene	3.797	3.713	2.2	78	0.00
86 T	p-Isopropyltoluene	3.215	3.126	2.8	77	0.00
87 T	1,3-Dichlorobenzene	1.739	1.748	-0.5	81	0.00
88 T	1,4-Dichlorobenzene	1.785	1.750	2.0	80	0.00
89 T	n-Butylbenzene	2.975	2.782	6.5	75	0.00
90 T	Hexachloroethane	0.564	0.565	-0.2	80	0.00
91 T	1,2-Dichlorobenzene	1.657	1.689	-1.9	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.233	0.250	-7.3	83	0.00
93 T	1,2,4-Trichlorobenzene	1.077	1.039	3.5	76	0.00
94 T	Hexachlorobutadiene	0.366	0.338	7.7	74	0.00
95 T	Naphthalene	3.279	3.435	-4.8	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
Data File : VX047795.D
Acq On : 25 Sep 2025 08:56
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 26 01:20:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.002	1.016	-1.4	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6