

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX091625

Review By	John Carlone	Review On	9/17/2025 8:27:00 AM
Supervise By	Mahesh Dadoda	Supervise On	9/18/2025 3:22:52 PM
SubDirectory	VX091625	HP Acquire Method	HP Processing Method 82X091625W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP135487		
Initial Calibration Stds	VP135489,VP135490,VP135491,VP135492,VP135493,VP135494		
CCC	VP135488		
Internal Standard/PEM			
ICV/I.BLK	VP135495		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047584.D	16 Sep 2025 08:56		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX047585.D	16 Sep 2025 09:23		JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX047586.D	16 Sep 2025 10:16	QR-58	JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX047587.D	16 Sep 2025 10:37		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX047588.D	16 Sep 2025 10:58		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX047589.D	16 Sep 2025 11:40		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX047590.D	16 Sep 2025 12:01		JC/MD	Ok,M
8	IBLK	IBLK	VX047591.D	16 Sep 2025 12:22		JC/MD	Ok
9	VSTDICV050	ICVVX091625	VX047592.D	16 Sep 2025 13:35		JC/MD	Ok,M
10	VX0916MBL01	VX0916MBL01	VX047593.D	16 Sep 2025 14:03		JC/MD	Ok
11	VX0916WBL01	VX0916WBL01	VX047594.D	16 Sep 2025 14:53		JC/MD	Ok
12	VX0916WBS01	VX0916WBS01	VX047595.D	16 Sep 2025 15:21		JC/MD	Ok,M
13	VX0916WBSD01	VX0916WBSD01	VX047596.D	16 Sep 2025 15:46		JC/MD	Ok,M
14	Q3015-01	WP0925-PT-VOA-WP	VX047597.D	16 Sep 2025 16:07		JC/MD	Ok,M
15	VIBLK	VIBLK	VX047598.D	16 Sep 2025 16:30		JC/MD	Ok
16	VSTDIC050	VSTDIC050EC	VX047599.D	16 Sep 2025 16:51		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX092325

Review By	John Carlone	Review On	9/25/2025 10:19:51 AM
Supervise By	Maresh Dadoda	Supervise On	9/26/2025 1:43:07 AM
SubDirectory	VX092325	HP Acquire Method	HP Processing Method 82X091625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135588,VP135591
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135589,VP135590,VP135593,VP135594

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047718.D	23 Sep 2025 08:05		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX047719.D	23 Sep 2025 09:02	pH#Lot#V12668	JC/MD	Ok,M
3	VX0923MBL01	VX0923MBL01	VX047720.D	23 Sep 2025 09:31		JC/MD	Ok
4	VX0923WBL01	VX0923WBL01	VX047721.D	23 Sep 2025 10:22		JC/MD	Ok
5	VX0923WBS01	VX0923WBS01	VX047722.D	23 Sep 2025 10:48	Recovery fail	JC/MD	Not Ok
6	VX0923WBS02	VX0923WBS02	VX047723.D	23 Sep 2025 11:15		JC/MD	Ok,M
7	Q3097-01DL	RE103D2-20250911DL	VX047724.D	23 Sep 2025 11:36	vial B pH<2	JC/MD	Ok
8	Q3141-01	BP-VPB-184-TB-20250	VX047725.D	23 Sep 2025 12:22	vial A pH<2 TB	JC/MD	Ok
9	VX0923WBSD02	VX0923WBSD02	VX047726.D	23 Sep 2025 12:43		JC/MD	Ok,M
10	Q3141-02	BP-VPB-184-EB-20250	VX047727.D	23 Sep 2025 13:04	vial A pH<2 EB;Hit of Com.#16,25	JC/MD	Not Ok
11	Q3097-18	TB05-20250912	VX047728.D	23 Sep 2025 13:25	vial A pH<2 TB	JC/MD	Ok
12	Q3097-16	TB03-20250911	VX047729.D	23 Sep 2025 13:46	vial A pH<2 TB	JC/MD	Ok
13	Q3141-03	BP-VPB-184-GW-60-62	VX047730.D	23 Sep 2025 14:07	vial A pH<2	JC/MD	Ok
14	Q3141-04	BP-VPB-184-GW-100-1	VX047731.D	23 Sep 2025 14:28	vial A pH<2 Surrogate Fail	JC/MD	Not Ok
15	Q3141-05	BP-VPB-184-GW-150-1	VX047732.D	23 Sep 2025 14:49	vial A pH<2	JC/MD	Ok
16	Q3141-06	BP-VPB-184-GW-200-2	VX047733.D	23 Sep 2025 15:10	vial A pH<2	JC/MD	Ok
17	Q3141-07	BP-VPB-184-GW-220-2	VX047734.D	23 Sep 2025 15:31	vial A pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX092325

Review By	John Carlone	Review On	9/25/2025 10:19:51 AM
Supervise By	Maresh Dadoda	Supervise On	9/26/2025 1:43:07 AM
SubDirectory	VX092325	HP Acquire Method	HP Processing Method 82X091625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135588,VP135591		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135589,VP135590,VP135593,VP135594		

18	Q3097-15	RE117D2-20250911	VX047735.D	23 Sep 2025 15:52	vial A pH<2	JC/MD	Ok
19	Q3097-17	TB04-20250911	VX047736.D	23 Sep 2025 16:13	vial A pH<2 TB	JC/MD	Ok
20	Q3097-19	RE120D1-20250912	VX047737.D	23 Sep 2025 16:34	vial A pH<2 Need 10X	JC/MD	Dilution
21	Q3097-20	DUP06-20250912	VX047738.D	23 Sep 2025 16:55	vial A pH<2 Need 10X	JC/MD	Dilution
22	Q3097-21	RE120D2-20250912	VX047739.D	23 Sep 2025 17:16	vial A pH<2 Need 5X	JC/MD	Dilution
23	Q3097-22	RE134D1-20250912	VX047740.D	23 Sep 2025 17:36	vial A pH<2	JC/MD	Ok
24	Q3097-23	RE134D4-20250912	VX047741.D	23 Sep 2025 17:57	vial A pH<2 Need 10X	JC/MD	Dilution
25	Q3097-24	RE134D2-20250912	VX047742.D	23 Sep 2025 18:18	vial A pH<2	JC/MD	Ok
26	Q3151-02	1402	VX047743.D	23 Sep 2025 18:39	vial A pH<2	JC/MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VX047744.D	23 Sep 2025 19:00		JC/MD	Ok,M
28	BFB	BFB	VX047745.D	24 Sep 2025 01:30		JC/MD	Ok
29	VSTDCCC050	VSTDCCC050	VX047746.D	24 Sep 2025 02:08	CCC failed High For Com.#02	JC/MD	Ok,M
30	VX0923WBL02	VX0923WBL02	VX047747.D	24 Sep 2025 03:11		JC/MD	Ok
31	VX0923WBS03	VX0923WBS03	VX047748.D	24 Sep 2025 03:53		JC/MD	Ok,M
32	VX0923WBSD03	VX0923WBSD03	VX047749.D	24 Sep 2025 04:14		JC/MD	Ok,M
33	Q3127-03	RE131D2-20250915	VX047750.D	24 Sep 2025 04:52	vial A pH<2 Need 5X	JC/MD	Dilution
34	Q3127-05	TT158I1-20250915	VX047751.D	24 Sep 2025 05:13	vial A pH<2	JC/MD	Ok
35	Q3127-06	TT180D2-20250915	VX047752.D	24 Sep 2025 05:34	vial A pH<2 Need 10X	JC/MD	Dilution
36	Q3127-08	RE132D6-20250916	VX047753.D	24 Sep 2025 05:55	vial A pH<2 Need 50X	JC/MD	Dilution

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX092325

Review By	John Carlone	Review On	9/25/2025 10:19:51 AM
Supervise By	Maresh Dadoda	Supervise On	9/26/2025 1:43:07 AM
SubDirectory	VX092325	HP Acquire Method	HP Processing Method 82X091625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135588,VP135591		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135589,VP135590,VP135593,VP135594		

37	Q3127-09	TT101D2-20250915	VX047754.D	24 Sep 2025 06:16	vial A pH<2 Need 50X	JC/MD	Dilution
38	Q3127-10	TT101D-20250915	VX047755.D	24 Sep 2025 06:37	vial A pH<2 E flag of previous sample;CCAL Failed High	JC/MD	ReRun
39	Q3127-12	RE132D7-20250916	VX047756.D	24 Sep 2025 06:58	vial A pH<2 Need 50X	JC/MD	Dilution
40	Q3127-16	TT101D1-20250915	VX047757.D	24 Sep 2025 07:19	vial A pH<2 CCC Failed High; Need 5X	JC/MD	Dilution
41	Q3127-18	BPOW1-8-20250917	VX047758.D	24 Sep 2025 07:40	vial A pH<2 E flag of previous sample	JC/MD	ReRun
42	Q3127-07	RE135D2-20250915	VX047759.D	24 Sep 2025 08:01	vial A pH<2 Confirm Concentration	JC/MD	Not Ok
43	Q3141-04	BP-VPB-184-GW-100-1	VX047760.D	24 Sep 2025 08:22	vial B pH<2	JC/MD	Ok
44	Q3141-02	BP-VPB-184-EB-20250	VX047761.D	24 Sep 2025 08:43	vial B pH<2 EB	JC/MD	Ok
45	Q3127-04	DUP07-20250915	VX047762.D	24 Sep 2025 09:04	vial A pH<2 Need 5X	JC/MD	Dilution
46	Q3097-25	RE134D3-20250912	VX047763.D	24 Sep 2025 09:25	vial A pH<2 Need 10X	JC/MD	Dilution
47	Q3127-13	TT119S1-20250916	VX047764.D	24 Sep 2025 09:46	vial A pH<2 E flag of previous sample	JC/MD	ReRun
48	Q3127-14MS	TT119S1-20250916MS	VX047765.D	24 Sep 2025 10:07	vial A pH<2	JC/MD	Ok,M
49	Q3127-15MSD	TT119S1-20250916MS	VX047766.D	24 Sep 2025 10:29	vial A pH<2	JC/MD	Ok,M
50	VSTDCCC050	VSTDCCC050EC	VX047767.D	24 Sep 2025 10:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX092425

Review By	John Carlone	Review On	9/25/2025 10:58:56 AM
Supervise By	Mahesh Dadoda	Supervise On	9/26/2025 1:47:46 AM
SubDirectory	VX092425	HP Acquire Method	HP Processing Method 82X091625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135605,VP135608
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135606,VP135607,VP135609,VP135610

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047768.D	24 Sep 2025 11:16		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX047769.D	24 Sep 2025 11:46	pH#Lot#V12668	JC/MD	Ok,M
3	VX0924MBL01	VX0924MBL01	VX047770.D	24 Sep 2025 12:14		JC/MD	Ok
4	VX0924WBL01	VX0924WBL01	VX047771.D	24 Sep 2025 13:04		JC/MD	Ok
5	VX0924WBS01	VX0924WBS01	VX047772.D	24 Sep 2025 13:59		JC/MD	Ok,M
6	Q3097-21DL	RE120D2-20250912DL	VX047773.D	24 Sep 2025 14:20	vial B pH<2	JC/MD	Ok
7	VX0924WBSD01	VX0924WBSD01	VX047774.D	24 Sep 2025 14:41		JC/MD	Ok,M
8	Q3097-19DL	RE120D1-20250912DL	VX047775.D	24 Sep 2025 15:02	vial B pH<2	JC/MD	Ok
9	Q3097-20DL	DUP06-20250912DL	VX047776.D	24 Sep 2025 15:23	vial B pH<2	JC/MD	Ok
10	Q3097-23DL	RE134D4-20250912DL	VX047777.D	24 Sep 2025 15:44	vial B pH<2	JC/MD	Ok
11	Q3097-25DL	RE134D3-20250912DL	VX047778.D	24 Sep 2025 16:05	vial B pH<2	JC/MD	Ok
12	Q3097-22RE	RE134D1-20250912RE	VX047779.D	24 Sep 2025 16:25	vial B pH<2 End CCC Missing;Hit Confirmed	JC/MD	Not Ok
13	Q3127-03DL	RE131D2-20250915DL	VX047780.D	24 Sep 2025 16:46	vial B pH<2	JC/MD	Ok
14	Q3127-16DL	TT101D1-20250915DL	VX047781.D	24 Sep 2025 17:07	vial B pH<2	JC/MD	Ok
15	Q3127-06DL	TT180D2-20250915DL	VX047782.D	24 Sep 2025 17:28	vial B pH<2	JC/MD	Ok
16	Q3127-08DL	RE132D6-20250916DL	VX047783.D	24 Sep 2025 17:49	vial B pH<2	JC/MD	Ok
17	Q3127-09DL	TT101D2-20250915DL	VX047784.D	24 Sep 2025 18:09	vial B pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX092425

Review By	John Carlone	Review On	9/25/2025 10:58:56 AM
Supervise By	Maresh Dadoda	Supervise On	9/26/2025 1:47:46 AM
SubDirectory	VX092425	HP Acquire Method	HP Processing Method 82X091625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135605,VP135608		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135606,VP135607,VP135609,VP135610		

18	Q3127-12DL	RE132D7-20250916DL	VX047785.D	24 Sep 2025 18:30	vial B pH<2	JC/MD	Ok
19	Q3127-07	RE135D2-20250915	VX047786.D	24 Sep 2025 18:51	vial B pH<2	JC/MD	Ok
20	Q3127-10RE	TT101D-20250915RE	VX047787.D	24 Sep 2025 19:12	vial B pH<2 End CCC Missing	JC/MD	Confirms
21	Q3127-18	BPOW1-8-20250917	VX047788.D	24 Sep 2025 19:33	vial B pH<2	JC/MD	Ok
22	Q3127-02	RE131D1-20250915	VX047789.D	24 Sep 2025 19:54	vial A pH<2 Need 5X	JC/MD	Dilution
23	Q3127-19	BPOW1-7-20250917	VX047790.D	24 Sep 2025 20:14	End CCC Missing	JC/MD	Not Ok

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX092525

Review By	Semsettin Yesilyurt	Review On	9/29/2025 1:10:31 AM
Supervise By	Mahesh Dadoda	Supervise On	9/29/2025 1:11:53 AM
SubDirectory	VX092525	HP Acquire Method	HP Processing Method 82X091625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135628,VP135646
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135629,VP135630,VP135647,VP135648

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047794.D	25 Sep 2025 08:24		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX047795.D	25 Sep 2025 08:56	pH#Lot#V12668	JC/MD	Ok,M
3	VX0925MBL01	VX0925MBL01	VX047796.D	25 Sep 2025 09:25		JC/MD	Ok
4	VX0925WBL01	VX0925WBL01	VX047797.D	25 Sep 2025 09:46		JC/MD	Ok
5	VX0925WBS01	VX0925WBS01	VX047798.D	25 Sep 2025 10:13		JC/MD	Ok,M
6	VX0925WBSD01	VX0925WBSD01	VX047799.D	25 Sep 2025 10:42		JC/MD	Ok,M
7	Q3165-02	VPB184-HYD-2025091	VX047800.D	25 Sep 2025 11:03	vial A pH<2	JC/MD	Ok
8	Q3165-01	BP-VPB-184-TB-20250	VX047801.D	25 Sep 2025 11:24	vial A pH<2 TB	JC/MD	Ok
9	Q3165-03	BP-VPB-184-GW-240-2	VX047802.D	25 Sep 2025 11:46	vial A pH<2	JC/MD	Ok
10	Q3165-04	BP-VPB-184-GW-260-2	VX047803.D	25 Sep 2025 12:07	vial A pH<2	JC/MD	Ok
11	Q3165-05	BP-VPB-184-GW-280-2	VX047804.D	25 Sep 2025 12:28	vial A pH<2	JC/MD	Ok
12	Q3165-06	BP-VPB-184-DUP-2025	VX047805.D	25 Sep 2025 12:49	vial A pH<2	JC/MD	Ok
13	Q3165-07	BP-VPB-184-GW-300-3	VX047806.D	25 Sep 2025 13:10	vial A pH<2	JC/MD	Ok
14	Q3127-04DL	DUP07-20250915DL	VX047807.D	25 Sep 2025 13:31	vial B pH<2	JC/MD	Ok
15	Q3127-17	TB06-20250912	VX047808.D	25 Sep 2025 13:53	vial A pH<2 TB	JC/MD	Ok
16	Q3127-13	TT119S1-20250916	VX047809.D	25 Sep 2025 14:14	vial B pH<2	JC/MD	Ok
17	Q3128-07	BPOW6-2-20250916	VX047810.D	25 Sep 2025 14:35	vial A pH<2	JC/MD	Ok
18	Q3128-08	BPOW6-10-20250916	VX047811.D	25 Sep 2025 14:56	vial A pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX092525

Review By	Semsettin Yesilyurt	Review On	9/29/2025 1:10:31 AM
Supervise By	Maresh Dadoda	Supervise On	9/29/2025 1:11:53 AM
SubDirectory	VX092525	HP Acquire Method	HP Processing Method 82X091625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135628,VP135646		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135629,VP135630,VP135647,VP135648		

19	Q3128-09	BPOW6-11-20250916	VX047812.D	25 Sep 2025 15:17	vial A pH<2	JC/MD	Ok
20	Q3128-10	BPOW6-8-20250916	VX047813.D	25 Sep 2025 15:38	vial A pH<2	JC/MD	Ok
21	Q3128-11	BPOW6-7-20250917	VX047814.D	25 Sep 2025 15:59	vial A pH<2	JC/MD	Ok
22	Q3128-12	BPOW6-6-20250917	VX047815.D	25 Sep 2025 16:21	vial A pH<2	JC/MD	Ok
23	Q3128-13	BPOW6-5-20250917	VX047816.D	25 Sep 2025 16:42	vial A pH<2	JC/MD	Ok
24	Q3127-01	DUP08-20250915	VX047817.D	25 Sep 2025 17:03	vial A pH<2	JC/MD	Ok
25	Q3127-11	RE135D3-20250915	VX047818.D	25 Sep 2025 17:24	vial A pH<2	JC/MD	Ok
26	Q3127-19	BPOW1-7-20250917	VX047819.D	25 Sep 2025 17:45	vial A pH<2	JC/MD	Ok
27	Q3165-08	BP-VPB-184-GW-320-3	VX047820.D	25 Sep 2025 18:06	vial A pH<2	JC/MD	Ok
28	Q3165-09MS	Q3165-08MSMS	VX047821.D	25 Sep 2025 18:27	vial A pH<2	JC/MD	Ok,M
29	Q3165-10MSD	Q3165-08MSDMSD	VX047822.D	25 Sep 2025 18:48	vial A pH<2	JC/MD	Ok,M
30	VSTDCCC050	VSTDCCC050EC	VX047823.D	25 Sep 2025 19:09		JC/MD	Ok,M
31	BFB	BFB	VX047824.D	26 Sep 2025 00:46		JC/MD	Ok
32	VSTDCCC050	VSTDCCC050	VX047825.D	26 Sep 2025 01:28		JC/MD	Ok,M
33	VX0925WBL02	VX0925WBL02	VX047826.D	26 Sep 2025 02:10		JC/MD	Ok
34	VX0925WBS02	VX0925WBS02	VX047827.D	26 Sep 2025 02:31		JC/MD	Ok,M
35	VX0925WBSD02	VX0925WBSD02	VX047828.D	26 Sep 2025 02:52	Surrogate Fail	JC/MD	Not Ok
36	Q3156-02	RE10-MW01S-20250918	VX047829.D	26 Sep 2025 03:31	vial A pH<2	JC/MD	Ok
37	Q3156-03	RE131D3-20250918	VX047830.D	26 Sep 2025 03:52	vial A pH<2 Need 10X	JC/MD	Dilution

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX092525

Review By	Semsettin Yesilyurt	Review On	9/29/2025 1:10:31 AM
Supervise By	Mahesh Dadoda	Supervise On	9/29/2025 1:11:53 AM
SubDirectory	VX092525	HP Acquire Method	HP Processing Method 82X091625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135628,VP135646		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135629,VP135630,VP135647,VP135648		

38	Q3156-04	TT190D1-20250918	VX047831.D	26 Sep 2025 04:13	vial A pH<2 Need 5X	JC/MD	Dilution
39	Q3156-05	RE132D2-20250918	VX047832.D	26 Sep 2025 04:34	vial A pH<2 E flag of previous sample	JC/MD	ReRun
40	Q3156-06	RE132D4-20250918	VX047833.D	26 Sep 2025 04:55	vial A pH<2 Need 5X	JC/MD	Dilution
41	Q3156-07	TT189D2-20250918	VX047834.D	26 Sep 2025 05:16	vial A pH<2	JC/MD	Ok
42	Q3156-08	TT149I1-20250918	VX047835.D	26 Sep 2025 05:37	vial A pH<2	JC/MD	Ok
43	Q3156-09	RE132D3-20250918	VX047836.D	26 Sep 2025 05:58	vial A pH<2	JC/MD	Ok
44	Q3156-11	RE132D1-20250918	VX047837.D	26 Sep 2025 06:19	vial A pH<2	JC/MD	Ok
45	Q3156-12	TT190D2-20250918	VX047838.D	26 Sep 2025 06:40	vial A pH<2	JC/MD	Ok
46	Q3128-04	BPOW6-4-20250916	VX047839.D	26 Sep 2025 07:01	vial A pH<2	JC/MD	Ok
47	Q3128-05	BPOW6-9-20250916	VX047840.D	26 Sep 2025 07:22	vial A pH<2	JC/MD	Ok
48	Q3128-06	BPOW6-3-20250916	VX047841.D	26 Sep 2025 07:43	vial A pH<2	JC/MD	Ok
49	Q3128-14	TB07-20250912	VX047842.D	26 Sep 2025 08:04	vial A pH<2 TB	JC/MD	Ok
50	Q3128-01	BPOW6-1-20250916	VX047843.D	26 Sep 2025 08:25	vial A pH<2	JC/MD	Ok
51	Q3128-02MS	BPOW6-1-20250916MS	VX047844.D	26 Sep 2025 08:46	vial A pH<2	JC/MD	Ok,M
52	Q3128-03MSD	BPOW6-1-20250916MS	VX047845.D	26 Sep 2025 09:07	vial A pH<2	JC/MD	Ok,M
53	VSTDCCC050	VSTDCCC050EC	VX047846.D	26 Sep 2025 09:28		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX092625

Review By	John Carlone	Review On	9/29/2025 8:28:35 AM
Supervise By	Maresh Dadoda	Supervise On	9/30/2025 2:20:54 AM
SubDirectory	VX092625	HP Acquire Method	HP Processing Method 82X091625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135664
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135665,VP135666

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047847.D	26 Sep 2025 09:54		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX047848.D	26 Sep 2025 10:54	pH#Lot#V12668	JC/MD	Ok,M
3	VX0926MBL01	VX0926MBL01	VX047849.D	26 Sep 2025 11:22		JC/MD	Ok
4	VX0926WBL01	VX0926WBL01	VX047850.D	26 Sep 2025 11:43	see non con	JC/MD	Ok
5	VX0926WBS01	VX0926WBS01	VX047851.D	26 Sep 2025 12:40		JC/MD	Ok,M
6	Q3127-02DL	RE131D1-20250915DL	VX047852.D	26 Sep 2025 13:01	vial B pH<2	JC/MD	Ok
7	VX0926WBSD01	VX0926WBSD01	VX047853.D	26 Sep 2025 13:22		JC/MD	Ok,M
8	Q3156-04DL	TT190D1-20250918DL	VX047854.D	26 Sep 2025 13:43	vial B pH<2	JC/MD	Ok
9	Q3156-06DL	RE132D4-20250918DL	VX047855.D	26 Sep 2025 14:04	vial B pH<2	JC/MD	Ok
10	Q3156-03DL	RE131D3-20250918DL	VX047856.D	26 Sep 2025 14:25	vial B pH<2	JC/MD	Ok
11	Q3156-05	RE132D2-20250918	VX047857.D	26 Sep 2025 14:46	vial B pH<2	JC/MD	Ok
12	Q3143-01	STORAGE-BLANK-SO	VX047858.D	26 Sep 2025 15:07	vial A pH<2	JC/MD	Ok
13	Q3143-02	STORAGE-BLANK-WA	VX047859.D	26 Sep 2025 15:28	vial A pH<2	JC/MD	Ok
14	Q3143-03	STORAGE-BLANK-WA	VX047860.D	26 Sep 2025 15:49	vial A pH<2	JC/MD	Ok
15	Q3143-04	STORAGE-BLANK-SAI	VX047861.D	26 Sep 2025 16:10	vial A pH<2	JC/MD	Ok
16	Q3156-14	EB03-20250918	VX047862.D	26 Sep 2025 16:31	vial A pH<2 EB;Hit of Com.#16,End CCC Missing	JC/MD	ReRun
17	Q3156-15	TB03-20250912	VX047863.D	26 Sep 2025 16:52	TB;End CCC Missing	JC/MD	Not Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX092625

Review By	John Carlone	Review On	9/29/2025 8:28:35 AM
Supervise By	Maresh Dadoda	Supervise On	9/30/2025 2:20:54 AM
SubDirectory	VX092625	HP Acquire Method	HP Processing Method 82X091625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135664
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135665,VP135666

18	Q3156-10	TT180D1-20250918	VX047864.D	26 Sep 2025 17:13	End CCC Missing	JC/MD	Not Ok
19	Q3156-16	RW9-MW01D1-202509	VX047865.D	26 Sep 2025 17:34	End CCC Missing	JC/MD	Not Ok
20	Q3156-17	TT191D2-20250918	VX047866.D	26 Sep 2025 17:54	Not purge;End CCC Missing	JC/MD	Not Ok

M : Manual Integration