

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15
 Lab Code: ACE SDG NO.: Q3127
 Lab File ID: VX047769.D Date Analyzed: 09/24/2025
 Instrument ID: MSVOA_X Time Analyzed: 11:46
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	144936	5.53	262651	6.75	234932	10.04
UPPER LIMIT	289872	6.032	525302	7.245	469864	10.537
LOWER LIMIT	72468	5.032	131326	6.245	117466	9.537
EPA SAMPLE NO.						
RE131D1-20250915	126132	5.54	270621	6.75	270032	10.04
RE131D2-20250915DL	136285	5.54	290275	6.75	295686	10.04
TT180D2-20250915DL	128490	5.54	270303	6.75	276987	10.04
RE135D2-20250915	128068	5.54	276429	6.75	279711	10.04
RE132D6-20250916DL	139103	5.54	297996	6.75	305668	10.04
TT101D2-20250915DL	143989	5.54	304217	6.75	310951	10.04
TT101D-20250915RE	143742	5.54	305012	6.75	309845	10.04
RE132D7-20250916DL	133982	5.54	282418	6.75	290709	10.04
TT101D1-20250915DL	135889	5.54	288957	6.75	293313	10.04
BPOW1-8-20250917	128659	5.54	271871	6.75	278272	10.04
VX0924WBL01	121515	5.53	250677	6.75	250409	10.04
VX0924WBS01	160966	5.54	287921	6.75	264714	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.