

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15  
 Lab Code: ACE SDG NO.: Q3127  
 Lab File ID: VX047795.D Date Analyzed: 09/25/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 08:56  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                  | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|------------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD      | 136107        | 5.53  | 245940        | 6.75  | 220262        | 10.04  |
| UPPER LIMIT      | 272214        | 6.032 | 491880        | 7.245 | 440524        | 10.537 |
| LOWER LIMIT      | 68053.5       | 5.032 | 122970        | 6.245 | 110131        | 9.537  |
| EPA SAMPLE NO.   |               |       |               |       |               |        |
| DUP08-20250915   | 113877        | 5.54  | 245918        | 6.75  | 250155        | 10.04  |
| DUP07-20250915DL | 121002        | 5.54  | 255503        | 6.75  | 261171        | 10.04  |
| RE135D3-20250915 | 186170        | 5.54  | 398423        | 6.75  | 394132        | 10.04  |
| TT119S1-20250916 | 122798        | 5.54  | 255684        | 6.75  | 260354        | 10.04  |
| TB06-20250912    | 125551        | 5.54  | 262060        | 6.75  | 271171        | 10.04  |
| BPOW1-7-20250917 | 187193        | 5.54  | 406514        | 6.75  | 412777        | 10.04  |
| VX0925WBL01      | 117291        | 5.54  | 245721        | 6.75  | 251657        | 10.04  |
| VX0925WBS01      | 146604        | 5.54  | 268605        | 6.75  | 252515        | 10.04  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.