

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15
 Lab Code: ACE SDG NO.: Q3127
 Lab File ID: VX047848.D Date Analyzed: 09/26/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:54
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	157794	5.53	276647	6.75	247479	10.04
UPPER LIMIT	315588	6.031	553294	7.245	494958	10.537
LOWER LIMIT	78897	5.031	138324	6.245	123740	9.537
EPA SAMPLE NO.						
RE131D1-20250915DL	121604	5.54	248889	6.75	250346	10.04
VX0926WBL01	182475	5.54	379369	6.75	380700	10.04
VX0926WBS01	157707	5.54	267870	6.74	239551	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.