

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3127</u>
Lab File ID:	<u>VX047746.D</u>	Date Analyzed:	<u>09/24/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>02:08</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	121948	12.00€				
UPPER LIMIT	243896	12.50€				
LOWER LIMIT	60974	11.50€				
EPA SAMPLE NO.						
RE131D2-20250915	127680	12.01				
DUP07-20250915	124709	12.01				
TT158I1-20250915	195779	12.01				
TT180D2-20250915	134510	12.01				
RE132D6-20250916	154008	12.01				
TT101D2-20250915	144556	12.01				
TT101D-20250915	153772	12.01				
RE132D7-20250916	131134	12.01				
TT119S1-20250916MS	112436	12.01				
TT119S1-20250916MSD	118812	12.01				
TT101D1-20250915	132133	12.01				
VX0923WBL02	171478	12.01				
VX0923WBS03	123765	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.