

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3127</u>
Lab File ID:	<u>VX047769.D</u>	Date Analyzed:	<u>09/24/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>11:46</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	114093	12.00€				
UPPER LIMIT	228186	12.50€				
LOWER LIMIT	57046.5	11.50€				
EPA SAMPLE NO.						
RE131D1-20250915	134111	12.01				
RE131D2-20250915DL	149631	12.01				
TT180D2-20250915DL	135362	12.01				
RE135D2-20250915	142150	12.01				
RE132D6-20250916DL	158858	12.01				
TT101D2-20250915DL	160599	12.01				
TT101D-20250915RE	159800	12.01				
RE132D7-20250916DL	149326	12.01				
TT101D1-20250915DL	150005	12.01				
BPOW1-8-20250917	143672	12.01				
VX0924WBL01	123109	12.01				
VX0924WBS01	129236	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.