

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15  
 Lab Code: ACE SDG NO.: Q3127  
 Lab File ID: VX047795.D Date Analyzed: 09/25/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 08:56  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                  | IS4<br>AREA # | RT #   |  |  |  |  |
|------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD      | 107538        | 12.00€ |  |  |  |  |
| UPPER LIMIT      | 215076        | 12.50€ |  |  |  |  |
| LOWER LIMIT      | 53769         | 11.50€ |  |  |  |  |
| EPA SAMPLE NO.   |               |        |  |  |  |  |
| DUP08-20250915   | 122323        | 12.01  |  |  |  |  |
| DUP07-20250915DL | 129528        | 12.01  |  |  |  |  |
| RE135D3-20250915 | 195911        | 12.01  |  |  |  |  |
| TT119S1-20250916 | 131454        | 12.01  |  |  |  |  |
| TB06-20250912    | 135532        | 12.01  |  |  |  |  |
| BPOW1-7-20250917 | 206806        | 12.01  |  |  |  |  |
| VX0925WBL01      | 126490        | 12.01  |  |  |  |  |
| VX0925WBS01      | 131521        | 12.01  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.