

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance

Contract: AECO15

Lab Code: ACE

SDG No.: Q3128

Instrument ID: MSVOA_X

Calibration Date/Time: 09/25/2025 08:56

Lab File ID: VX047795.D

Init. Calib. Date(s): 09/16/2025 09/16/2025

Heated Purge: (Y/N) N

Init. Calib. Time(s): 09:23 12:01

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.627		21.04	20
Chloromethane	0.680	0.789	0.1	16.03	20
Vinyl Chloride	0.666	0.779		16.97	20
Bromomethane	0.422	0.471		11.61	20
Chloroethane	0.445	0.495		11.24	20
Trichlorofluoromethane	0.989	1.083		9.51	20
1,1,2-Trichlorotrifluoroethane	0.578	0.618		6.92	20
1,1-Dichloroethene	0.594	0.663		11.62	20
Acetone	0.346	0.336		-2.89	20
Carbon Disulfide	1.626	1.765		8.55	20
Methyl tert-butyl Ether	2.169	2.414		11.3	20
Methyl Acetate	0.770	0.985		27.92	20
Methylene Chloride	0.732	0.817		11.61	20
trans-1,2-Dichloroethene	0.640	0.703		9.84	20
1,1-Dichloroethane	1.263	1.370	0.1	8.47	20
Cyclohexane	1.052	1.056		0.38	20
2-Butanone	0.432	0.496		14.81	20
Carbon Tetrachloride	0.532	0.538		1.13	20
cis-1,2-Dichloroethene	0.783	0.854		9.07	20
Bromochloromethane	0.551	0.621		12.7	20
Chloroform	1.276	1.421		11.36	20
1,1,1-Trichloroethane	1.082	1.142		5.55	20
Methylcyclohexane	0.556	0.530		-4.68	20
Benzene	1.488	1.598		7.39	20
1,2-Dichloroethane	0.566	0.594		4.95	20
Trichloroethene	0.360	0.380		5.56	20
1,2-Dichloropropane	0.381	0.418		9.71	20
Bromodichloromethane	0.572	0.613		7.17	20
4-Methyl-2-Pentanone	0.477	0.566		18.66	20
Toluene	0.917	0.962		4.91	20
t-1,3-Dichloropropene	0.584	0.544		-6.85	20
cis-1,3-Dichloropropene	0.624	0.601		-3.69	20
1,1,2-Trichloroethane	0.354	0.390		10.17	20
2-Hexanone	0.343	0.392		14.29	20
Dibromochloromethane	0.407	0.449		10.32	20
1,2-Dibromoethane	0.360	0.409		13.61	20
Tetrachloroethene	0.326	0.353		8.28	20
Chlorobenzene	1.136	1.192	0.3	4.93	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3128</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/25/2025 08:56</u>
Lab File ID:	<u>VX047795.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	2.012		2.65	20
m/p-Xylenes	0.729	0.757		3.84	20
o-Xylene	0.706	0.737		4.39	20
Styrene	1.236	1.303		5.42	20
Bromoform	0.293	0.318	0.1	8.53	20
Isopropylbenzene	3.801	3.828		0.71	20
1,1,2,2-Tetrachloroethane	1.150	1.273	0.3	10.7	20
1,3-Dichlorobenzene	1.739	1.748		0.52	20
1,4-Dichlorobenzene	1.785	1.750		-1.96	20
1,2-Dichlorobenzene	1.657	1.689		1.93	20
1,2-Dibromo-3-Chloropropane	0.233	0.250		7.3	20
1,2,4-Trichlorobenzene	1.077	1.039		-3.53	20
1,2,3-Trichlorobenzene	1.002	1.016		1.4	20
1,2-Dichloroethane-d4	0.796	0.817		2.64	20
Dibromofluoromethane	0.335	0.351		4.78	20
Toluene-d8	1.160	1.182		1.9	20
4-Bromofluorobenzene	0.440	0.443		0.68	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.