

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3128</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/26/2025 01:28</u>
Lab File ID:	<u>VX047825.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.564		8.88	20
Chloromethane	0.680	0.683	0.1	0.44	20
Vinyl Chloride	0.666	0.692		3.9	20
Bromomethane	0.422	0.381		-9.72	20
Chloroethane	0.445	0.437		-1.8	20
Trichlorofluoromethane	0.989	0.984		-0.51	20
1,1,2-Trichlorotrifluoroethane	0.578	0.573		-0.87	20
1,1-Dichloroethene	0.594	0.601		1.18	20
Acetone	0.346	0.311		-10.12	20
Carbon Disulfide	1.626	1.681		3.38	20
Methyl tert-butyl Ether	2.169	2.253		3.87	20
Methyl Acetate	0.770	0.935		21.43	20
Methylene Chloride	0.732	0.749		2.32	20
trans-1,2-Dichloroethene	0.640	0.660		3.13	20
1,1-Dichloroethane	1.263	1.302	0.1	3.09	20
Cyclohexane	1.052	1.000		-4.94	20
2-Butanone	0.432	0.473		9.49	20
Carbon Tetrachloride	0.532	0.512		-3.76	20
cis-1,2-Dichloroethene	0.783	0.806		2.94	20
Bromochloromethane	0.551	0.585		6.17	20
Chloroform	1.276	1.331		4.31	20
1,1,1-Trichloroethane	1.082	1.084		0.19	20
Methylcyclohexane	0.556	0.521		-6.3	20
Benzene	1.488	1.511		1.55	20
1,2-Dichloroethane	0.566	0.556		-1.77	20
Trichloroethene	0.360	0.364		1.11	20
1,2-Dichloropropane	0.381	0.399		4.72	20
Bromodichloromethane	0.572	0.592		3.5	20
4-Methyl-2-Pentanone	0.477	0.550		15.3	20
Toluene	0.917	0.932		1.64	20
t-1,3-Dichloropropene	0.584	0.564		-3.42	20
cis-1,3-Dichloropropene	0.624	0.599		-4.01	20
1,1,2-Trichloroethane	0.354	0.378		6.78	20
2-Hexanone	0.343	0.388		13.12	20
Dibromochloromethane	0.407	0.446		9.58	20
1,2-Dibromoethane	0.360	0.396		10	20
Tetrachloroethene	0.326	0.335		2.76	20
Chlorobenzene	1.136	1.147	0.3	0.97	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3128</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/26/2025</u> <u>01:28</u>
Lab File ID:	<u>VX047825.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	1.966		0.31	20
m/p-Xylenes	0.729	0.737		1.1	20
o-Xylene	0.706	0.722		2.27	20
Styrene	1.236	1.274		3.07	20
Bromoform	0.293	0.319	0.1	8.87	20
Isopropylbenzene	3.801	3.659		-3.74	20
1,1,2,2-Tetrachloroethane	1.150	1.237	0.3	7.57	20
1,3-Dichlorobenzene	1.739	1.702		-2.13	20
1,4-Dichlorobenzene	1.785	1.726		-3.31	20
1,2-Dichlorobenzene	1.657	1.670		0.79	20
1,2-Dibromo-3-Chloropropane	0.233	0.253		8.58	20
1,2,4-Trichlorobenzene	1.077	1.031		-4.27	20
1,2,3-Trichlorobenzene	1.002	1.021		1.9	20
1,2-Dichloroethane-d4	0.796	0.769		-3.39	20
Dibromofluoromethane	0.335	0.335		0	20
Toluene-d8	1.160	1.135		-2.15	20
4-Bromofluorobenzene	0.440	0.436		-0.91	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.