

LAB CHRONICLE

OrderID:	Q3128	OrderDate:	9/17/2025 1:13:00 PM
Client:	AECOM Technical Services, Inc.	Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258
Contact:	Eleanor Vivadou	Location:	VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3128-01	BPOW6-1-20250916	Water	VOCMS Group1	8260-Low	09/16/25		09/26/25	09/17/25
Q3128-04	BPOW6-4-20250916	Water	VOCMS Group1	8260-Low	09/16/25		09/26/25	09/17/25
Q3128-05	BPOW6-9-20250916	Water	VOCMS Group1	8260-Low	09/16/25		09/26/25	09/17/25
Q3128-06	BPOW6-3-20250916	Water	VOCMS Group1	8260-Low	09/16/25		09/26/25	09/17/25
Q3128-07	BPOW6-2-20250916	Water	VOCMS Group1	8260-Low	09/16/25		09/25/25	09/17/25
Q3128-08	BPOW6-10-20250916	Water	VOCMS Group1	8260-Low	09/16/25		09/25/25	09/17/25
Q3128-09	BPOW6-11-20250916	Water	VOCMS Group1	8260-Low	09/16/25		09/25/25	09/17/25
Q3128-10	BPOW6-8-20250916	Water	VOCMS Group1	8260-Low	09/16/25		09/25/25	09/17/25
Q3128-11	BPOW6-7-20250917	Water	VOCMS Group1	8260-Low	09/17/25		09/25/25	09/17/25
Q3128-12	BPOW6-6-20250917	Water	VOCMS Group1	8260-Low	09/17/25		09/25/25	09/17/25
Q3128-13	BPOW6-5-20250917	Water	VOCMS Group1	8260-Low	09/17/25		09/25/25	09/17/25
Q3128-14	TB07-20250912	Water			09/12/25			09/17/25

LAB CHRONICLE

VOCMS Group1 8260-Low 09/26/25

Hit Summary Sheet
SW-846

SDG No.: Q3128
Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:	BPOW6-1-20250916								
Q3128-01	BPOW6-1-2025091	Water	Acetone	1.90	J	1.50	3.80	5.00	ug/L
			Total Voc :	1.90					
			Total Concentration:	1.90					
Client ID:	BPOW6-9-20250916								
Q3128-05	BPOW6-9-2025091	Water	Acetone	2.00	J	1.50	3.80	5.00	ug/L
Q3128-05	BPOW6-9-2025091	Water	Trichloroethene	0.29	J	0.090	0.75	1.00	ug/L
			Total Voc :	2.29					
			Total Concentration:	2.29					
Client ID:	BPOW6-3-20250916								
Q3128-06	BPOW6-3-2025091	Water	Acetone	1.60	J	1.50	3.80	5.00	ug/L
			Total Voc :	1.60					
			Total Concentration:	1.60					
Client ID:	BPOW6-6-20250917								
Q3128-12	BPOW6-6-2025091	Water	m/p-Xylenes	0.33	J	0.24	1.00	2.00	ug/L
Q3128-12	BPOW6-6-2025091	Water	o-Xylene	0.25	J	0.12	0.50	1.00	ug/L
			Total Voc :	0.58					
			Total Concentration:	0.58					



SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-1-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047843.D	1	09/26/25 08:25	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	1.90	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-1-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047843.D	1	09/26/25 08:25	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.3		81 - 118		109%	SPK: 50
1868-53-7	Dibromofluoromethane	49.1		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	48.6		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.7		85 - 114		109%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	118000	5.543				
540-36-3	1,4-Difluorobenzene	250000	6.751				
3114-55-4	Chlorobenzene-d5	255000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	128000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-1-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047843.D	1	09/26/25 08:25	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-4-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047839.D	1	09/26/25 07:01	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-4-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047839.D	1	09/26/25 07:01	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.9		81 - 118		102%	SPK: 50
1868-53-7	Dibromofluoromethane	48.5		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	48.0		89 - 112		96%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.0		85 - 114		106%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	120000	5.544				
540-36-3	1,4-Difluorobenzene	253000	6.751				
3114-55-4	Chlorobenzene-d5	258000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	127000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-4-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047839.D	1	09/26/25 07:01	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-9-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-05		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047840.D	1	09/26/25 07:22	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	2.00	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.29	J	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-9-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-05		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047840.D	1	09/26/25 07:22	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.5		81 - 118		109%	SPK: 50
1868-53-7	Dibromofluoromethane	50.2		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	49.4		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	55.3		85 - 114		111%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	119000	5.538				
540-36-3	1,4-Difluorobenzene	251000	6.751				
3114-55-4	Chlorobenzene-d5	261000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	133000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-9-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-05		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047840.D	1	09/26/25 07:22	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-3-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-06		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047841.D	1	09/26/25 07:43	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	1.60	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-3-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-06		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047841.D	1	09/26/25 07:43	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.6		81 - 118		109%	SPK: 50
1868-53-7	Dibromofluoromethane	50.6		80 - 119		101%	SPK: 50
2037-26-5	Toluene-d8	49.5		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	56.3		85 - 114		113%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	133000	5.544				
540-36-3	1,4-Difluorobenzene	283000	6.751				
3114-55-4	Chlorobenzene-d5	293000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	148000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-3-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-06		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047841.D	1	09/26/25 07:43	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-2-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-07		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047810.D	1	09/25/25 14:35	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-2-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-07		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047810.D	1	09/25/25 14:35	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.3		81 - 118		105%	SPK: 50
1868-53-7	Dibromofluoromethane	48.9		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	49.0		89 - 112		98%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.0		85 - 114		106%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	119000	5.538				
540-36-3	1,4-Difluorobenzene	255000	6.745				
3114-55-4	Chlorobenzene-d5	260000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	131000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-2-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-07		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047810.D	1	09/25/25 14:35	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-10-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-08		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047811.D	1	09/25/25 14:56	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-10-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-08		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047811.D	1	09/25/25 14:56	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.4		81 - 118		103%	SPK: 50
1868-53-7	Dibromofluoromethane	48.7		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	48.4		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.0		85 - 114		108%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	124000	5.538				
540-36-3	1,4-Difluorobenzene	261000	6.751				
3114-55-4	Chlorobenzene-d5	264000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	132000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-10-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-08		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047811.D	1	09/25/25 14:56	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-11-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-09		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047812.D	1	09/25/25 15:17	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-11-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-09		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047812.D	1	09/25/25 15:17	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.2		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	49.4		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	49.1		89 - 112		98%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.4		85 - 114		109%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	116000	5.544				
540-36-3	1,4-Difluorobenzene	248000	6.751				
3114-55-4	Chlorobenzene-d5	257000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	128000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-11-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-09		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047812.D	1	09/25/25 15:17	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-8-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-10		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047813.D	1	09/25/25 15:38	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-8-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-10		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047813.D	1	09/25/25 15:38	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.0		81 - 118		104%	SPK: 50
1868-53-7	Dibromofluoromethane	49.9		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	48.7		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.7		85 - 114		107%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	119000	5.538				
540-36-3	1,4-Difluorobenzene	250000	6.745				
3114-55-4	Chlorobenzene-d5	258000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	128000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-8-20250916		SDG No.:	Q3128	
Lab Sample ID:	Q3128-10		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047813.D	1	09/25/25 15:38	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/17/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-7-20250917		SDG No.:	Q3128	
Lab Sample ID:	Q3128-11		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047814.D	1	09/25/25 15:59	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/17/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-7-20250917		SDG No.:	Q3128	
Lab Sample ID:	Q3128-11		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047814.D	1	09/25/25 15:59	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.4		81 - 118		101%	SPK: 50
1868-53-7	Dibromofluoromethane	49.5		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	49.4		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.8		85 - 114		108%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	120000	5.537				
540-36-3	1,4-Difluorobenzene	249000	6.751				
3114-55-4	Chlorobenzene-d5	255000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	128000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/17/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-7-20250917		SDG No.:	Q3128	
Lab Sample ID:	Q3128-11		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047814.D	1	09/25/25 15:59	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/17/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-6-20250917		SDG No.:	Q3128	
Lab Sample ID:	Q3128-12		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047815.D	1	09/25/25 16:21	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/17/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-6-20250917		SDG No.:	Q3128	
Lab Sample ID:	Q3128-12		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047815.D	1	09/25/25 16:21	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	0.33	J	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.25	J	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.9		81 - 118		110%	SPK: 50
1868-53-7	Dibromofluoromethane	50.2		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	49.6		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	56.1		85 - 114		112%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	128000	5.537				
540-36-3	1,4-Difluorobenzene	270000	6.745				
3114-55-4	Chlorobenzene-d5	279000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	144000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/17/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-6-20250917		SDG No.:	Q3128	
Lab Sample ID:	Q3128-12		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047815.D	1	09/25/25 16:21	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/17/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-5-20250917		SDG No.:	Q3128	
Lab Sample ID:	Q3128-13		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047816.D	1	09/25/25 16:42	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/17/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-5-20250917		SDG No.:	Q3128	
Lab Sample ID:	Q3128-13		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047816.D	1	09/25/25 16:42	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.1		81 - 118		104%	SPK: 50
1868-53-7	Dibromofluoromethane	48.4		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	47.3		89 - 112		95%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.0		85 - 114		104%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	122000	5.537				
540-36-3	1,4-Difluorobenzene	261000	6.745				
3114-55-4	Chlorobenzene-d5	264000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	132000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/17/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-5-20250917		SDG No.:	Q3128	
Lab Sample ID:	Q3128-13		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047816.D	1	09/25/25 16:42	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/12/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	TB07-20250912		SDG No.:	Q3128	
Lab Sample ID:	Q3128-14		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047842.D	1	09/26/25 08:04	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/12/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	TB07-20250912		SDG No.:	Q3128	
Lab Sample ID:	Q3128-14		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047842.D	1	09/26/25 08:04	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.3		81 - 118		103%	SPK: 50
1868-53-7	Dibromofluoromethane	48.8		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	48.6		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.1		85 - 114		108%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	132000	5.538				
540-36-3	1,4-Difluorobenzene	276000	6.751				
3114-55-4	Chlorobenzene-d5	283000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	141000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/12/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	TB07-20250912		SDG No.:	Q3128	
Lab Sample ID:	Q3128-14		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047842.D	1	09/26/25 08:04	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: **Q3128**

Client: **AECOM Technical Services, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3128-01	BPOW6-1-20250916	1,2-Dichloroethane-d4	50	54.3	109		81	118
		Dibromofluoromethane	50	49.1	98		80	119
		Toluene-d8	50	48.6	97		89	112
		4-Bromofluorobenzene	50	54.7	109		85	114
Q3128-02MS	BPOW6-1-20250916MS	1,2-Dichloroethane-d4	50	55.4	111		81	118
		Dibromofluoromethane	50	54.1	108		80	119
		Toluene-d8	50	51.7	103		89	112
		4-Bromofluorobenzene	50	52.9	106		85	114
Q3128-03MSD	BPOW6-1-20250916MSD	1,2-Dichloroethane-d4	50	49.1	98		81	118
		Dibromofluoromethane	50	47.9	96		80	119
		Toluene-d8	50	47.6	95		89	112
		4-Bromofluorobenzene	50	47.9	96		85	114
Q3128-04	BPOW6-4-20250916	1,2-Dichloroethane-d4	50	50.9	102		81	118
		Dibromofluoromethane	50	48.5	97		80	119
		Toluene-d8	50	48.0	96		89	112
		4-Bromofluorobenzene	50	53.0	106		85	114
Q3128-05	BPOW6-9-20250916	1,2-Dichloroethane-d4	50	54.5	109		81	118
		Dibromofluoromethane	50	50.2	100		80	119
		Toluene-d8	50	49.4	99		89	112
		4-Bromofluorobenzene	50	55.3	111		85	114
Q3128-06	BPOW6-3-20250916	1,2-Dichloroethane-d4	50	54.6	109		81	118
		Dibromofluoromethane	50	50.6	101		80	119
		Toluene-d8	50	49.5	99		89	112
		4-Bromofluorobenzene	50	56.3	113		85	114
Q3128-07	BPOW6-2-20250916	1,2-Dichloroethane-d4	50	52.3	105		81	118
		Dibromofluoromethane	50	48.9	98		80	119
		Toluene-d8	50	49.0	98		89	112
		4-Bromofluorobenzene	50	53.0	106		85	114
Q3128-08	BPOW6-10-20250916	1,2-Dichloroethane-d4	50	51.4	103		81	118
		Dibromofluoromethane	50	48.7	97		80	119
		Toluene-d8	50	48.4	97		89	112
		4-Bromofluorobenzene	50	54.0	108		85	114
Q3128-09	BPOW6-11-20250916	1,2-Dichloroethane-d4	50	53.2	106		81	118
		Dibromofluoromethane	50	49.4	99		80	119
		Toluene-d8	50	49.1	98		89	112
		4-Bromofluorobenzene	50	54.5	109		85	114
Q3128-10	BPOW6-8-20250916	1,2-Dichloroethane-d4	50	52.0	104		81	118
		Dibromofluoromethane	50	49.9	100		80	119
		Toluene-d8	50	48.7	97		89	112
		4-Bromofluorobenzene	50	53.7	107		85	114
Q3128-11	BPOW6-7-20250917	1,2-Dichloroethane-d4	50	50.4	101		81	118
		Dibromofluoromethane	50	49.5	99		80	119
		Toluene-d8	50	49.4	99		89	112
		4-Bromofluorobenzene	50	53.8	108		85	114
Q3128-12	BPOW6-6-20250917	1,2-Dichloroethane-d4	50	54.9	110		81	118
		Dibromofluoromethane	50	50.2	100		80	119
		Toluene-d8	50	49.6	99		89	112
		4-Bromofluorobenzene	50	56.1	112		85	114
Q3128-13	BPOW6-5-20250917	1,2-Dichloroethane-d4	50	52.1	104		81	118
		Dibromofluoromethane	50	48.4	97		80	119

Surrogate Summary

SDG No.: **Q3128**

Client: **AECOM Technical Services, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3128-13	BPOW6-5-20250917	Toluene-d8	50	47.3	95		89	112
		4-Bromofluorobenzene	50	52.0	104		85	114
Q3128-14	TB07-20250912	1,2-Dichloroethane-d4	50	51.3	103		81	118
		Dibromofluoromethane	50	48.8	98		80	119
		Toluene-d8	50	48.6	97		89	112
		4-Bromofluorobenzene	50	54.1	108		85	114
VX0925WBL01	VX0925WBL01	1,2-Dichloroethane-d4	50	51.1	102		81	118
		Dibromofluoromethane	50	48.4	97		80	119
		Toluene-d8	50	48.3	97		89	112
		4-Bromofluorobenzene	50	53.3	107		85	114
VX0925WBL02	VX0925WBL02	1,2-Dichloroethane-d4	50	54.4	109		81	118
		Dibromofluoromethane	50	48.8	98		80	119
		Toluene-d8	50	48.7	97		89	112
		4-Bromofluorobenzene	50	54.8	110		85	114
VX0925WBS01	VX0925WBS01	1,2-Dichloroethane-d4	50	53.3	107		81	118
		Dibromofluoromethane	50	52.0	104		80	119
		Toluene-d8	50	50.2	100		89	112
		4-Bromofluorobenzene	50	53.1	106		85	114
VX0925WBS02	VX0925WBS02	1,2-Dichloroethane-d4	50	53.5	107		81	118
		Dibromofluoromethane	50	54.5	109		80	119
		Toluene-d8	50	53.6	107		89	112
		4-Bromofluorobenzene	50	54.6	109		85	114

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3128

Analytical Method:

SW8260-Low

Client: AECOM Technical Services, Inc.

Datafile :

VX047844.D

		Sample				Rec	RPD	Limits			
Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID :	Q3128-02MS	Client Sample ID :	BPOW6-1-20250916MS								
Dichlorodifluoromethane	50	0	54.8	ug/L	110				32	152	
Chloromethane	50	0	55.6	ug/L	111				50	139	
Vinyl chloride	50	0	57.3	ug/L	115				58	137	
Bromomethane	50	0	49.6	ug/L	99				53	141	
Chloroethane	50	0	55.0	ug/L	110				60	138	
Trichlorofluoromethane	50	0	53.5	ug/L	107				65	141	
1,1,2-Trichlorotrifluoroethane	50	0	50.3	ug/L	101				70	136	
1,1-Dichloroethene	50	0	54.6	ug/L	109				71	131	
Acetone	250	1.90	240	ug/L	95				39	160	
Carbon disulfide	50	0	55.0	ug/L	110				64	133	
Methyl tert-butyl Ether	50	0	56.0	ug/L	112				71	124	
Methyl Acetate	50	0	51.9	ug/L	104				56	136	
Methylene Chloride	50	0	56.0	ug/L	112				74	124	
trans-1,2-Dichloroethene	50	0	56.0	ug/L	112				75	124	
1,1-Dichloroethane	50	0	55.9	ug/L	112				77	125	
Cyclohexane	50	0	48.2	ug/L	96				71	130	
2-Butanone	250	0	290	ug/L	116				56	143	
Carbon Tetrachloride	50	0	49.9	ug/L	100				72	136	
cis-1,2-Dichloroethene	50	0	56.0	ug/L	112				78	123	
Bromochloromethane	50	0	55.8	ug/L	112				78	123	
Chloroform	50	0	56.8	ug/L	114				79	124	
1,1,1-Trichloroethane	50	0	53.8	ug/L	108				74	131	
Methylcyclohexane	50	0	43.4	ug/L	87				72	132	
Benzene	50	0	53.1	ug/L	106				79	120	
1,2-Dichloroethane	50	0	52.0	ug/L	104				73	128	
Trichloroethene	50	0	52.0	ug/L	104				79	123	
1,2-Dichloropropane	50	0	54.2	ug/L	108				78	122	
Bromodichloromethane	50	0	54.2	ug/L	108				79	125	
4-Methyl-2-Pentanone	250	0	290	ug/L	116				67	130	
Toluene	50	0	52.9	ug/L	106				80	121	
t-1,3-Dichloropropene	50	0	47.1	ug/L	94				73	127	
cis-1,3-Dichloropropene	50	0	46.3	ug/L	93				75	124	
1,1,2-Trichloroethane	50	0	55.5	ug/L	111				80	119	
2-Hexanone	250	0	280	ug/L	112				57	139	
Dibromochloromethane	50	0	55.8	ug/L	112				74	126	
1,2-Dibromoethane	50	0	56.3	ug/L	113				77	121	
Tetrachloroethene	50	0	46.9	ug/L	94				74	129	
Chlorobenzene	50	0	50.3	ug/L	101				82	118	
Ethyl Benzene	50	0	49.2	ug/L	98				79	121	
m/p-Xylenes	100	0	99.6	ug/L	100				80	121	
o-Xylene	50	0	50.2	ug/L	100				78	122	
Styrene	50	0	50.0	ug/L	100				78	123	
Bromoform	50	0	52.0	ug/L	104				66	130	
Isopropylbenzene	50	0	46.8	ug/L	94				72	131	
1,1,2,2-Tetrachloroethane	50	0	52.0	ug/L	104				71	121	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3128

Analytical Method: SW8260-Low

Client: AECOM Technical Services, Inc.

Datafile : VX047844.D

Parameter	Spike	Sample	Result	Units	Rec			RPD	Limits		
		Result			Rec	Qual	RPD	Qual	Low	High	RPD
1,3-Dichlorobenzene	50	0	46.8	ug/L	94				80	119	
1,4-Dichlorobenzene	50	0	46.3	ug/L	93				79	118	
1,2-Dichlorobenzene	50	0	48.4	ug/L	97				80	119	
1,2-Dibromo-3-Chloropropane	50	0	49.9	ug/L	100				62	128	
1,2,4-Trichlorobenzene	50	0	45.2	ug/L	90				69	130	
1,2,3-Trichlorobenzene	50	0	47.3	ug/L	95				69	129	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3128

Analytical Method: SW8260-Low

Client: AECOM Technical Services, Inc.

Datafile : VX047845.D

Parameter	Spike	Sample	Result	Units	Rec		RPD		Limits		RPD
		Result			Qual	RPD	Qual	Low	High		
Lab Sample ID :	Q3128-03MSD	Client Sample ID :	BPOW6-1-20250916MSD								
Dichlorodifluoromethane	50	0	51.3	ug/L	103		7		32	152	20
Chloromethane	50	0	50.3	ug/L	101		10		50	139	20
Vinyl chloride	50	0	51.5	ug/L	103		11		58	137	20
Bromomethane	50	0	48.2	ug/L	96		3		53	141	20
Chloroethane	50	0	50.3	ug/L	101		9		60	138	20
Trichlorofluoromethane	50	0	49.2	ug/L	98		8		65	141	20
1,1,2-Trichlorotrifluoroethane	50	0	47.7	ug/L	95		5		70	136	20
1,1-Dichloroethene	50	0	50.6	ug/L	101		8		71	131	20
Acetone	250	1.90	230	ug/L	91		4		39	160	20
Carbon disulfide	50	0	50.2	ug/L	100		9		64	133	20
Methyl tert-butyl Ether	50	0	52.3	ug/L	105		7		71	124	20
Methyl Acetate	50	0	47.4	ug/L	95		9		56	136	20
Methylene Chloride	50	0	51.9	ug/L	104		8		74	124	20
trans-1,2-Dichloroethene	50	0	51.5	ug/L	103		8		75	124	20
1,1-Dichloroethane	50	0	51.7	ug/L	103		8		77	125	20
Cyclohexane	50	0	45.8	ug/L	92		5		71	130	20
2-Butanone	250	0	280	ug/L	112		4		56	143	20
Carbon Tetrachloride	50	0	45.9	ug/L	92		8		72	136	20
cis-1,2-Dichloroethene	50	0	52.3	ug/L	105		7		78	123	20
Bromochloromethane	50	0	51.9	ug/L	104		7		78	123	20
Chloroform	50	0	52.8	ug/L	106		7		79	124	20
1,1,1-Trichloroethane	50	0	50.4	ug/L	101		7		74	131	20
Methylcyclohexane	50	0	42.0	ug/L	84		3		72	132	20
Benzene	50	0	50.3	ug/L	101		5		79	120	20
1,2-Dichloroethane	50	0	49.7	ug/L	99		5		73	128	20
Trichloroethene	50	0	50.4	ug/L	101		3		79	123	20
1,2-Dichloropropane	50	0	51.3	ug/L	103		5		78	122	20
Bromodichloromethane	50	0	51.1	ug/L	102		6		79	125	20
4-Methyl-2-Pentanone	250	0	280	ug/L	112		4		67	130	20
Toluene	50	0	50.5	ug/L	101		5		80	121	20
t-1,3-Dichloropropene	50	0	46.0	ug/L	92		2		73	127	20
cis-1,3-Dichloropropene	50	0	45.2	ug/L	90		2		75	124	20
1,1,2-Trichloroethane	50	0	54.1	ug/L	108		3		80	119	20
2-Hexanone	250	0	280	ug/L	112		0		57	139	20
Dibromochloromethane	50	0	53.7	ug/L	107		4		74	126	20
1,2-Dibromoethane	50	0	55.4	ug/L	111		2		77	121	20
Tetrachloroethene	50	0	45.5	ug/L	91		3		74	129	20
Chlorobenzene	50	0	48.5	ug/L	97		4		82	118	20
Ethyl Benzene	50	0	47.7	ug/L	95		3		79	121	20
m/p-Xylenes	100	0	97.2	ug/L	97		2		80	121	20
o-Xylene	50	0	49.0	ug/L	98		2		78	122	20
Styrene	50	0	48.5	ug/L	97		3		78	123	20
Bromoform	50	0	51.0	ug/L	102		2		66	130	20
Isopropylbenzene	50	0	45.6	ug/L	91		3		72	131	20
1,1,2,2-Tetrachloroethane	50	0	51.6	ug/L	103		1		71	121	20

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3128

Analytical Method: SW8260-Low

Client: AECOM Technical Services, Inc.

Datafile : VX047845.D

Parameter	Spike	Sample Result	Result	Units	Rec			RPD		Limits	
					Rec	Qual	RPD	Qual	Low	High	RPD
1,3-Dichlorobenzene	50	0	47.2	ug/L	94		1		80	119	20
1,4-Dichlorobenzene	50	0	46.1	ug/L	92		0		79	118	20
1,2-Dichlorobenzene	50	0	48.3	ug/L	97		0		80	119	20
1,2-Dibromo-3-Chloropropane	50	0	50.6	ug/L	101		1		62	128	20
1,2,4-Trichlorobenzene	50	0	46.3	ug/L	93		2		69	130	20
1,2,3-Trichlorobenzene	50	0	47.7	ug/L	95		1		69	129	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3128</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>AECOM Technical Services, Inc.</u>	Datafile :	<u>VX047798.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0925WBS01	Dichlorodifluoromethane	20	22.1	ug/L	111			32	152	
	Chloromethane	20	20.6	ug/L	103			50	139	
	Vinyl chloride	20	20.6	ug/L	103			58	137	
	Bromomethane	20	20.8	ug/L	104			53	141	
	Chloroethane	20	20.2	ug/L	101			60	138	
	Trichlorofluoromethane	20	19.8	ug/L	99			65	141	
	1,1,2-Trichlorotrifluoroethane	20	20.3	ug/L	102			70	136	
	1,1-Dichloroethene	20	19.3	ug/L	97			71	131	
	Acetone	100	96.4	ug/L	96			39	160	
	Carbon disulfide	20	19.6	ug/L	98			64	133	
	Methyl tert-butyl Ether	20	20.0	ug/L	100			71	124	
	Methyl Acetate	20	22.6	ug/L	113			56	136	
	Methylene Chloride	20	20.6	ug/L	103			74	124	
	trans-1,2-Dichloroethene	20	20.2	ug/L	101			75	124	
	1,1-Dichloroethane	20	20.4	ug/L	102			77	125	
	Cyclohexane	20	19.2	ug/L	96			71	130	
	2-Butanone	100	110	ug/L	110			56	143	
	Carbon Tetrachloride	20	18.2	ug/L	91			72	136	
	cis-1,2-Dichloroethene	20	20.3	ug/L	102			78	123	
	Bromochloromethane	20	20.0	ug/L	100			78	123	
	Chloroform	20	20.7	ug/L	104			79	124	
	1,1,1-Trichloroethane	20	19.8	ug/L	99			74	131	
	Methylcyclohexane	20	17.8	ug/L	89			72	132	
	Benzene	20	19.9	ug/L	100			79	120	
	1,2-Dichloroethane	20	19.5	ug/L	98			73	128	
	Trichloroethene	20	19.7	ug/L	99			79	123	
	1,2-Dichloropropane	20	20.1	ug/L	101			78	122	
	Bromodichloromethane	20	19.5	ug/L	98			79	125	
	4-Methyl-2-Pentanone	100	110	ug/L	110			67	130	
	Toluene	20	20.0	ug/L	100			80	121	
	t-1,3-Dichloropropene	20	18.9	ug/L	95			73	127	
	cis-1,3-Dichloropropene	20	19.3	ug/L	97			75	124	
	1,1,2-Trichloroethane	20	21.0	ug/L	105			80	119	
	2-Hexanone	100	110	ug/L	110			57	139	
	Dibromochloromethane	20	20.5	ug/L	103			74	126	
	1,2-Dibromoethane	20	20.8	ug/L	104			77	121	
	Tetrachloroethene	20	18.7	ug/L	94			74	129	
	Chlorobenzene	20	19.0	ug/L	95			82	118	
	Ethyl Benzene	20	18.1	ug/L	91			79	121	
	m/p-Xylenes	40	37.6	ug/L	94			80	121	
	o-Xylene	20	18.5	ug/L	93			78	122	
	Styrene	20	18.9	ug/L	95			78	123	
	Bromoform	20	19.6	ug/L	98			66	130	
	Isopropylbenzene	20	16.8	ug/L	84			72	131	
	1,1,2,2-Tetrachloroethane	20	18.9	ug/L	95			71	121	
	1,3-Dichlorobenzene	20	17.5	ug/L	88			80	119	
	1,4-Dichlorobenzene	20	17.8	ug/L	89			79	118	
	1,2-Dichlorobenzene	20	17.8	ug/L	89			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3128</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>AECOM Technical Services, Inc.</u>	Datafile :	<u>VX047798.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0925WBS01	1,2-Dibromo-3-Chloropropane	20	17.6	ug/L	88			62	128	
	1,2,4-Trichlorobenzene	20	16.2	ug/L	81			69	130	
	1,2,3-Trichlorobenzene	20	16.9	ug/L	85			69	129	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3128</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>AECOM Technical Services, Inc.</u>	Datafile :	<u>VX047827.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0925WBS02	Dichlorodifluoromethane	20	21.6	ug/L	108			32	152	
	Chloromethane	20	20.8	ug/L	104			50	139	
	Vinyl chloride	20	20.6	ug/L	103			58	137	
	Bromomethane	20	20.2	ug/L	101			53	141	
	Chloroethane	20	20.6	ug/L	103			60	138	
	Trichlorofluoromethane	20	20.1	ug/L	101			65	141	
	1,1,2-Trichlorotrifluoroethane	20	19.7	ug/L	99			70	136	
	1,1-Dichloroethene	20	20.6	ug/L	103			71	131	
	Acetone	100	93.0	ug/L	93			39	160	
	Carbon disulfide	20	19.8	ug/L	99			64	133	
	Methyl tert-butyl Ether	20	20.9	ug/L	104			71	124	
	Methyl Acetate	20	24.7	ug/L	124			56	136	
	Methylene Chloride	20	21.6	ug/L	108			74	124	
	trans-1,2-Dichloroethene	20	20.8	ug/L	104			75	124	
	1,1-Dichloroethane	20	21.0	ug/L	105			77	125	
	Cyclohexane	20	19.3	ug/L	97			71	130	
	2-Butanone	100	110	ug/L	110			56	143	
	Carbon Tetrachloride	20	19.1	ug/L	96			72	136	
	cis-1,2-Dichloroethene	20	21.1	ug/L	106			78	123	
	Bromochloromethane	20	20.3	ug/L	102			78	123	
	Chloroform	20	21.6	ug/L	108			79	124	
	1,1,1-Trichloroethane	20	20.5	ug/L	103			74	131	
	Methylcyclohexane	20	17.7	ug/L	89			72	132	
	Benzene	20	20.5	ug/L	103			79	120	
	1,2-Dichloroethane	20	19.9	ug/L	100			73	128	
	Trichloroethene	20	20.1	ug/L	101			79	123	
	1,2-Dichloropropane	20	21.0	ug/L	105			78	122	
	Bromodichloromethane	20	20.6	ug/L	103			79	125	
	4-Methyl-2-Pentanone	100	120	ug/L	120			67	130	
	Toluene	20	20.3	ug/L	102			80	121	
	t-1,3-Dichloropropene	20	18.5	ug/L	93			73	127	
	cis-1,3-Dichloropropene	20	18.7	ug/L	94			75	124	
	1,1,2-Trichloroethane	20	21.6	ug/L	108			80	119	
	2-Hexanone	100	110	ug/L	110			57	139	
	Dibromochloromethane	20	21.5	ug/L	108			74	126	
	1,2-Dibromoethane	20	22.5	ug/L	113			77	121	
	Tetrachloroethene	20	20.2	ug/L	101			74	129	
	Chlorobenzene	20	19.8	ug/L	99			82	118	
	Ethyl Benzene	20	19.0	ug/L	95			79	121	
	m/p-Xylenes	40	39.4	ug/L	99			80	121	
	o-Xylene	20	19.5	ug/L	98			78	122	
	Styrene	20	19.7	ug/L	99			78	123	
	Bromoform	20	20.8	ug/L	104			66	130	
	Isopropylbenzene	20	18.5	ug/L	93			72	131	
	1,1,2,2-Tetrachloroethane	20	21.2	ug/L	106			71	121	
	1,3-Dichlorobenzene	20	19.5	ug/L	98			80	119	
	1,4-Dichlorobenzene	20	19.0	ug/L	95			79	118	
	1,2-Dichlorobenzene	20	19.7	ug/L	99			80	119	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q3128	Analytical Method:	SW8260-Low
Client:	AECOM Technical Services, Inc.	Datafile :	VX047827.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0925WBS02	1,2-Dibromo-3-Chloropropane	20	19.5	ug/L	98			62	128	
	1,2,4-Trichlorobenzene	20	18.1	ug/L	91			69	130	
	1,2,3-Trichlorobenzene	20	19.0	ug/L	95			69	129	

VOLATILE METHOD BLANK SUMMARY

Client ID

VX0925WBL01

Lab Name: Alliance

Contract: AECO15

Lab Code: ACE

SDG NO.: Q3128

Lab File ID: VX047797.D

Lab Sample ID: VX0925WBL01

Date Analyzed: 09/25/2025

Time Analyzed: 09:46

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0925WBS01	VX0925WBS01	VX047798.D	09/25/2025
BPOW6-2-20250916	Q3128-07	VX047810.D	09/25/2025
BPOW6-10-20250916	Q3128-08	VX047811.D	09/25/2025
BPOW6-11-20250916	Q3128-09	VX047812.D	09/25/2025
BPOW6-8-20250916	Q3128-10	VX047813.D	09/25/2025
BPOW6-7-20250917	Q3128-11	VX047814.D	09/25/2025
BPOW6-6-20250917	Q3128-12	VX047815.D	09/25/2025
BPOW6-5-20250917	Q3128-13	VX047816.D	09/25/2025

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

Client ID

VX0925WBL02

Lab Name: Alliance

Contract: AECO15

Lab Code: ACE

SDG NO.: Q3128

Lab File ID: VX047826.D

Lab Sample ID: VX0925WBL02

Date Analyzed: 09/26/2025

Time Analyzed: 02:10

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0925WBS02	VX0925WBS02	VX047827.D	09/26/2025
BPOW6-4-20250916	Q3128-04	VX047839.D	09/26/2025
BPOW6-9-20250916	Q3128-05	VX047840.D	09/26/2025
BPOW6-3-20250916	Q3128-06	VX047841.D	09/26/2025
TB07-20250912	Q3128-14	VX047842.D	09/26/2025
BPOW6-1-20250916	Q3128-01	VX047843.D	09/26/2025
BPOW6-1-20250916MS	Q3128-02MS	VX047844.D	09/26/2025
BPOW6-1-20250916MSD	Q3128-03MSD	VX047845.D	09/26/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: AECO15

Lab Code: ACE

SDG NO.: Q3128

Lab File ID: VX047584.D

BFB Injection Date: 09/16/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:56

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20
75	30.0 - 60.0% of mass 95	54
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	67.3
175	5.0 - 9.0% of mass 174	5.4 (8.1) 1
176	95.0 - 101.0% of mass 174	66.3 (98.6) 1
177	5.0 - 9.0% of mass 176	4.3 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX047585.D	09/16/2025	09:23
VSTDIC005	VSTDIC005	VX047586.D	09/16/2025	10:16
VSTDIC020	VSTDIC020	VX047587.D	09/16/2025	10:37
VSTDIC050	VSTDIC050	VX047588.D	09/16/2025	10:58
VSTDIC100	VSTDIC100	VX047589.D	09/16/2025	11:40
VSTDIC150	VSTDIC150	VX047590.D	09/16/2025	12:01

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: AECO15

Lab Code: ACE

SDG NO.: Q3128

Lab File ID: VX047794.D

BFB Injection Date: 09/25/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:24

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.1
75	30.0 - 60.0% of mass 95	52
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	69.9
175	5.0 - 9.0% of mass 174	5.4 (7.8) 1
176	95.0 - 101.0% of mass 174	67 (95.7) 1
177	5.0 - 9.0% of mass 176	4.6 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX047795.D	09/25/2025	08:56
VX0925WBL01	VX0925WBL01	VX047797.D	09/25/2025	09:46
VX0925WBS01	VX0925WBS01	VX047798.D	09/25/2025	10:13
BPOW6-2-20250916	Q3128-07	VX047810.D	09/25/2025	14:35
BPOW6-10-20250916	Q3128-08	VX047811.D	09/25/2025	14:56
BPOW6-11-20250916	Q3128-09	VX047812.D	09/25/2025	15:17
BPOW6-8-20250916	Q3128-10	VX047813.D	09/25/2025	15:38
BPOW6-7-20250917	Q3128-11	VX047814.D	09/25/2025	15:59
BPOW6-6-20250917	Q3128-12	VX047815.D	09/25/2025	16:21
BPOW6-5-20250917	Q3128-13	VX047816.D	09/25/2025	16:42
VSTDCCC050EC	VSTDCCC050	VX047823.D	09/25/2025	19:09

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: VX047824.D
Instrument ID: MSVOA_X
GC Column: DB-624UI ID: 0.18 (mm)

Contract: AECO15
SDG NO.: Q3128
BFB Injection Date: 09/26/2025
BFB Injection Time: 00:46
Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.8
75	30.0 - 60.0% of mass 95	51.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.4 (0.5) 1
174	50.0 - 100.0% of mass 95	76.6
175	5.0 - 9.0% of mass 174	5.9 (7.7) 1
176	95.0 - 101.0% of mass 174	73.5 (95.9) 1
177	5.0 - 9.0% of mass 176	4.9 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX047825.D	09/26/2025	01:28
VX0925WBL02	VX0925WBL02	VX047826.D	09/26/2025	02:10
VX0925WBS02	VX0925WBS02	VX047827.D	09/26/2025	02:31
BPOW6-4-20250916	Q3128-04	VX047839.D	09/26/2025	07:01
BPOW6-9-20250916	Q3128-05	VX047840.D	09/26/2025	07:22
BPOW6-3-20250916	Q3128-06	VX047841.D	09/26/2025	07:43
TB07-20250912	Q3128-14	VX047842.D	09/26/2025	08:04
BPOW6-1-20250916	Q3128-01	VX047843.D	09/26/2025	08:25
BPOW6-1-20250916MS	Q3128-02MS	VX047844.D	09/26/2025	08:46
BPOW6-1-20250916MSD	Q3128-03MSD	VX047845.D	09/26/2025	09:07
VSTDCCC050EC	VSTDCCC050	VX047846.D	09/26/2025	09:28

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15
Lab Code: ACE SDG NO.: Q3128
Lab File ID: VX047795.D Date Analyzed: 09/25/2025
Instrument ID: MSVOA_X Time Analyzed: 08:56
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	136107	5.53	245940	6.75	220262	10.04
UPPER LIMIT	272214	6.032	491880	7.245	440524	10.537
LOWER LIMIT	68053.5	5.032	122970	6.245	110131	9.537
EPA SAMPLE NO.						
BPOW6-2-20250916	119164	5.54	254523	6.75	260449	10.04
BPOW6-10-20250916	124352	5.54	260863	6.75	264166	10.04
BPOW6-11-20250916	116305	5.54	248243	6.75	257163	10.04
BPOW6-8-20250916	119014	5.54	249986	6.75	257664	10.04
BPOW6-7-20250917	119857	5.54	249353	6.75	255198	10.04
BPOW6-6-20250917	127711	5.54	269892	6.75	279053	10.04
BPOW6-5-20250917	121834	5.54	260923	6.75	263550	10.04
VX0925WBL01	117291	5.54	245721	6.75	251657	10.04
VX0925WBS01	146604	5.54	268605	6.75	252515	10.04

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3128</u>
Lab File ID:	<u>VX047795.D</u>	Date Analyzed:	<u>09/25/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>08:56</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	107538	12.00€				
UPPER LIMIT	215076	12.50€				
LOWER LIMIT	53769	11.50€				
EPA SAMPLE NO.						
BPOW6-2-20250916	130810	12.01				
BPOW6-10-20250916	132252	12.01				
BPOW6-11-20250916	128330	12.01				
BPOW6-8-20250916	128316	12.01				
BPOW6-7-20250917	128055	12.01				
BPOW6-6-20250917	143881	12.01				
BPOW6-5-20250917	131658	12.01				
VX0925WBL01	126490	12.01				
VX0925WBS01	131521	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15
Lab Code: ACE SDG NO.: Q3128
Lab File ID: VX047825.D Date Analyzed: 09/26/2025
Instrument ID: MSVOA_X Time Analyzed: 01:28
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	136987	5.54	247384	6.75	222999	10.04
UPPER LIMIT	273974	6.038	494768	7.245	445998	10.537
LOWER LIMIT	68493.5	5.038	123692	6.245	111500	9.537
EPA SAMPLE NO.						
BPOW6-1-20250916	118010	5.54	250283	6.75	255095	10.04
BPOW6-1-20250916MS	132069	5.54	246705	6.75	231070	10.04
BPOW6-1-20250916MSD	130228	5.54	242264	6.75	226841	10.04
BPOW6-4-20250916	120108	5.54	253048	6.75	258299	10.04
BPOW6-9-20250916	118685	5.54	250694	6.75	261456	10.04
BPOW6-3-20250916	132775	5.54	283230	6.75	293419	10.04
TB07-20250912	132283	5.54	276404	6.75	283292	10.04
VX0925WBL02	121983	5.54	259174	6.75	269510	10.04
VX0925WBS02	143765	5.54	265116	6.75	243871	10.04

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15
 Lab Code: ACE SDG NO.: Q3128
 Lab File ID: VX047825.D Date Analyzed: 09/26/2025
 Instrument ID: MSVOA_X Time Analyzed: 01:28
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	110381	12.00€				
UPPER LIMIT	220762	12.50€				
LOWER LIMIT	55190.5	11.50€				
EPA SAMPLE NO.						
BPOW6-1-20250916	127833	12.01				
BPOW6-1-20250916MS	116018	12.01				
BPOW6-1-20250916MSD	112665	12.01				
BPOW6-4-20250916	127208	12.01				
BPOW6-9-20250916	133160	12.01				
BPOW6-3-20250916	148155	12.01				
TB07-20250912	140911	12.01				
VX0925WBL02	138899	12.01				
VX0925WBS02	122300	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0925WBL01		SDG No.:	Q3128
Lab Sample ID:	VX0925WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047797.D	1	09/25/25 09:46	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0925WBL01		SDG No.:	Q3128
Lab Sample ID:	VX0925WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047797.D	1	09/25/25 09:46	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.1		81 - 118		102%	SPK: 50
1868-53-7	Dibromofluoromethane	48.4		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	48.3		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.3		85 - 114		107%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	117000	5.538				
540-36-3	1,4-Difluorobenzene	246000	6.745				
3114-55-4	Chlorobenzene-d5	252000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	126000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0925WBL01		SDG No.:	Q3128
Lab Sample ID:	VX0925WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047797.D	1	09/25/25 09:46	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0925WBL02		SDG No.:	Q3128
Lab Sample ID:	VX0925WBL02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047826.D	1	09/26/25 02:10	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0925WBL02		SDG No.:	Q3128
Lab Sample ID:	VX0925WBL02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047826.D	1	09/26/25 02:10	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.3		81 - 118		109%	SPK: 50
1868-53-7	Dibromofluoromethane	48.8		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	48.7		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.8		85 - 114		110%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	122000	5.538				
540-36-3	1,4-Difluorobenzene	259000	6.745				
3114-55-4	Chlorobenzene-d5	270000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	139000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0925WBL02		SDG No.:	Q3128
Lab Sample ID:	VX0925WBL02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047826.D	1	09/26/25 02:10	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0925WBS01		SDG No.:	Q3128
Lab Sample ID:	VX0925WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047798.D	1	09/25/25 10:13	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	22.1		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	20.6		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	20.6		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	20.8		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	20.2		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.8		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	20.3		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.3		0.23	0.75	1.00	ug/L
67-64-1	Acetone	96.4		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	19.6		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	20.0		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	22.6		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	20.6		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	20.2		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	20.4		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	19.2		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	110		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.2		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	20.3		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	20.0		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	20.7		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.8		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	17.8		0.16	0.50	1.00	ug/L
71-43-2	Benzene	19.9		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.5		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.7		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.1		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	19.5		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.68	2.50	5.00	ug/L
108-88-3	Toluene	20.0		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0925WBS01		SDG No.:	Q3128
Lab Sample ID:	VX0925WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047798.D	1	09/25/25 10:13	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	18.9		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.3		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.0		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	20.5		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.8		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	18.7		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.0		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	18.1		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	37.6		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	18.5		0.12	0.50	1.00	ug/L
100-42-5	Styrene	18.9		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	19.6		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	16.8		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	18.9		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	17.5		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	17.8		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	17.8		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	17.6		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	16.2		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	16.9		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.3		81 - 118		107%	SPK: 50
1868-53-7	Dibromofluoromethane	52.0		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	50.2		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.1		85 - 114		106%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	147000	5.538				
540-36-3	1,4-Difluorobenzene	269000	6.745				
3114-55-4	Chlorobenzene-d5	253000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	132000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0925WBS01		SDG No.:	Q3128
Lab Sample ID:	VX0925WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047798.D	1	09/25/25 10:13	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0925WBS02		SDG No.:	Q3128
Lab Sample ID:	VX0925WBS02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047827.D	1	09/26/25 02:31	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	21.6		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	20.8		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	20.6		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	20.2		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	20.6		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	20.1		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.7		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	20.6		0.23	0.75	1.00	ug/L
67-64-1	Acetone	93.0		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	19.8		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	20.9		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	24.7		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	21.6		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	20.8		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	21.0		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	19.3		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	110		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.1		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	21.1		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	20.3		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	21.6		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	20.5		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	17.7		0.16	0.50	1.00	ug/L
71-43-2	Benzene	20.5		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.9		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	20.1		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	21.0		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	20.6		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	120		0.68	2.50	5.00	ug/L
108-88-3	Toluene	20.3		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0925WBS02		SDG No.:	Q3128
Lab Sample ID:	VX0925WBS02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047827.D	1	09/26/25 02:31	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	18.5		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	18.7		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.6		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	21.5		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	22.5		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	20.2		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.8		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.0		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	39.4		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	19.5		0.12	0.50	1.00	ug/L
100-42-5	Styrene	19.7		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	20.8		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	18.5		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	21.2		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.5		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.0		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.7		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	19.5		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	18.1		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	19.0		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.5		81 - 118		107%	SPK: 50
1868-53-7	Dibromofluoromethane	54.5		80 - 119		109%	SPK: 50
2037-26-5	Toluene-d8	53.6		89 - 112		107%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.6		85 - 114		109%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	144000	5.538				
540-36-3	1,4-Difluorobenzene	265000	6.745				
3114-55-4	Chlorobenzene-d5	244000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	122000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0925WBS02		SDG No.:	Q3128
Lab Sample ID:	VX0925WBS02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047827.D	1	09/26/25 02:31	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-1-20250916MS		SDG No.:	Q3128	
Lab Sample ID:	Q3128-02MS		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047844.D	1	09/26/25 08:46	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	54.8		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	55.6		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	57.3		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	49.6		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	55.0		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	53.5		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	50.3		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	54.6		0.23	0.75	1.00	ug/L
67-64-1	Acetone	240		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	55.0		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	56.0		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	51.9		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	56.0		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	56.0		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	55.9		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	48.2		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	290		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	49.9		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	56.0		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	55.8		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	56.8		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	53.8		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	43.4		0.16	0.50	1.00	ug/L
71-43-2	Benzene	53.1		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	52.0		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	52.0		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	54.2		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	54.2		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	290		0.68	2.50	5.00	ug/L
108-88-3	Toluene	52.9		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-1-20250916MS		SDG No.:	Q3128	
Lab Sample ID:	Q3128-02MS		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047844.D	1	09/26/25 08:46	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	47.1		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	46.3		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	55.5		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	280		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	55.8		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	56.3		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	46.9		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	50.3		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	49.2		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	99.6		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	50.2		0.12	0.50	1.00	ug/L
100-42-5	Styrene	50.0		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	52.0		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	46.8		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	52.0		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	46.8		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	46.3		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	48.4		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	49.9		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	45.2		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	47.3		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.4		81 - 118		111%	SPK: 50
1868-53-7	Dibromofluoromethane	54.1		80 - 119		108%	SPK: 50
2037-26-5	Toluene-d8	51.7		89 - 112		103%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.9		85 - 114		106%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	132000	5.544				
540-36-3	1,4-Difluorobenzene	247000	6.751				
3114-55-4	Chlorobenzene-d5	231000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	116000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-1-20250916MS		SDG No.:	Q3128	
Lab Sample ID:	Q3128-02MS		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047844.D	1	09/26/25 08:46	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-1-20250916MSD		SDG No.:	Q3128	
Lab Sample ID:	Q3128-03MSD		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047845.D	1	09/26/25 09:07	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	51.3		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	50.3		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	51.5		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	48.2		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	50.3		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	49.2		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	47.7		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	50.6		0.23	0.75	1.00	ug/L
67-64-1	Acetone	230		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	50.2		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	52.3		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	47.4		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	51.9		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	51.5		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	51.7		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	45.8		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	280		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	45.9		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	52.3		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	51.9		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	52.8		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	50.4		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	42.0		0.16	0.50	1.00	ug/L
71-43-2	Benzene	50.3		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	49.7		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	50.4		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	51.3		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	51.1		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	280		0.68	2.50	5.00	ug/L
108-88-3	Toluene	50.5		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-1-20250916MSD		SDG No.:	Q3128	
Lab Sample ID:	Q3128-03MSD		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047845.D	1	09/26/25 09:07	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	46.0		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	45.2		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	54.1		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	280		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	53.7		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	55.4		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	45.5		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	48.5		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	47.7		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	97.2		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	49.0		0.12	0.50	1.00	ug/L
100-42-5	Styrene	48.5		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	51.0		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	45.6		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	51.6		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	47.2		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	46.1		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	48.3		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	50.6		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	46.3		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	47.7		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	49.1		81 - 118		98%	SPK: 50
1868-53-7	Dibromofluoromethane	47.9		80 - 119		96%	SPK: 50
2037-26-5	Toluene-d8	47.6		89 - 112		95%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.9		85 - 114		96%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	130000	5.538				
540-36-3	1,4-Difluorobenzene	242000	6.745				
3114-55-4	Chlorobenzene-d5	227000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	113000	12.006				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	09/16/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	09/17/25	
Client Sample ID:	BPOW6-1-20250916MSD		SDG No.:	Q3128	
Lab Sample ID:	Q3128-03MSD		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047845.D	1	09/26/25 09:07	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3128</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Calibration Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

LAB FILE ID:		RRF001 = VX047585.D		RRF005 = VX047586.D		RRF020 = VX047587.D		
		RRF050 = VX047588.D		RRF100 = VX047589.D		RRF150 = VX047590.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.447	0.575	0.582	0.501	0.510	0.492	0.518	10
Chloromethane	0.610	0.749	0.739	0.661	0.677	0.647	0.680	7.9
Vinyl Chloride	0.572	0.713	0.716	0.661	0.677	0.658	0.666	7.9
Bromomethane		0.493	0.451	0.424	0.423	0.321	0.422	15
Chloroethane	0.433	0.450	0.469	0.443	0.450	0.426	0.445	3.4
Trichlorofluoromethane	0.865	1.033	1.038	0.995	1.018	0.986	0.989	6.5
1,1,2-Trichlorotrifluoroethane	0.519	0.591	0.603	0.583	0.590	0.583	0.578	5.2
1,1-Dichloroethene	0.526	0.612	0.618	0.604	0.609	0.596	0.594	5.7
Acetone	0.343	0.416	0.400	0.317	0.302	0.299	0.346	14.7
Carbon Disulfide	1.555	1.735	1.758	1.566	1.589	1.555	1.626	5.8
Methyl tert-butyl Ether	1.884	2.238	2.221	2.247	2.212	2.209	2.169	6.5
Methyl Acetate	0.612	0.673	0.714	0.876	0.865	0.881	0.770	15.4
Methylene Chloride	0.718	0.772	0.754	0.730	0.716	0.705	0.732	3.5
trans-1,2-Dichloroethene	0.543	0.703	0.665	0.649	0.643	0.635	0.640	8.3
1,1-Dichloroethane	1.090	1.325	1.299	1.297	1.287	1.278	1.263	6.8
Cyclohexane		1.161	1.092	1.030	0.984	0.993	1.052	7.1
2-Butanone	0.370	0.470	0.473	0.439	0.414	0.424	0.432	9
Carbon Tetrachloride	0.482	0.574	0.555	0.539	0.526	0.518	0.532	6
cis-1,2-Dichloroethene	0.698	0.806	0.803	0.808	0.787	0.797	0.783	5.4
Bromochloromethane	0.457	0.607	0.450	0.577	0.587	0.630	0.551	14.2
Chloroform	1.034	1.364	1.366	1.323	1.282	1.288	1.276	9.7
1,1,1-Trichloroethane	0.901	1.123	1.141	1.127	1.100	1.103	1.082	8.3
Methylcyclohexane	0.490	0.589	0.591	0.565	0.548	0.554	0.556	6.6
Benzene	1.330	1.577	1.566	1.523	1.473	1.459	1.488	6.1
1,2-Dichloroethane	0.513	0.595	0.600	0.582	0.554	0.553	0.566	5.8
Trichloroethene	0.296	0.382	0.382	0.370	0.364	0.363	0.360	8.9
1,2-Dichloropropane	0.326	0.403	0.394	0.397	0.382	0.384	0.381	7.4
Bromodichloromethane	0.461	0.599	0.601	0.607	0.579	0.586	0.572	9.7
4-Methyl-2-Pentanone	0.390	0.493	0.503	0.511	0.479	0.487	0.477	9.3
Toluene	0.841	0.967	0.960	0.940	0.897	0.897	0.917	5.2

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>AECO15</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q3128</u>
Instrument ID: <u>MSVOA_X</u>	Calibration Date(s): <u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>09:23</u> <u>12:01</u>
GC Column: <u>DB-624UI</u> ID: <u>0.18</u> (mm)	

LAB FILE ID:								
RRF001 = VX047585.D			RRF005 = VX047586.D			RRF020 = VX047587.D		
RRF050 = VX047588.D			RRF100 = VX047589.D			RRF150 = VX047590.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.487	0.588	0.602	0.619	0.595	0.611	0.584	8.3
cis-1,3-Dichloropropene	0.513	0.652	0.644	0.653	0.637	0.644	0.624	8.7
1,1,2-Trichloroethane	0.323	0.365	0.371	0.367	0.344	0.351	0.354	5.1
2-Hexanone	0.274	0.374	0.372	0.360	0.333	0.343	0.343	10.8
Dibromochloromethane	0.321	0.414	0.431	0.435	0.421	0.422	0.407	10.5
1,2-Dibromoethane	0.287	0.379	0.384	0.379	0.364	0.367	0.360	10.1
Tetrachloroethene	0.299	0.354	0.342	0.326	0.320	0.314	0.326	6
Chlorobenzene	0.998	1.182	1.177	1.178	1.139	1.142	1.136	6.2
Ethyl Benzene	1.625	2.077	2.049	2.039	1.996	1.973	1.960	8.6
m/p-Xylenes	0.604	0.774	0.765	0.765	0.739	0.730	0.729	8.8
o-Xylene	0.576	0.734	0.734	0.749	0.721	0.719	0.706	9.1
Styrene	1.019	1.288	1.290	1.299	1.257	1.265	1.236	8.7
Bromoform	0.221	0.293	0.301	0.313	0.312	0.317	0.293	12.4
Isopropylbenzene	3.124	3.907	3.986	4.047	3.907	3.839	3.801	8.9
1,1,2,2-Tetrachloroethane	0.965	1.182	1.203	1.225	1.152	1.173	1.150	8.2
1,3-Dichlorobenzene	1.462	1.800	1.829	1.817	1.768	1.758	1.739	8
1,4-Dichlorobenzene	1.570	1.890	1.868	1.832	1.767	1.783	1.785	6.5
1,2-Dichlorobenzene	1.359	1.691	1.757	1.746	1.696	1.692	1.657	9
1,2-Dibromo-3-Chloropropane	0.173	0.218	0.242	0.254	0.250	0.260	0.233	14.1
1,2,4-Trichlorobenzene	0.831	1.099	1.107	1.156	1.140	1.126	1.077	11.3
1,2,3-Trichlorobenzene	0.753	0.986	1.031	1.088	1.078	1.073	1.002	12.7
1,2-Dichloroethane-d4		0.884	0.647	0.770	0.815	0.863	0.796	11.8
Dibromofluoromethane		0.356	0.266	0.331	0.355	0.368	0.335	12.3
Toluene-d8		1.237	0.943	1.147	1.211	1.263	1.160	11.1
4-Bromofluorobenzene		0.478	0.360	0.427	0.452	0.484	0.440	11.4

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3128</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/25/2025 08:56</u>
Lab File ID:	<u>VX047795.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.627		21.04	20
Chloromethane	0.680	0.789	0.1	16.03	20
Vinyl Chloride	0.666	0.779		16.97	20
Bromomethane	0.422	0.471		11.61	20
Chloroethane	0.445	0.495		11.24	20
Trichlorofluoromethane	0.989	1.083		9.51	20
1,1,2-Trichlorotrifluoroethane	0.578	0.618		6.92	20
1,1-Dichloroethene	0.594	0.663		11.62	20
Acetone	0.346	0.336		-2.89	20
Carbon Disulfide	1.626	1.765		8.55	20
Methyl tert-butyl Ether	2.169	2.414		11.3	20
Methyl Acetate	0.770	0.985		27.92	20
Methylene Chloride	0.732	0.817		11.61	20
trans-1,2-Dichloroethene	0.640	0.703		9.84	20
1,1-Dichloroethane	1.263	1.370	0.1	8.47	20
Cyclohexane	1.052	1.056		0.38	20
2-Butanone	0.432	0.496		14.81	20
Carbon Tetrachloride	0.532	0.538		1.13	20
cis-1,2-Dichloroethene	0.783	0.854		9.07	20
Bromochloromethane	0.551	0.621		12.7	20
Chloroform	1.276	1.421		11.36	20
1,1,1-Trichloroethane	1.082	1.142		5.55	20
Methylcyclohexane	0.556	0.530		-4.68	20
Benzene	1.488	1.598		7.39	20
1,2-Dichloroethane	0.566	0.594		4.95	20
Trichloroethene	0.360	0.380		5.56	20
1,2-Dichloropropane	0.381	0.418		9.71	20
Bromodichloromethane	0.572	0.613		7.17	20
4-Methyl-2-Pentanone	0.477	0.566		18.66	20
Toluene	0.917	0.962		4.91	20
t-1,3-Dichloropropene	0.584	0.544		-6.85	20
cis-1,3-Dichloropropene	0.624	0.601		-3.69	20
1,1,2-Trichloroethane	0.354	0.390		10.17	20
2-Hexanone	0.343	0.392		14.29	20
Dibromochloromethane	0.407	0.449		10.32	20
1,2-Dibromoethane	0.360	0.409		13.61	20
Tetrachloroethene	0.326	0.353		8.28	20
Chlorobenzene	1.136	1.192	0.3	4.93	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3128</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/25/2025 08:56</u>
Lab File ID:	<u>VX047795.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	2.012		2.65	20
m/p-Xylenes	0.729	0.757		3.84	20
o-Xylene	0.706	0.737		4.39	20
Styrene	1.236	1.303		5.42	20
Bromoform	0.293	0.318	0.1	8.53	20
Isopropylbenzene	3.801	3.828		0.71	20
1,1,2,2-Tetrachloroethane	1.150	1.273	0.3	10.7	20
1,3-Dichlorobenzene	1.739	1.748		0.52	20
1,4-Dichlorobenzene	1.785	1.750		-1.96	20
1,2-Dichlorobenzene	1.657	1.689		1.93	20
1,2-Dibromo-3-Chloropropane	0.233	0.250		7.3	20
1,2,4-Trichlorobenzene	1.077	1.039		-3.53	20
1,2,3-Trichlorobenzene	1.002	1.016		1.4	20
1,2-Dichloroethane-d4	0.796	0.817		2.64	20
Dibromofluoromethane	0.335	0.351		4.78	20
Toluene-d8	1.160	1.182		1.9	20
4-Bromofluorobenzene	0.440	0.443		0.68	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3128</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/25/2025 19:09</u>
Lab File ID:	<u>VX047823.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.570		10.04	50
Chloromethane	0.680	0.706	0.1	3.82	50
Vinyl Chloride	0.666	0.715		7.36	50
Bromomethane	0.422	0.476		12.8	50
Chloroethane	0.445	0.468		5.17	50
Trichlorofluoromethane	0.989	1.012		2.33	50
1,1,2-Trichlorotrifluoroethane	0.578	0.582		0.69	50
1,1-Dichloroethene	0.594	0.611		2.86	50
Acetone	0.346	0.295		-14.74	50
Carbon Disulfide	1.626	1.666		2.46	50
Methyl tert-butyl Ether	2.169	2.267		4.52	50
Methyl Acetate	0.770	0.897		16.49	50
Methylene Chloride	0.732	0.760		3.83	50
trans-1,2-Dichloroethene	0.640	0.669		4.53	50
1,1-Dichloroethane	1.263	1.314	0.1	4.04	50
Cyclohexane	1.052	1.017		-3.33	50
2-Butanone	0.432	0.457		5.79	50
Carbon Tetrachloride	0.532	0.519		-2.44	50
cis-1,2-Dichloroethene	0.783	0.823		5.11	50
Bromochloromethane	0.551	0.605		9.8	50
Chloroform	1.276	1.355		6.19	50
1,1,1-Trichloroethane	1.082	1.111		2.68	50
Methylcyclohexane	0.556	0.527		-5.22	50
Benzene	1.488	1.537		3.29	50
1,2-Dichloroethane	0.566	0.566		0	50
Trichloroethene	0.360	0.370		2.78	50
1,2-Dichloropropane	0.381	0.403		5.77	50
Bromodichloromethane	0.572	0.597		4.37	50
4-Methyl-2-Pentanone	0.477	0.541		13.42	50
Toluene	0.917	0.955		4.14	50
t-1,3-Dichloropropene	0.584	0.574		-1.71	50
cis-1,3-Dichloropropene	0.624	0.627		0.48	50
1,1,2-Trichloroethane	0.354	0.385		8.76	50
2-Hexanone	0.343	0.377		9.91	50
Dibromochloromethane	0.407	0.444		9.09	50
1,2-Dibromoethane	0.360	0.402		11.67	50
Tetrachloroethene	0.326	0.327		0.31	50
Chlorobenzene	1.136	1.165	0.3	2.55	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3128</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/25/2025 19:09</u>
Lab File ID:	<u>VX047823.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	1.999		1.99	50
m/p-Xylenes	0.729	0.748		2.61	50
o-Xylene	0.706	0.733		3.82	50
Styrene	1.236	1.280		3.56	50
Bromoform	0.293	0.319	0.1	8.87	50
Isopropylbenzene	3.801	3.759		-1.11	50
1,1,2,2-Tetrachloroethane	1.150	1.240	0.3	7.83	50
1,3-Dichlorobenzene	1.739	1.729		-0.57	50
1,4-Dichlorobenzene	1.785	1.757		-1.57	50
1,2-Dichlorobenzene	1.657	1.693		2.17	50
1,2-Dibromo-3-Chloropropane	0.233	0.237		1.72	50
1,2,4-Trichlorobenzene	1.077	1.045		-2.97	50
1,2,3-Trichlorobenzene	1.002	1.018		1.6	50
1,2-Dichloroethane-d4	0.796	0.769		-3.39	50
Dibromofluoromethane	0.335	0.335		0	50
Toluene-d8	1.160	1.142		-1.55	50
4-Bromofluorobenzene	0.440	0.431		-2.05	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3128</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/26/2025 01:28</u>
Lab File ID:	<u>VX047825.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.564		8.88	20
Chloromethane	0.680	0.683	0.1	0.44	20
Vinyl Chloride	0.666	0.692		3.9	20
Bromomethane	0.422	0.381		-9.72	20
Chloroethane	0.445	0.437		-1.8	20
Trichlorofluoromethane	0.989	0.984		-0.51	20
1,1,2-Trichlorotrifluoroethane	0.578	0.573		-0.87	20
1,1-Dichloroethene	0.594	0.601		1.18	20
Acetone	0.346	0.311		-10.12	20
Carbon Disulfide	1.626	1.681		3.38	20
Methyl tert-butyl Ether	2.169	2.253		3.87	20
Methyl Acetate	0.770	0.935		21.43	20
Methylene Chloride	0.732	0.749		2.32	20
trans-1,2-Dichloroethene	0.640	0.660		3.13	20
1,1-Dichloroethane	1.263	1.302	0.1	3.09	20
Cyclohexane	1.052	1.000		-4.94	20
2-Butanone	0.432	0.473		9.49	20
Carbon Tetrachloride	0.532	0.512		-3.76	20
cis-1,2-Dichloroethene	0.783	0.806		2.94	20
Bromochloromethane	0.551	0.585		6.17	20
Chloroform	1.276	1.331		4.31	20
1,1,1-Trichloroethane	1.082	1.084		0.19	20
Methylcyclohexane	0.556	0.521		-6.3	20
Benzene	1.488	1.511		1.55	20
1,2-Dichloroethane	0.566	0.556		-1.77	20
Trichloroethene	0.360	0.364		1.11	20
1,2-Dichloropropane	0.381	0.399		4.72	20
Bromodichloromethane	0.572	0.592		3.5	20
4-Methyl-2-Pentanone	0.477	0.550		15.3	20
Toluene	0.917	0.932		1.64	20
t-1,3-Dichloropropene	0.584	0.564		-3.42	20
cis-1,3-Dichloropropene	0.624	0.599		-4.01	20
1,1,2-Trichloroethane	0.354	0.378		6.78	20
2-Hexanone	0.343	0.388		13.12	20
Dibromochloromethane	0.407	0.446		9.58	20
1,2-Dibromoethane	0.360	0.396		10	20
Tetrachloroethene	0.326	0.335		2.76	20
Chlorobenzene	1.136	1.147	0.3	0.97	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3128</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/26/2025</u> <u>01:28</u>
Lab File ID:	<u>VX047825.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	1.966		0.31	20
m/p-Xylenes	0.729	0.737		1.1	20
o-Xylene	0.706	0.722		2.27	20
Styrene	1.236	1.274		3.07	20
Bromoform	0.293	0.319	0.1	8.87	20
Isopropylbenzene	3.801	3.659		-3.74	20
1,1,2,2-Tetrachloroethane	1.150	1.237	0.3	7.57	20
1,3-Dichlorobenzene	1.739	1.702		-2.13	20
1,4-Dichlorobenzene	1.785	1.726		-3.31	20
1,2-Dichlorobenzene	1.657	1.670		0.79	20
1,2-Dibromo-3-Chloropropane	0.233	0.253		8.58	20
1,2,4-Trichlorobenzene	1.077	1.031		-4.27	20
1,2,3-Trichlorobenzene	1.002	1.021		1.9	20
1,2-Dichloroethane-d4	0.796	0.769		-3.39	20
Dibromofluoromethane	0.335	0.335		0	20
Toluene-d8	1.160	1.135		-2.15	20
4-Bromofluorobenzene	0.440	0.436		-0.91	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3128</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/26/2025 09:28</u>
Lab File ID:	<u>VX047846.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.606		16.99	50
Chloromethane	0.680	0.741	0.1	8.97	50
Vinyl Chloride	0.666	0.758		13.81	50
Bromomethane	0.422	0.484		14.69	50
Chloroethane	0.445	0.501		12.58	50
Trichlorofluoromethane	0.989	1.077		8.9	50
1,1,2-Trichlorotrifluoroethane	0.578	0.638		10.38	50
1,1-Dichloroethene	0.594	0.672		13.13	50
Acetone	0.346	0.348		0.58	50
Carbon Disulfide	1.626	1.770		8.86	50
Methyl tert-butyl Ether	2.169	2.523		16.32	50
Methyl Acetate	0.770	1.045		35.71	50
Methylene Chloride	0.732	0.842		15.03	50
trans-1,2-Dichloroethene	0.640	0.723		12.97	50
1,1-Dichloroethane	1.263	1.439	0.1	13.94	50
Cyclohexane	1.052	1.095		4.09	50
2-Butanone	0.432	0.519		20.14	50
Carbon Tetrachloride	0.532	0.548		3.01	50
cis-1,2-Dichloroethene	0.783	0.895		14.3	50
Bromochloromethane	0.551	0.605		9.8	50
Chloroform	1.276	1.490		16.77	50
1,1,1-Trichloroethane	1.082	1.212		12.02	50
Methylcyclohexane	0.556	0.546		-1.8	50
Benzene	1.488	1.635		9.88	50
1,2-Dichloroethane	0.566	0.617		9.01	50
Trichloroethene	0.360	0.393		9.17	50
1,2-Dichloropropane	0.381	0.434		13.91	50
Bromodichloromethane	0.572	0.648		13.29	50
4-Methyl-2-Pentanone	0.477	0.597		25.16	50
Toluene	0.917	1.016		10.8	50
t-1,3-Dichloropropene	0.584	0.586		0.34	50
cis-1,3-Dichloropropene	0.624	0.630		0.96	50
1,1,2-Trichloroethane	0.354	0.418		18.08	50
2-Hexanone	0.343	0.416		21.28	50
Dibromochloromethane	0.407	0.487		19.66	50
1,2-Dibromoethane	0.360	0.431		19.72	50
Tetrachloroethene	0.326	0.356		9.2	50
Chlorobenzene	1.136	1.232	0.3	8.45	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3128</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/26/2025</u> <u>09:28</u>
Lab File ID:	<u>VX047846.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	2.082		6.22	50
m/p-Xylenes	0.729	0.787		7.96	50
o-Xylene	0.706	0.772		9.35	50
Styrene	1.236	1.375		11.25	50
Bromoform	0.293	0.341	0.1	16.38	50
Isopropylbenzene	3.801	3.894		2.45	50
1,1,2,2-Tetrachloroethane	1.150	1.325	0.3	15.22	50
1,3-Dichlorobenzene	1.739	1.838		5.69	50
1,4-Dichlorobenzene	1.785	1.846		3.42	50
1,2-Dichlorobenzene	1.657	1.794		8.27	50
1,2-Dibromo-3-Chloropropane	0.233	0.266		14.16	50
1,2,4-Trichlorobenzene	1.077	1.131		5.01	50
1,2,3-Trichlorobenzene	1.002	1.103		10.08	50
1,2-Dichloroethane-d4	0.796	0.855		7.41	50
Dibromofluoromethane	0.335	0.359		7.16	50
Toluene-d8	1.160	1.223		5.43	50
4-Bromofluorobenzene	0.440	0.470		6.82	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.