

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3128

Analytical Method:

SW8260-Low

Client: AECOM Technical Services, Inc.

Datafile :

VX047844.D

		Sample				Rec	RPD		Limits		
Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID :	Q3128-02MS	Client Sample ID :	BPOW6-1-20250916MS								
Dichlorodifluoromethane	50	0	54.8	ug/L	110				32	152	
Chloromethane	50	0	55.6	ug/L	111				50	139	
Vinyl chloride	50	0	57.3	ug/L	115				58	137	
Bromomethane	50	0	49.6	ug/L	99				53	141	
Chloroethane	50	0	55.0	ug/L	110				60	138	
Trichlorofluoromethane	50	0	53.5	ug/L	107				65	141	
1,1,2-Trichlorotrifluoroethane	50	0	50.3	ug/L	101				70	136	
1,1-Dichloroethene	50	0	54.6	ug/L	109				71	131	
Acetone	250	1.90	240	ug/L	95				39	160	
Carbon disulfide	50	0	55.0	ug/L	110				64	133	
Methyl tert-butyl Ether	50	0	56.0	ug/L	112				71	124	
Methyl Acetate	50	0	51.9	ug/L	104				56	136	
Methylene Chloride	50	0	56.0	ug/L	112				74	124	
trans-1,2-Dichloroethene	50	0	56.0	ug/L	112				75	124	
1,1-Dichloroethane	50	0	55.9	ug/L	112				77	125	
Cyclohexane	50	0	48.2	ug/L	96				71	130	
2-Butanone	250	0	290	ug/L	116				56	143	
Carbon Tetrachloride	50	0	49.9	ug/L	100				72	136	
cis-1,2-Dichloroethene	50	0	56.0	ug/L	112				78	123	
Bromochloromethane	50	0	55.8	ug/L	112				78	123	
Chloroform	50	0	56.8	ug/L	114				79	124	
1,1,1-Trichloroethane	50	0	53.8	ug/L	108				74	131	
Methylcyclohexane	50	0	43.4	ug/L	87				72	132	
Benzene	50	0	53.1	ug/L	106				79	120	
1,2-Dichloroethane	50	0	52.0	ug/L	104				73	128	
Trichloroethene	50	0	52.0	ug/L	104				79	123	
1,2-Dichloropropane	50	0	54.2	ug/L	108				78	122	
Bromodichloromethane	50	0	54.2	ug/L	108				79	125	
4-Methyl-2-Pentanone	250	0	290	ug/L	116				67	130	
Toluene	50	0	52.9	ug/L	106				80	121	
t-1,3-Dichloropropene	50	0	47.1	ug/L	94				73	127	
cis-1,3-Dichloropropene	50	0	46.3	ug/L	93				75	124	
1,1,2-Trichloroethane	50	0	55.5	ug/L	111				80	119	
2-Hexanone	250	0	280	ug/L	112				57	139	
Dibromochloromethane	50	0	55.8	ug/L	112				74	126	
1,2-Dibromoethane	50	0	56.3	ug/L	113				77	121	
Tetrachloroethene	50	0	46.9	ug/L	94				74	129	
Chlorobenzene	50	0	50.3	ug/L	101				82	118	
Ethyl Benzene	50	0	49.2	ug/L	98				79	121	
m/p-Xylenes	100	0	99.6	ug/L	100				80	121	
o-Xylene	50	0	50.2	ug/L	100				78	122	
Styrene	50	0	50.0	ug/L	100				78	123	
Bromoform	50	0	52.0	ug/L	104				66	130	
Isopropylbenzene	50	0	46.8	ug/L	94				72	131	
1,1,2,2-Tetrachloroethane	50	0	52.0	ug/L	104				71	121	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3128

Analytical Method: SW8260-Low

Client: AECOM Technical Services, Inc.

Datafile : VX047844.D

Parameter	Spike	Sample	Result	Units	Rec			RPD		Limits		
		Result			Rec	Qual	RPD	Qual	Low	High	RPD	
1,3-Dichlorobenzene	50	0	46.8	ug/L	94					80	119	
1,4-Dichlorobenzene	50	0	46.3	ug/L	93					79	118	
1,2-Dichlorobenzene	50	0	48.4	ug/L	97					80	119	
1,2-Dibromo-3-Chloropropane	50	0	49.9	ug/L	100					62	128	
1,2,4-Trichlorobenzene	50	0	45.2	ug/L	90					69	130	
1,2,3-Trichlorobenzene	50	0	47.3	ug/L	95					69	129	