

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15
Lab Code: ACE SDG NO.: Q3128
Lab File ID: VX047825.D Date Analyzed: 09/26/2025
Instrument ID: MSVOA_X Time Analyzed: 01:28
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	136987	5.54	247384	6.75	222999	10.04
UPPER LIMIT	273974	6.038	494768	7.245	445998	10.537
LOWER LIMIT	68493.5	5.038	123692	6.245	111500	9.537
EPA SAMPLE NO.						
BPOW6-1-20250916	118010	5.54	250283	6.75	255095	10.04
BPOW6-1-20250916MS	132069	5.54	246705	6.75	231070	10.04
BPOW6-1-20250916MSD	130228	5.54	242264	6.75	226841	10.04
BPOW6-4-20250916	120108	5.54	253048	6.75	258299	10.04
BPOW6-9-20250916	118685	5.54	250694	6.75	261456	10.04
BPOW6-3-20250916	132775	5.54	283230	6.75	293419	10.04
TB07-20250912	132283	5.54	276404	6.75	283292	10.04
VX0925WBL02	121983	5.54	259174	6.75	269510	10.04
VX0925WBS02	143765	5.54	265116	6.75	243871	10.04

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.