

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15
 Lab Code: ACE SDG NO.: Q3128
 Lab File ID: VX047795.D Date Analyzed: 09/25/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:56
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	107538	12.00€				
UPPER LIMIT	215076	12.50€				
LOWER LIMIT	53769	11.50€				
EPA SAMPLE NO.						
BPOW6-2-20250916	130810	12.01				
BPOW6-10-20250916	132252	12.01				
BPOW6-11-20250916	128330	12.01				
BPOW6-8-20250916	128316	12.01				
BPOW6-7-20250917	128055	12.01				
BPOW6-6-20250917	143881	12.01				
BPOW6-5-20250917	131658	12.01				
VX0925WBL01	126490	12.01				
VX0925WBS01	131521	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.