

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15
 Lab Code: ACE SDG NO.: Q3128
 Lab File ID: VX047825.D Date Analyzed: 09/26/2025
 Instrument ID: MSVOA_X Time Analyzed: 01:28
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	110381	12.00€				
UPPER LIMIT	220762	12.50€				
LOWER LIMIT	55190.5	11.50€				
EPA SAMPLE NO.						
BPOW6-1-20250916	127833	12.01				
BPOW6-1-20250916MS	116018	12.01				
BPOW6-1-20250916MSD	112665	12.01				
BPOW6-4-20250916	127208	12.01				
BPOW6-9-20250916	133160	12.01				
BPOW6-3-20250916	148155	12.01				
TB07-20250912	140911	12.01				
VX0925WBL02	138899	12.01				
VX0925WBS02	122300	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.