

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: RTPE01

Lab Code: ACE

SDG No.: Q3130

Instrument ID: MSVOA_L

Calibration Date(s): 09/22/2025 09/22/2025

Heated Purge: (Y/N) N

Calibration Time(s): 09:39 13:05

GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:		RRF010 = VL042950.D		RRF002 = VL042951.D		RRF001 = VL042952.D		
		RRF0.5 = VL042953.D		RRF0.1 = VL042954.D		RRF.03 = VL042955.D		
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF.03	RRF	% RSD
Dichlorodifluoromethane	1.248	1.401	1.452	1.391			1.345	7.3
Chloromethane	0.560	0.624	0.622	0.693			0.609	9.7
Vinyl Chloride	0.470	0.524	0.567	0.529	0.562	0.729	0.550	15.9
Bromomethane	0.166	0.200	0.215	0.224			0.194	13.8
Chloroethane	0.196	0.212	0.238	0.234			0.214	10.3
Tetrahydrofuran	0.389	0.419	0.407	0.427			0.403	5.6
Trichlorofluoromethane	1.199	1.346	1.376	1.354			1.297	6.6
1,1,2-Trichlorotrifluoroethane	0.963	1.068	1.117	1.130			1.050	7.5
Dichlorotetrafluoroethane	1.047	1.180	1.239	1.227			1.148	8.2
tert-Butyl alcohol	1.351	1.398	1.544	1.587			1.429	9.5
Heptane	1.870	1.961	2.060	2.102			1.946	7.6
1,1-Dichloroethene	0.446	0.485	0.503	0.549			0.486	8.9
Acetone	0.993	1.475	1.552	1.490			1.302	21.5
Carbon Disulfide	1.271	1.396	1.436	1.395			1.346	6.6
Methyl tert-Butyl Ether	0.638	0.742	0.773	0.953			0.780	14.6
Methylene Chloride	0.388	0.510	0.617	0.832			0.545	34.4
trans-1,2-Dichloroethene	0.507	0.556	0.612	0.589			0.552	9
1,1-Dichloroethane	0.993	1.109	1.172	1.147			1.083	7.8
Cyclohexane	1.193	1.280	1.333	1.312			1.255	6.2
2-Butanone	0.655	0.741	0.731	0.744			0.701	7.5
Carbon Tetrachloride	0.610	0.657	0.665	0.660	0.616	0.734	0.651	6.6
cis-1,2-Dichloroethene	1.357	1.465	1.464	1.443			1.406	5.3
Chloroform	1.848	2.038	2.042	2.078			1.962	6.4
1,1,1-Trichloroethane	1.902	2.071	2.164	2.017	2.087	2.724	2.119	13.5
2,2,4-Trimethylpentane	1.563	1.690	1.660	1.705			1.622	5.6
Benzene	0.898	0.972	0.970	0.974			0.939	4.8
1,2-Dichloroethane	0.470	0.507	0.517	0.502			0.494	4.3
Trichloroethene	0.426	0.452	0.467	0.481	0.495	0.527	0.467	8.2
1,2-Dichloropropane	0.327	0.356	0.361	0.344			0.343	4.8
Bromodichloromethane	0.681	0.702	0.699	0.689			0.688	2.1

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: RTPE01

Lab Code: ACE

SDG No.: Q3130

Instrument ID: MSVOA_L

Calibration Date(s): 09/22/2025 09/22/2025

Heated Purge: (Y/N) N

Calibration Time(s): 09:39 13:05

GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:		RRF010 = VL042950.D		RRF002 = VL042951.D		RRF001 = VL042952.D		RRF0.5 = VL042953.D		RRF0.1 = VL042954.D		RRF.03 = VL042955.D	
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF.03	RRF	% RSD					
4-Methyl-2-Pentanone	0.934	0.988	0.949	0.928			0.939	3.5					
Toluene	1.078	1.156	1.137	1.155			1.111	5					
t-1,3-Dichloropropene	0.409	0.421	0.414	0.403			0.410	2					
cis-1,3-Dichloropropene	0.523	0.537	0.528	0.522			0.522	2.6					
1,1,2-Trichloroethane	0.344	0.371	0.377	0.382			0.362	5.7					
Dibromochloromethane	0.571	0.600	0.586	0.572			0.580	2.4					
1,2-Dibromoethane	0.490	0.525	0.520	0.520	0.520		0.510	3.5					
Tetrachloroethene	0.325	0.351	0.355	0.356	0.404	0.436	0.363	11.8					
Chlorobenzene	0.853	0.944	0.978	0.968			0.913	7.7					
Ethyl Benzene	1.570	1.737	1.715	1.657			1.642	5.5					
m/p-Xylene	1.225	1.361	1.365	1.332			1.295	6.3					
o-Xylene	1.221	1.359	1.379	1.335			1.294	7					
Styrene	0.602	0.648	0.619	0.605			0.612	3.9					
Bromoform	0.508	0.534	0.507	0.477			0.504	4.2					
1,1,2,2-Tetrachloroethane	0.610	0.673	0.675	0.688	0.722	0.806	0.680	10.7					
2-Chlorotoluene	1.373	1.528	1.504	1.543			1.456	6.6					
1,3,5-Trimethylbenzene	1.263	1.399	1.346	1.333			1.314	5.1					
1,2,4-Trimethylbenzene	1.349	1.558	1.522	1.534			1.451	8.3					
1,3-Dichlorobenzene	0.845	0.956	0.954	0.918			0.900	6.7					
1,4-Dichlorobenzene	0.845	0.959	0.914	0.922			0.894	6.1					
1,2-Dichlorobenzene	0.796	0.918	0.914	0.874			0.856	7.6					
1,2,4-Trichlorobenzene	0.718	0.777	0.691	0.557			0.686	11.7					
Hexachloro-1,3-Butadiene	0.575	0.728	0.742	0.755			0.670	14.8					
1,3-Butadiene	0.520	0.607	0.675	0.751			0.614	16.4					
Naphthalene	1.373	1.474	1.320	1.102	1.093		1.283	12					
4-Ethyltoluene	1.523	1.703	1.686	1.715			1.621	6.9					
1-Bromo-4-Fluorobenzene	0.763	0.747	0.743	0.749	0.738	0.736	0.746	1.2					
Hexane	1.398	1.553	1.614	1.621			1.501	9.1					
Allyl Chloride	0.857	0.966	0.971	0.904			0.907	6.7					
1,4-Dioxane	147.887	162.799	172.031	172.980			159.404	9					

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: RTPE01
 Lab Code: ACE SDG No.: Q3130
 Instrument ID: MSVOA_L Calibration Date(s): 09/22/2025 09/22/2025
 Heated Purge: (Y/N) N Calibration Time(s): 09:39 13:05
 GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:								
RRF010 = VL042950.D			RRF002 = VL042951.D			RRF001 = VL042952.D		
RRF0.5 = VL042953.D			RRF0.1 = VL042954.D			RRF.03 = VL042955.D		
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF.03	RRF	% RSD
Methyl Methacrylate	0.368	0.396	0.396	0.406			0.384	5.6

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.