

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: RTPE01
 Lab Code: ACE SDG No.: Q3132
 Instrument ID: MSVOA_L Calibration Date/Time: 09/22/2025 09:39
 Lab File ID: VL042950.D Init. Calib. Date(s): 09/22/2025 09/22/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 09:39 13:05
 GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	10.000	9.99	0.08000374	20