

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q3132	Analytical Method:	SWTO-15
Client:	RTP Environmental	Datafile :	VL042959.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0922ABS01	Dichlorodifluoromethane	10	9.30	ppbv	93			40 (70)	160 (130)	
	Chloromethane	10	8.70	ppbv	87			40 (70)	160 (130)	
	Vinyl Chloride	10	8.40	ppbv	84			70 (70)	130 (130)	
	Bromomethane	10	8.30	ppbv	83			40 (70)	160 (130)	
	Chloroethane	10	8.80	ppbv	88			40 (70)	160 (130)	
	Tetrahydrofuran	10	9.30	ppbv	93			70 (70)	130 (130)	
	Trichlorofluoromethane	10	9.10	ppbv	91			40 (70)	160 (130)	
	1,1,2-Trichlorotrifluoroethane	10	9.00	ppbv	90			70 (70)	130 (130)	
	Dichlorotetrafluoroethane	10	9.00	ppbv	90			70 (70)	130 (130)	
	tert-Butyl Alcohol	10	9.00	ppbv	90			70 (70)	130 (130)	
	Heptane	10	9.00	ppbv	90			70 (70)	130 (130)	
	1,1-Dichloroethene	10	9.00	ppbv	90			70 (70)	130 (130)	
	Acetone	10	7.40	ppbv	74			40 (70)	160 (130)	
	Carbon disulfide	10	8.80	ppbv	88			40 (70)	160 (130)	
	Methyl tert-butyl Ether	10	7.80	ppbv	78			70 (70)	130 (130)	
	Methylene Chloride	10	9.80	ppbv	98			70 (70)	130 (130)	
	trans-1,2-Dichloroethene	10	8.90	ppbv	89			70 (70)	130 (130)	
	1,1-Dichloroethane	10	9.00	ppbv	90			70 (70)	130 (130)	
	Cyclohexane	10	9.30	ppbv	93			70 (70)	130 (130)	
	2-Butanone	10	9.20	ppbv	92			40 (70)	160 (130)	
	Carbon Tetrachloride	10	9.70	ppbv	97			70 (70)	130 (130)	
	cis-1,2-Dichloroethene	10	9.30	ppbv	93			70 (70)	130 (130)	
	Chloroform	10	9.50	ppbv	95			70 (70)	130 (130)	
	1,1,1-Trichloroethane	10	9.00	ppbv	90			70 (70)	130 (130)	
	2,2,4-Trimethylpentane	10	9.40	ppbv	94			70 (70)	130 (130)	
	Benzene	10	9.60	ppbv	96			70 (70)	130 (130)	
	1,2-Dichloroethane	10	9.90	ppbv	99			70 (70)	130 (130)	
	Trichloroethene	10	9.20	ppbv	92			70 (70)	130 (130)	
	1,2-Dichloropropane	10	9.70	ppbv	97			70 (70)	130 (130)	
	Bromodichloromethane	10	9.90	ppbv	99			70 (70)	130 (130)	
	4-Methyl-2-Pentanone	10	9.60	ppbv	96			40 (70)	160 (130)	
	Toluene	10	9.50	ppbv	95			70 (70)	130 (130)	
	t-1,3-Dichloropropene	10	9.50	ppbv	95			70 (70)	130 (130)	
	cis-1,3-Dichloropropene	10	9.70	ppbv	97			70 (70)	130 (130)	
	1,1,2-Trichloroethane	10	9.60	ppbv	96			70 (70)	130 (130)	
	Dibromochloromethane	10	10.0	ppbv	100			70 (70)	130 (130)	
	1,2-Dibromoethane	10	9.70	ppbv	97			70 (70)	130 (130)	
	Tetrachloroethene	10	8.70	ppbv	87			70 (70)	130 (130)	
	Chlorobenzene	10	9.40	ppbv	94			70 (70)	130 (130)	
	Ethyl Benzene	10	9.50	ppbv	95			70 (70)	130 (130)	
	m/p-Xylene	20	18.9	ppbv	95			70 (70)	130 (130)	
	o-Xylene	10	9.40	ppbv	94			70 (70)	130 (130)	
	Styrene	10	9.60	ppbv	96			70 (70)	130 (130)	
	Bromoform	10	9.80	ppbv	98			70 (70)	130 (130)	
	1,1,2,2-Tetrachloroethane	10	8.90	ppbv	89			70 (70)	130 (130)	
	2-Chlorotoluene	10	9.30	ppbv	93			70 (70)	130 (130)	
	1,3,5-Trimethylbenzene	10	9.50	ppbv	95			70 (70)	130 (130)	
	1,2,4-Trimethylbenzene	10	9.30	ppbv	93			70 (70)	130 (130)	

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3132</u>	Analytical Method:	<u>SWTO-15</u>
Client:	<u>RTP Environmental</u>	Datafile :	<u>VL042959.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0922ABS01	1,3-Dichlorobenzene	10	9.30	ppbv	93			70 (70)	130 (130)	
	1,4-Dichlorobenzene	10	9.40	ppbv	94			70 (70)	130 (130)	
	1,2-Dichlorobenzene	10	9.30	ppbv	93			70 (70)	130 (130)	
	1,2,4-Trichlorobenzene	10	10.0	ppbv	100			70 (70)	130 (130)	
	Hexachloro-1,3-butadiene	10	8.40	ppbv	84			70 (70)	130 (130)	
	Naphthalene	10	10.3	ppbv	103			40 (70)	160 (130)	
	1,3-Butadiene	10	8.10	ppbv	81			70 (70)	130 (130)	
	4-Ethyltoluene	10	9.30	ppbv	93			70 (70)	130 (130)	
	Hexane	10	9.10	ppbv	91			70 (70)	130 (130)	
	Allyl Chloride	10	9.00	ppbv	90			70 (70)	130 (130)	
	1,4-Dioxane	10	9.30	ppbv	93			40 (70)	160 (130)	
	Methyl methacrylate	10	9.30	ppbv	93			70 (70)	130 (130)	