

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3141

Client ID : SSTDCCC0.4

Date Analyzed: 09/25/2025

Lab File ID: BN037883.D

Time Analyzed: 11:31

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	5686	7.818	16943	10.63	8938	14.47
UPPER LIMIT	11372	8.318	33886	11.126	17876	14.973
LOWER LIMIT	2843	7.318	8471.5	10.126	4469	13.973
EPA SAMPLE NO.						
01 BP-VPB-184-GW-60-62	6755	7.82	21716	10.63	14305	14.47
02 BP-VPB-184-GW-150-152	8047	7.82	25971	10.63	15628	14.47
03 BP-VPB-184-GW-220-222	9376	7.82	28623	10.63	16601	14.47
04 PB169793BL	6424	7.82	17072	10.63	8141	14.47
05 PB169793BS	8698	7.82	23812	10.63	11731	14.47
06 PB169793BSD	6010	7.82	16163	10.63	7938	14.47
07 BP-VPB-184-EB-20250918	5890	7.83	15941	10.63	8922	14.47

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.