

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3141</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/24/2025 02:08</u>
Lab File ID:	<u>VX047746.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.672		29.73	20
Chloromethane	0.680	0.783	0.1	15.15	20
Vinyl Chloride	0.666	0.789		18.47	20
Bromomethane	0.422	0.468		10.9	20
Chloroethane	0.445	0.492		10.56	20
Trichlorofluoromethane	0.989	1.101		11.32	20
1,1,2-Trichlorotrifluoroethane	0.578	0.655		13.32	20
1,1-Dichloroethene	0.594	0.663		11.62	20
Acetone	0.346	0.298		-13.87	20
Carbon Disulfide	1.626	1.796		10.45	20
Methyl tert-butyl Ether	2.169	2.253		3.87	20
Methyl Acetate	0.770	0.868		12.73	20
Methylene Chloride	0.732	0.771		5.33	20
trans-1,2-Dichloroethene	0.640	0.695		8.59	20
1,1-Dichloroethane	1.263	1.349	0.1	6.81	20
Cyclohexane	1.052	1.077		2.38	20
2-Butanone	0.432	0.438		1.39	20
Carbon Tetrachloride	0.532	0.554		4.14	20
cis-1,2-Dichloroethene	0.783	0.815		4.09	20
Bromochloromethane	0.551	0.629		14.16	20
Chloroform	1.276	1.328		4.07	20
1,1,1-Trichloroethane	1.082	1.123		3.79	20
Methylcyclohexane	0.556	0.576		3.6	20
Benzene	1.488	1.584		6.45	20
1,2-Dichloroethane	0.566	0.569		0.53	20
Trichloroethene	0.360	0.387		7.5	20
1,2-Dichloropropane	0.381	0.401		5.25	20
Bromodichloromethane	0.572	0.594		3.85	20
4-Methyl-2-Pentanone	0.477	0.516		8.18	20
Toluene	0.917	0.950		3.6	20
t-1,3-Dichloropropene	0.584	0.561		-3.94	20
cis-1,3-Dichloropropene	0.624	0.614		-1.6	20
1,1,2-Trichloroethane	0.354	0.369		4.24	20
2-Hexanone	0.343	0.357		4.08	20
Dibromochloromethane	0.407	0.431		5.9	20
1,2-Dibromoethane	0.360	0.382		6.11	20
Tetrachloroethene	0.326	0.346		6.14	20
Chlorobenzene	1.136	1.179	0.3	3.79	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3141</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/24/2025</u> <u>02:08</u>
Lab File ID:	<u>VX047746.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	2.043		4.24	20
m/p-Xylenes	0.729	0.773		6.04	20
o-Xylene	0.706	0.741		4.96	20
Styrene	1.236	1.308		5.82	20
Bromoform	0.293	0.317	0.1	8.19	20
Isopropylbenzene	3.801	3.949		3.89	20
1,1,2,2-Tetrachloroethane	1.150	1.228	0.3	6.78	20
1,3-Dichlorobenzene	1.739	1.760		1.21	20
1,4-Dichlorobenzene	1.785	1.788		0.17	20
1,2-Dichlorobenzene	1.657	1.706		2.96	20
1,2-Dibromo-3-Chloropropane	0.233	0.234		0.43	20
1,2,4-Trichlorobenzene	1.077	1.065		-1.11	20
1,2,3-Trichlorobenzene	1.002	1.020		1.8	20
1,2-Dichloroethane-d4	0.796	0.863		8.42	20
Dibromofluoromethane	0.335	0.398		18.81	20
Toluene-d8	1.160	1.325		14.22	20
4-Bromofluorobenzene	0.440	0.493		12.05	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.