

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3141</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/24/2025 10:50</u>
Lab File ID:	<u>VX047767.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.593		14.48	50
Chloromethane	0.680	0.735	0.1	8.09	50
Vinyl Chloride	0.666	0.726		9.01	50
Bromomethane	0.422	0.463		9.72	50
Chloroethane	0.445	0.466		4.72	50
Trichlorofluoromethane	0.989	1.013		2.43	50
1,1,2-Trichlorotrifluoroethane	0.578	0.586		1.38	50
1,1-Dichloroethene	0.594	0.631		6.23	50
Acetone	0.346	0.323		-6.65	50
Carbon Disulfide	1.626	1.720		5.78	50
Methyl tert-butyl Ether	2.169	2.352		8.44	50
Methyl Acetate	0.770	0.940		22.08	50
Methylene Chloride	0.732	0.779		6.42	50
trans-1,2-Dichloroethene	0.640	0.685		7.03	50
1,1-Dichloroethane	1.263	1.341	0.1	6.18	50
Cyclohexane	1.052	1.066		1.33	50
2-Butanone	0.432	0.500		15.74	50
Carbon Tetrachloride	0.532	0.515		-3.19	50
cis-1,2-Dichloroethene	0.783	0.835		6.64	50
Bromochloromethane	0.551	0.591		7.26	50
Chloroform	1.276	1.376		7.84	50
1,1,1-Trichloroethane	1.082	1.134		4.81	50
Methylcyclohexane	0.556	0.525		-5.58	50
Benzene	1.488	1.534		3.09	50
1,2-Dichloroethane	0.566	0.562		-0.71	50
Trichloroethene	0.360	0.371		3.06	50
1,2-Dichloropropane	0.381	0.401		5.25	50
Bromodichloromethane	0.572	0.591		3.32	50
4-Methyl-2-Pentanone	0.477	0.565		18.45	50
Toluene	0.917	0.943		2.84	50
t-1,3-Dichloropropene	0.584	0.540		-7.53	50
cis-1,3-Dichloropropene	0.624	0.587		-5.93	50
1,1,2-Trichloroethane	0.354	0.379		7.06	50
2-Hexanone	0.343	0.396		15.45	50
Dibromochloromethane	0.407	0.437		7.37	50
1,2-Dibromoethane	0.360	0.406		12.78	50
Tetrachloroethene	0.326	0.340		4.29	50
Chlorobenzene	1.136	1.145	0.3	0.79	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3141</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/24/2025</u> <u>10:50</u>
Lab File ID:	<u>VX047767.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	1.949		-0.56	50
m/p-Xylenes	0.729	0.741		1.65	50
o-Xylene	0.706	0.724		2.55	50
Styrene	1.236	1.277		3.32	50
Bromoform	0.293	0.316	0.1	7.85	50
Isopropylbenzene	3.801	3.694		-2.82	50
1,1,2,2-Tetrachloroethane	1.150	1.264	0.3	9.91	50
1,3-Dichlorobenzene	1.739	1.732		-0.4	50
1,4-Dichlorobenzene	1.785	1.729		-3.14	50
1,2-Dichlorobenzene	1.657	1.696		2.35	50
1,2-Dibromo-3-Chloropropane	0.233	0.254		9.01	50
1,2,4-Trichlorobenzene	1.077	1.059		-1.67	50
1,2,3-Trichlorobenzene	1.002	1.049		4.69	50
1,2-Dichloroethane-d4	0.796	0.866		8.79	50
Dibromofluoromethane	0.335	0.362		8.06	50
Toluene-d8	1.160	1.237		6.64	50
4-Bromofluorobenzene	0.440	0.475		7.95	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.