

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
 Data File : VX047719.D
 Acq On : 23 Sep 2025 09:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 24 01:25:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	-0.01
2 T	Dichlorodifluoromethane	0.518	0.656	-26.6#	110	0.00
3 P	Chloromethane	0.680	0.808	-18.8	103	0.00
4 C	Vinyl Chloride	0.666	0.805	-20.9#	102	0.00
5 T	Bromomethane	0.422	0.530	-25.6#	105	0.00
6 T	Chloroethane	0.445	0.514	-15.5	98	0.00
7 T	Trichlorofluoromethane	0.989	1.104	-11.6	93	0.00
8 T	Diethyl Ether	0.405	0.463	-14.3	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.578	0.650	-12.5	94	0.00
10 T	Methyl Iodide	0.902	1.047	-16.1	97	0.00
11 T	Tert butyl alcohol	0.089	0.087	2.2	82	0.00
12 CM	1,1-Dichloroethene	0.594	0.677	-14.0#	94	0.00
13 T	Acrolein	0.082	0.092	-12.2	95	0.00
14 T	Allyl chloride	1.226	1.342	-9.5	93	0.00
15 T	Acrylonitrile	0.328	0.378	-15.2	91	0.00
16 T	Acetone	0.346	0.388	-12.1	103	0.00
17 T	Carbon Disulfide	1.626	1.899	-16.8	102	0.00
18 T	Methyl Acetate	0.770	0.873	-13.4	84	0.00
19 T	Methyl tert-butyl Ether	2.169	2.452	-13.0	92	0.00
20 T	Methylene Chloride	0.732	0.832	-13.7	96	0.00
21 T	trans-1,2-Dichloroethene	0.640	0.741	-15.8	96	0.00
22 T	Diisopropyl ether	2.376	2.700	-13.6	92	0.00
23 T	Vinyl Acetate	1.901	2.204	-15.9	93	0.00
24 P	1,1-Dichloroethane	1.263	1.428	-13.1	93	0.00
25 T	2-Butanone	0.432	0.477	-10.4	91	0.00
26 T	2,2-Dichloropropane	1.050	1.149	-9.4	90	0.00
27 T	cis-1,2-Dichloroethene	0.783	0.903	-15.3	94	-0.01
28 T	Bromochloromethane	0.551	0.612	-11.1	89	-0.01
29 T	Tetrahydrofuran	0.260	0.285	-9.6	87	0.00
30 C	Chloroform	1.276	1.444	-13.2#	92	0.00
31 T	Cyclohexane	1.052	1.142	-8.6	93	0.00
32 T	1,1,1-Trichloroethane	1.082	1.211	-11.9	90	0.00
33 S	1,2-Dichloroethane-d4	0.796	0.816	-2.5	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	86	0.00
35 S	Dibromofluoromethane	0.335	0.352	-5.1	92	0.00
36 T	1,1-Dichloropropene	0.490	0.516	-5.3	91	0.00
37 T	Ethyl Acetate	0.525	0.559	-6.5	91	0.00
38 T	Carbon Tetrachloride	0.532	0.561	-5.5	90	0.00
39 T	Methylcyclohexane	0.556	0.599	-7.7	91	0.00
40 TM	Benzene	1.488	1.647	-10.7	93	0.00
41 T	Methacrylonitrile	0.271	0.298	-10.0	90	0.00
42 TM	1,2-Dichloroethane	0.566	0.606	-7.1	90	0.00
43 T	Isopropyl Acetate	0.860	0.931	-8.3	90	0.00
44 TM	Trichloroethene	0.360	0.403	-11.9	94	0.00
45 C	1,2-Dichloropropane	0.381	0.426	-11.8#	92	0.00
46 T	Dibromomethane	0.277	0.307	-10.8	92	0.00
47 T	Bromodichloromethane	0.572	0.634	-10.8	90	0.00
48 T	Methyl methacrylate	0.429	0.458	-6.8	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
 Data File : VX047719.D
 Acq On : 23 Sep 2025 09:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 24 01:25:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.004	0.005	-25.0	98	0.00
50 S	Toluene-d8	1.160	1.200	-3.4	90	0.00
51 T	4-Methyl-2-Pentanone	0.477	0.525	-10.1	89	0.00
52 CM	Toluene	0.917	1.016	-10.8#	93	0.00
53 T	t-1,3-Dichloropropene	0.584	0.658	-12.7	92	0.00
54 T	cis-1,3-Dichloropropene	0.624	0.704	-12.8	93	0.00
55 T	1,1,2-Trichloroethane	0.354	0.392	-10.7	92	0.00
56 T	Ethyl methacrylate	0.555	0.630	-13.5	91	0.00
57 T	1,3-Dichloropropane	0.607	0.689	-13.5	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.227	0.294	-29.5#	93	0.00
59 T	2-Hexanone	0.343	0.374	-9.0	90	0.00
60 T	Dibromochloromethane	0.407	0.464	-14.0	92	0.00
61 T	1,2-Dibromoethane	0.360	0.410	-13.9	93	0.00
62 S	4-Bromofluorobenzene	0.440	0.459	-4.3	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.326	0.353	-8.3	94	0.00
65 PM	Chlorobenzene	1.136	1.246	-9.7	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.392	0.440	-12.2	93	0.00
67 C	Ethyl Benzene	1.960	2.150	-9.7#	92	0.00
68 T	m/p-Xylenes	0.729	0.807	-10.7	92	0.00
69 T	o-Xylene	0.706	0.790	-11.9	92	0.00
70 T	Styrene	1.236	1.382	-11.8	93	0.00
71 P	Bromoform	0.293	0.332	-13.3	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	3.801	4.072	-7.1	91	0.00
74 T	N-amyl acetate	1.733	1.886	-8.8	91	0.00
75 P	1,1,2,2-Tetrachloroethane	1.150	1.253	-9.0	92	0.00
76 T	1,2,3-Trichloropropane	1.012	0.990	2.2	79	0.00
77 T	Bromobenzene	0.936	1.030	-10.0	94	0.00
78 T	n-propylbenzene	4.494	4.846	-7.8	91	0.00
79 T	2-Chlorotoluene	2.752	2.937	-6.7	91	0.00
80 T	1,3,5-Trimethylbenzene	3.123	3.365	-7.7	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.397	0.417	-5.0	91	0.00
82 T	4-Chlorotoluene	3.251	3.448	-6.1	90	0.00
83 T	tert-Butylbenzene	3.197	3.395	-6.2	91	0.00
84 T	1,2,4-Trimethylbenzene	3.166	3.429	-8.3	92	0.00
85 T	sec-Butylbenzene	3.797	4.071	-7.2	91	0.00
86 T	p-Isopropyltoluene	3.215	3.438	-6.9	91	0.00
87 T	1,3-Dichlorobenzene	1.739	1.887	-8.5	94	0.00
88 T	1,4-Dichlorobenzene	1.785	1.893	-6.1	93	0.00
89 T	n-Butylbenzene	2.975	3.171	-6.6	91	0.00
90 T	Hexachloroethane	0.564	0.609	-8.0	92	0.00
91 T	1,2-Dichlorobenzene	1.657	1.825	-10.1	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.233	0.241	-3.4	85	0.00
93 T	1,2,4-Trichlorobenzene	1.077	1.150	-6.8	90	0.00
94 T	Hexachlorobutadiene	0.366	0.387	-5.7	91	0.00
95 T	Naphthalene	3.279	3.524	-7.5	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
Data File : VX047719.D
Acq On : 23 Sep 2025 09:02
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 24 01:25:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.002	1.103	-10.1	91	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6