

LAB CHRONICLE

OrderID:	Q3141	OrderDate:	9/18/2025 3:41:00 PM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	J11,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3141-01	BP-VPB-184-TB-2025 0917	Water			09/17/25			09/18/25
			VOC-TCLVOA-10	8260-Low			09/23/25	
Q3141-02	BP-VPB-184-EB-2025 0918	Water			09/18/25			09/18/25
			VOC-TCLVOA-10	8260-Low			09/24/25	
Q3141-03	BP-VPB-184-GW-60-6 2	Water			09/17/25			09/18/25
			VOC-TCLVOA-10	8260-Low			09/23/25	
Q3141-04	BP-VPB-184-GW-100- 102	Water			09/17/25			09/18/25
			VOC-TCLVOA-10	8260-Low			09/24/25	
Q3141-05	BP-VPB-184-GW-150- 152	Water			09/18/25			09/18/25
			VOC-TCLVOA-10	8260-Low			09/23/25	
Q3141-06	BP-VPB-184-GW-200- 202	Water			09/18/25			09/18/25
			VOC-TCLVOA-10	8260-Low			09/23/25	
Q3141-07	BP-VPB-184-GW-220- 222	Water			09/18/25			09/18/25
			VOC-TCLVOA-10	8260-Low			09/23/25	

Hit Summary Sheet
SW-846

SDG No.: Q3141
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:	BP-VPB-184-EB-20250918								
Q3141-02	BP-VPB-184-EB-20 Water	Acetone		64.5		1.50	3.80	5.00	ug/L
Q3141-02	BP-VPB-184-EB-20 Water	2-Butanone		7.80		0.98	2.50	5.00	ug/L
		Total Voc :		72.3					
		Total Concentration:		72.3					
Client ID:	BP-VPB-184-GW-60-62								
Q3141-03	BP-VPB-184-GW-6 Water	Chloromethane		0.35	J	0.32	0.50	1.00	ug/L
Q3141-03	BP-VPB-184-GW-6 Water	Acetone		10.6		1.50	3.80	5.00	ug/L
Q3141-03	BP-VPB-184-GW-6 Water	2-Butanone		2.90	J	0.98	2.50	5.00	ug/L
Q3141-03	BP-VPB-184-GW-6 Water	Chloroform		1.60		0.25	0.50	1.00	ug/L
Q3141-03	BP-VPB-184-GW-6 Water	Benzene		0.76	J	0.15	0.50	1.00	ug/L
Q3141-03	BP-VPB-184-GW-6 Water	Toluene		0.39	J	0.14	0.50	1.00	ug/L
Q3141-03	BP-VPB-184-GW-6 Water	Dibromochloromethane		0.36	J	0.18	0.50	1.00	ug/L
Q3141-03	BP-VPB-184-GW-6 Water	Chlorobenzene		0.28	J	0.12	0.50	1.00	ug/L
		Total Voc :		17.2					
		Total Concentration:		17.2					
Client ID:	BP-VPB-184-GW-100-102								
Q3141-04	BP-VPB-184-GW-1 Water	Acetone		10.9		1.50	3.80	5.00	ug/L
Q3141-04	BP-VPB-184-GW-1 Water	Chloroform		0.67	J	0.25	0.50	1.00	ug/L
Q3141-04	BP-VPB-184-GW-1 Water	Trichloroethene		0.94	J	0.090	0.75	1.00	ug/L
Q3141-04	BP-VPB-184-GW-1 Water	Bromodichloromethane		0.35	J	0.22	0.50	1.00	ug/L
Q3141-04	BP-VPB-184-GW-1 Water	Toluene		0.34	J	0.14	0.50	1.00	ug/L
Q3141-04	BP-VPB-184-GW-1 Water	Dibromochloromethane		0.45	J	0.18	0.50	1.00	ug/L
Q3141-04	BP-VPB-184-GW-1 Water	Tetrachloroethene		0.26	J	0.23	0.50	1.00	ug/L
		Total Voc :		13.9					
		Total Concentration:		13.9					
Client ID:	BP-VPB-184-GW-150-152								
Q3141-05	BP-VPB-184-GW-1 Water	Acetone		6.40		1.50	3.80	5.00	ug/L
Q3141-05	BP-VPB-184-GW-1 Water	Carbon Disulfide		0.37	J	0.21	0.75	1.00	ug/L
Q3141-05	BP-VPB-184-GW-1 Water	Benzene		0.60	J	0.15	0.50	1.00	ug/L
		Total Voc :		7.37					
		Total Concentration:		7.37					
Client ID:	BP-VPB-184-GW-200-202								
Q3141-06	BP-VPB-184-GW-2 Water	Acetone		10.1		1.50	3.80	5.00	ug/L
		Total Voc :		10.1					
		Total Concentration:		10.1					
Client ID:	BP-VPB-184-GW-220-222								
Q3141-07	BP-VPB-184-GW-2 Water	Chloromethane		0.36	J	0.32	0.50	1.00	ug/L
Q3141-07	BP-VPB-184-GW-2 Water	Acetone		7.80		1.50	3.80	5.00	ug/L

Hit Summary Sheet
SW-846

SDG No.: Q3141

Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Total Voc :				8.16					
Total Concentration:				8.16					



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/17/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/18/25	
Client Sample ID:	BP-VPB-184-TB-20250917		SDG No.:	Q3141	
Lab Sample ID:	Q3141-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047725.D	1	09/23/25 12:22	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/17/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/18/25
Client Sample ID:	BP-VPB-184-TB-20250917	SDG No.:	Q3141
Lab Sample ID:	Q3141-01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047725.D	1	09/23/25 12:22	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	47.5		81 - 118		95%	SPK: 50
1868-53-7	Dibromofluoromethane	46.4		80 - 119		93%	SPK: 50
2037-26-5	Toluene-d8	47.5		89 - 112		95%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.1		85 - 114		104%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	131000	5.538				
540-36-3	1,4-Difluorobenzene	273000	6.745				
3114-55-4	Chlorobenzene-d5	269000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	128000	12.006				

TENTATIVE IDENTIFIED COMPOUNDS

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/17/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/18/25
Client Sample ID:	BP-VPB-184-TB-20250917		SDG No.:	Q3141
Lab Sample ID:	Q3141-01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047725.D	1	09/23/25 12:22	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/18/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/18/25	
Client Sample ID:	BP-VPB-184-EB-20250918		SDG No.:	Q3141	
Lab Sample ID:	Q3141-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047761.D	1	09/24/25 08:43	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	64.5		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	7.80		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/18/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/18/25
Client Sample ID:	BP-VPB-184-EB-20250918	SDG No.:	Q3141
Lab Sample ID:	Q3141-02	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047761.D	1	09/24/25 08:43	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.9		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	51.1		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	49.5		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.8		85 - 114		102%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	157000	5.538				
540-36-3	1,4-Difluorobenzene	294000	6.751				
3114-55-4	Chlorobenzene-d5	269000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	127000	12.006				

TENTATIVE IDENTIFIED COMPOUNDS

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/18/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/18/25	
Client Sample ID:	BP-VPB-184-EB-20250918		SDG No.:	Q3141	
Lab Sample ID:	Q3141-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047761.D	1	09/24/25 08:43	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/17/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/18/25	
Client Sample ID:	BP-VPB-184-GW-60-62		SDG No.:	Q3141	
Lab Sample ID:	Q3141-03		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047730.D	1	09/23/25 14:07	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.35	J	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	10.6		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.90	J	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	1.60		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.76	J	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.39	J	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	09/17/25
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	09/18/25
Client Sample ID:	BP-VPB-184-GW-60-62			SDG No.:	Q3141
Lab Sample ID:	Q3141-03			Matrix:	Water
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID :	0.18	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047730.D	1	09/23/25 14:07	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.36	J	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.28	J	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.5		81 - 118		105%	SPK: 50
1868-53-7	Dibromofluoromethane	48.6		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	48.3		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	55.6		85 - 114		111%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	142000	5.544				
540-36-3	1,4-Difluorobenzene	299000	6.751				
3114-55-4	Chlorobenzene-d5	301000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	156000	12.006				

TENTATIVE IDENTIFIED COMPOUNDS

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/17/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/18/25	
Client Sample ID:	BP-VPB-184-GW-60-62		SDG No.:	Q3141	
Lab Sample ID:	Q3141-03		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047730.D	1	09/23/25 14:07	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/17/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/18/25	
Client Sample ID:	BP-VPB-184-GW-100-102		SDG No.:	Q3141	
Lab Sample ID:	Q3141-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047760.D	1	09/24/25 08:22	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	10.9		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.67	J	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.94	J	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.35	J	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.34	J	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	09/17/25
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	09/18/25
Client Sample ID:	BP-VPB-184-GW-100-102			SDG No.:	Q3141
Lab Sample ID:	Q3141-04			Matrix:	Water
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID :	0.18	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047760.D	1	09/24/25 08:22	VX092325

[illegible]

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/17/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/18/25	
Client Sample ID:	BP-VPB-184-GW-100-102		SDG No.:	Q3141	
Lab Sample ID:	Q3141-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047760.D	1	09/24/25 08:22	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/18/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/18/25	
Client Sample ID:	BP-VPB-184-GW-150-152		SDG No.:	Q3141	
Lab Sample ID:	Q3141-05		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047732.D	1	09/23/25 14:49	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	6.40		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.37	J	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.60	J	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/18/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/18/25	
Client Sample ID:	BP-VPB-184-GW-150-152		SDG No.:	Q3141	
Lab Sample ID:	Q3141-05		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047732.D	1	09/23/25 14:49	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.8		81 - 118		102%	SPK: 50
1868-53-7	Dibromofluoromethane	48.4		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	48.7		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.6		85 - 114		109%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	145000	5.544				
540-36-3	1,4-Difluorobenzene	301000	6.751				
3114-55-4	Chlorobenzene-d5	306000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	153000	12.006				

TENTATIVE IDENTIFIED COMPOUNDS

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/18/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/18/25
Client Sample ID:	BP-VPB-184-GW-150-152		SDG No.:	Q3141
Lab Sample ID:	Q3141-05		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047732.D	1	09/23/25 14:49	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/18/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/18/25	
Client Sample ID:	BP-VPB-184-GW-200-202		SDG No.:	Q3141	
Lab Sample ID:	Q3141-06		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047733.D	1	09/23/25 15:10	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	10.1		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/18/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/18/25
Client Sample ID:	BP-VPB-184-GW-200-202	SDG No.:	Q3141
Lab Sample ID:	Q3141-06	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047733.D	1	09/23/25 15:10	VX092325

[illegible]

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/18/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/18/25	
Client Sample ID:	BP-VPB-184-GW-200-202		SDG No.:	Q3141	
Lab Sample ID:	Q3141-06		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047733.D	1	09/23/25 15:10	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/18/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/18/25	
Client Sample ID:	BP-VPB-184-GW-220-222		SDG No.:	Q3141	
Lab Sample ID:	Q3141-07		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047734.D	1	09/23/25 15:31	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.36	J	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	7.80		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/18/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/18/25
Client Sample ID:	BP-VPB-184-GW-220-222	SDG No.:	Q3141
Lab Sample ID:	Q3141-07	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047734.D	1	09/23/25 15:31	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.7		81 - 118		101%	SPK: 50
1868-53-7	Dibromofluoromethane	47.8		80 - 119		96%	SPK: 50
2037-26-5	Toluene-d8	47.7		89 - 112		95%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.0		85 - 114		108%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	147000	5.538				
540-36-3	1,4-Difluorobenzene	306000	6.745				
3114-55-4	Chlorobenzene-d5	308000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	158000	12.006				

TENTATIVE IDENTIFIED COMPOUNDS

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/18/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/18/25	
Client Sample ID:	BP-VPB-184-GW-220-222		SDG No.:	Q3141	
Lab Sample ID:	Q3141-07		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047734.D	1	09/23/25 15:31	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: **Q3141**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3141-01	BP-VPB-184-TB-20250917	1,2-Dichloroethane-d4	50	47.5	95		81	118
		Dibromofluoromethane	50	46.4	93		80	119
		Toluene-d8	50	47.5	95		89	112
		4-Bromofluorobenzene	50	52.1	104		85	114
Q3141-02	BP-VPB-184-EB-20250918	1,2-Dichloroethane-d4	50	53.0	106		81	118
		Dibromofluoromethane	50	51.1	102		80	119
		Toluene-d8	50	49.5	99		89	112
		4-Bromofluorobenzene	50	50.8	102		85	114
Q3141-03	BP-VPB-184-GW-60-62	1,2-Dichloroethane-d4	50	52.5	105		81	118
		Dibromofluoromethane	50	48.6	97		80	119
		Toluene-d8	50	48.3	97		89	112
		4-Bromofluorobenzene	50	55.6	111		85	114
Q3141-04	BP-VPB-184-GW-100-102	1,2-Dichloroethane-d4	50	56.9	114		81	118
		Dibromofluoromethane	50	56.6	113		80	119
		Toluene-d8	50	55.0	110		89	112
		4-Bromofluorobenzene	50	56.3	112		85	114
Q3141-05	BP-VPB-184-GW-150-152	1,2-Dichloroethane-d4	50	50.8	102		81	118
		Dibromofluoromethane	50	48.4	97		80	119
		Toluene-d8	50	48.7	97		89	112
		4-Bromofluorobenzene	50	54.6	109		85	114
Q3141-06	BP-VPB-184-GW-200-202	1,2-Dichloroethane-d4	50	52.0	104		81	118
		Dibromofluoromethane	50	48.4	97		80	119
		Toluene-d8	50	49.3	99		89	112
		4-Bromofluorobenzene	50	54.5	109		85	114
Q3141-07	BP-VPB-184-GW-220-222	1,2-Dichloroethane-d4	50	50.7	101		81	118
		Dibromofluoromethane	50	47.8	96		80	119
		Toluene-d8	50	47.7	95		89	112
		4-Bromofluorobenzene	50	54.0	108		85	114
VX0923WBL01	VX0923WBL01	1,2-Dichloroethane-d4	50	50.6	101		81	118
		Dibromofluoromethane	50	47.7	95		80	119
		Toluene-d8	50	48.0	96		89	112
		4-Bromofluorobenzene	50	56.5	113		85	114
VX0923WBL02	VX0923WBL02	1,2-Dichloroethane-d4	50	51.7	103		81	118
		Dibromofluoromethane	50	50.8	102		80	119
		Toluene-d8	50	50.2	100		89	112
		4-Bromofluorobenzene	50	51.8	103		85	114
VX0923WBS02	VX0923WBS02	1,2-Dichloroethane-d4	50	45.0	90		81	118
		Dibromofluoromethane	50	46.1	92		80	119
		Toluene-d8	50	46.4	93		89	112
		4-Bromofluorobenzene	50	46.8	94		85	114
VX0923WBS03	VX0923WBS03	1,2-Dichloroethane-d4	50	49.0	98		81	118
		Dibromofluoromethane	50	50.7	101		80	119
		Toluene-d8	50	50.6	101		89	112
		4-Bromofluorobenzene	50	51.1	102		85	114
VX0923WBSD0	VX0923WBSD03	1,2-Dichloroethane-d4	50	50.4	101		81	118
		Dibromofluoromethane	50	51.3	103		80	119
		Toluene-d8	50	52.1	104		89	112
		4-Bromofluorobenzene	50	52.1	104		85	114

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3141</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX047723.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0923WBS02	Dichlorodifluoromethane	20	19.2	ug/L	96			32	152	
	Chloromethane	20	17.6	ug/L	88			50	139	
	Vinyl chloride	20	17.4	ug/L	87			58	137	
	Bromomethane	20	18.5	ug/L	93			53	141	
	Chloroethane	20	16.5	ug/L	83			60	138	
	Trichlorofluoromethane	20	16.6	ug/L	83			65	141	
	1,1,2-Trichlorotrifluoroethane	20	17.1	ug/L	86			70	136	
	1,1-Dichloroethene	20	16.7	ug/L	84			71	131	
	Acetone	100	71.7	ug/L	72			39	160	
	Carbon disulfide	20	17.2	ug/L	86			64	133	
	Methyl tert-butyl Ether	20	16.3	ug/L	81			71	124	
	Methyl Acetate	20	16.6	ug/L	83			56	136	
	Methylene Chloride	20	16.4	ug/L	82			74	124	
	trans-1,2-Dichloroethene	20	17.0	ug/L	85			75	124	
	1,1-Dichloroethane	20	16.5	ug/L	83			77	125	
	Cyclohexane	20	16.7	ug/L	84			71	130	
	2-Butanone	100	78.7	ug/L	79			56	143	
	Carbon Tetrachloride	20	15.9	ug/L	79			72	136	
	cis-1,2-Dichloroethene	20	16.2	ug/L	81			78	123	
	Bromochloromethane	20	16.8	ug/L	84			78	123	
	Chloroform	20	16.6	ug/L	83			79	124	
	1,1,1-Trichloroethane	20	16.4	ug/L	82			74	131	
	Methylcyclohexane	20	17.3	ug/L	86			72	132	
	Benzene	20	16.7	ug/L	84			79	120	
	1,2-Dichloroethane	20	16.5	ug/L	83			73	128	
	Trichloroethene	20	17.5	ug/L	88			79	123	
	1,2-Dichloropropane	20	16.8	ug/L	84			78	122	
	Bromodichloromethane	20	16.6	ug/L	83			79	125	
	4-Methyl-2-Pentanone	100	86.0	ug/L	86			67	130	
	Toluene	20	16.9	ug/L	85			80	121	
	t-1,3-Dichloropropene	20	16.3	ug/L	81			73	127	
	cis-1,3-Dichloropropene	20	16.8	ug/L	84			75	124	
	1,1,2-Trichloroethane	20	17.2	ug/L	86			80	119	
	2-Hexanone	100	83.0	ug/L	83			57	139	
	Dibromochloromethane	20	17.2	ug/L	86			74	126	
	1,2-Dibromoethane	20	17.4	ug/L	87			77	121	
	Tetrachloroethene	20	17.1	ug/L	86			74	129	
	Chlorobenzene	20	16.7	ug/L	84			82	118	
	Ethyl Benzene	20	16.7	ug/L	84			79	121	
	m/p-Xylenes	40	33.9	ug/L	85			80	121	
	o-Xylene	20	16.9	ug/L	85			78	122	
	Styrene	20	16.6	ug/L	83			78	123	
	Bromoform	20	16.6	ug/L	83			66	130	
	Isopropylbenzene	20	16.4	ug/L	82			72	131	
	1,1,2,2-Tetrachloroethane	20	16.6	ug/L	83			71	121	
	1,3-Dichlorobenzene	20	16.9	ug/L	85			80	119	
	1,4-Dichlorobenzene	20	16.6	ug/L	83			79	118	
	1,2-Dichlorobenzene	20	16.7	ug/L	84			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3141</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX047723.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0923WBS02	1,2-Dibromo-3-Chloropropane	20	15.2	ug/L	76			62	128	
	1,2,4-Trichlorobenzene	20	16.2	ug/L	81			69	130	
	1,2,3-Trichlorobenzene	20	16.4	ug/L	82			69	129	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3141</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX047748.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0923WBS03	Dichlorodifluoromethane	20	23.6	ug/L	118			32	152	
	Chloromethane	20	21.7	ug/L	109			50	139	
	Vinyl chloride	20	22.4	ug/L	112			58	137	
	Bromomethane	20	23.5	ug/L	117			53	141	
	Chloroethane	20	21.0	ug/L	105			60	138	
	Trichlorofluoromethane	20	20.9	ug/L	104			65	141	
	1,1,2-Trichlorotrifluoroethane	20	21.0	ug/L	105			70	136	
	1,1-Dichloroethene	20	21.2	ug/L	106			71	131	
	Acetone	100	87.7	ug/L	88			39	160	
	Carbon disulfide	20	20.9	ug/L	104			64	133	
	Methyl tert-butyl Ether	20	20.6	ug/L	103			71	124	
	Methyl Acetate	20	22.4	ug/L	112			56	136	
	Methylene Chloride	20	20.7	ug/L	104			74	124	
	trans-1,2-Dichloroethene	20	21.4	ug/L	107			75	124	
	1,1-Dichloroethane	20	21.0	ug/L	105			77	125	
	Cyclohexane	20	20.9	ug/L	104			71	130	
	2-Butanone	100	110	ug/L	110			56	143	
	Carbon Tetrachloride	20	19.7	ug/L	99			72	136	
	cis-1,2-Dichloroethene	20	20.9	ug/L	104			78	123	
	Bromochloromethane	20	19.1	ug/L	96			78	123	
	Chloroform	20	21.3	ug/L	106			79	124	
	1,1,1-Trichloroethane	20	21.2	ug/L	106			74	131	
	Methylcyclohexane	20	20.1	ug/L	101			72	132	
	Benzene	20	20.9	ug/L	104			79	120	
	1,2-Dichloroethane	20	20.4	ug/L	102			73	128	
	Trichloroethene	20	21.3	ug/L	106			79	123	
	1,2-Dichloropropane	20	21.0	ug/L	105			78	122	
	Bromodichloromethane	20	20.7	ug/L	104			79	125	
	4-Methyl-2-Pentanone	100	110	ug/L	110			67	130	
	Toluene	20	21.1	ug/L	106			80	121	
	t-1,3-Dichloropropene	20	18.8	ug/L	94			73	127	
	cis-1,3-Dichloropropene	20	19.1	ug/L	96			75	124	
	1,1,2-Trichloroethane	20	21.7	ug/L	109			80	119	
	2-Hexanone	100	110	ug/L	110			57	139	
	Dibromochloromethane	20	21.2	ug/L	106			74	126	
	1,2-Dibromoethane	20	22.0	ug/L	110			77	121	
	Tetrachloroethene	20	21.4	ug/L	107			74	129	
	Chlorobenzene	20	20.7	ug/L	104			82	118	
	Ethyl Benzene	20	20.8	ug/L	104			79	121	
	m/p-Xylenes	40	42.3	ug/L	106			80	121	
	o-Xylene	20	20.9	ug/L	104			78	122	
	Styrene	20	20.7	ug/L	104			78	123	
	Bromoform	20	20.9	ug/L	104			66	130	
	Isopropylbenzene	20	20.5	ug/L	103			72	131	
	1,1,2,2-Tetrachloroethane	20	21.6	ug/L	108			71	121	
	1,3-Dichlorobenzene	20	20.6	ug/L	103			80	119	
	1,4-Dichlorobenzene	20	20.5	ug/L	103			79	118	
	1,2-Dichlorobenzene	20	20.7	ug/L	104			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3141</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX047748.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0923WBS03	1,2-Dibromo-3-Chloropropane	20	20.4	ug/L	102			62	128	
	1,2,4-Trichlorobenzene	20	19.8	ug/L	99			69	130	
	1,2,3-Trichlorobenzene	20	20.7	ug/L	104			69	129	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SDG No.:		Q3141	SW-846		Analytical Method:		SW8260-Low
Client:		Tetra Tech NUS, Inc.	Datafile :		VX047749.D		

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0923WBSD03	Dichlorodifluoromethane	20	19.6	ug/L	98	19		32	152	20
	Chloromethane	20	18.5	ug/L	93	16		50	139	20
	Vinyl chloride	20	18.4	ug/L	92	20		58	137	20
	Bromomethane	20	19.7	ug/L	99	17		53	141	20
	Chloroethane	20	17.6	ug/L	88	18		60	138	20
	Trichlorofluoromethane	20	17.6	ug/L	88	17		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	17.3	ug/L	86	20		70	136	20
	1,1-Dichloroethene	20	18.0	ug/L	90	16		71	131	20
	Acetone	100	75.9	ug/L	76	15		39	160	20
	Carbon disulfide	20	18.0	ug/L	90	14		64	133	20
	Methyl tert-butyl Ether	20	17.5	ug/L	88	16		71	124	20
	Methyl Acetate	20	19.8	ug/L	99	12		56	136	20
	Methylene Chloride	20	17.7	ug/L	89	16		74	124	20
	trans-1,2-Dichloroethene	20	17.8	ug/L	89	18		75	124	20
	1,1-Dichloroethane	20	17.6	ug/L	88	18		77	125	20
	Cyclohexane	20	17.4	ug/L	87	18		71	130	20
	2-Butanone	100	90.8	ug/L	91	19		56	143	20
	Carbon Tetrachloride	20	16.5	ug/L	83	18		72	136	20
	cis-1,2-Dichloroethene	20	17.9	ug/L	90	14		78	123	20
	Bromochloromethane	20	21.7	ug/L	109	13		78	123	20
	Chloroform	20	17.9	ug/L	90	16		79	124	20
	1,1,1-Trichloroethane	20	18.0	ug/L	90	16		74	131	20
	Methylcyclohexane	20	16.6	ug/L	83	20		72	132	20
	Benzene	20	17.7	ug/L	89	16		79	120	20
	1,2-Dichloroethane	20	16.8	ug/L	84	19		73	128	20
	Trichloroethene	20	17.9	ug/L	90	16		79	123	20
	1,2-Dichloropropane	20	17.9	ug/L	90	15		78	122	20
	Bromodichloromethane	20	17.5	ug/L	88	17		79	125	20
	4-Methyl-2-Pentanone	100	96.4	ug/L	96	14		67	130	20
	Toluene	20	17.7	ug/L	89	17		80	121	20
	t-1,3-Dichloropropene	20	16.0	ug/L	80	16		73	127	20
	cis-1,3-Dichloropropene	20	16.0	ug/L	80	18		75	124	20
	1,1,2-Trichloroethane	20	18.2	ug/L	91	18		80	119	20
	2-Hexanone	100	94.0	ug/L	94	16		57	139	20
	Dibromochloromethane	20	18.1	ug/L	91	15		74	126	20
	1,2-Dibromoethane	20	19.1	ug/L	96	14		77	121	20
	Tetrachloroethene	20	18.5	ug/L	93	14		74	129	20
	Chlorobenzene	20	17.7	ug/L	89	16		82	118	20
	Ethyl Benzene	20	17.2	ug/L	86	19		79	121	20
	m/p-Xylenes	40	35.3	ug/L	88	19		80	121	20
	o-Xylene	20	17.5	ug/L	88	17		78	122	20
	Styrene	20	17.5	ug/L	88	17		78	123	20
	Bromoform	20	17.5	ug/L	88	17		66	130	20
	Isopropylbenzene	20	16.9	ug/L	85	19		72	131	20
	1,1,2,2-Tetrachloroethane	20	18.5	ug/L	93	15		71	121	20
	1,3-Dichlorobenzene	20	17.4	ug/L	87	17		80	119	20
	1,4-Dichlorobenzene	20	17.1	ug/L	86	18		79	118	20
	1,2-Dichlorobenzene	20	17.5	ug/L	88	17		80	119	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3141</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX047749.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0923WBSD03	1,2-Dibromo-3-Chloropropane	20	17.6	ug/L	88	15		62	128	20
	1,2,4-Trichlorobenzene	20	16.4	ug/L	82	19		69	130	20
	1,2,3-Trichlorobenzene	20	17.1	ug/L	86	19		69	129	20

VOLATILE METHOD BLANK SUMMARY

Client ID

VX0923WBL01

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q3141

Lab File ID: VX047721.D

Lab Sample ID: VX0923WBL01

Date Analyzed: 09/23/2025

Time Analyzed: 10:22

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0923WBS02	VX0923WBS02	VX047723.D	09/23/2025
BP-VPB-184-TB-20250917	Q3141-01	VX047725.D	09/23/2025
BP-VPB-184-GW-60-62	Q3141-03	VX047730.D	09/23/2025
BP-VPB-184-GW-150-152	Q3141-05	VX047732.D	09/23/2025
BP-VPB-184-GW-200-202	Q3141-06	VX047733.D	09/23/2025
BP-VPB-184-GW-220-222	Q3141-07	VX047734.D	09/23/2025

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

Client ID

VX0923WBL02

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q3141

Lab File ID: VX047747.D

Lab Sample ID: VX0923WBL02

Date Analyzed: 09/24/2025

Time Analyzed: 03:11

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0923WBS03	VX0923WBS03	VX047748.D	09/24/2025
VX0923WBSD03	VX0923WBSD03	VX047749.D	09/24/2025
BP-VPB-184-GW-100-102	Q3141-04	VX047760.D	09/24/2025
BP-VPB-184-EB-20250918	Q3141-02	VX047761.D	09/24/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: VX047584.D
Instrument ID: MSVOA_X
GC Column: DB-624UI ID: 0.18 (mm)

Contract: TETR06
SDG NO.: Q3141
BFB Injection Date: 09/16/2025
BFB Injection Time: 08:56
Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20
75	30.0 - 60.0% of mass 95	54
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	67.3
175	5.0 - 9.0% of mass 174	5.4 (8.1) 1
176	95.0 - 101.0% of mass 174	66.3 (98.6) 1
177	5.0 - 9.0% of mass 176	4.3 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX047585.D	09/16/2025	09:23
VSTDIC005	VSTDIC005	VX047586.D	09/16/2025	10:16
VSTDIC020	VSTDIC020	VX047587.D	09/16/2025	10:37
VSTDIC050	VSTDIC050	VX047588.D	09/16/2025	10:58
VSTDIC100	VSTDIC100	VX047589.D	09/16/2025	11:40
VSTDIC150	VSTDIC150	VX047590.D	09/16/2025	12:01

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q3141

Lab File ID: VX047718.D

BFB Injection Date: 09/23/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:05

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.6
75	30.0 - 60.0% of mass 95	51.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.8 (1.2) 1
174	50.0 - 100.0% of mass 95	68.2
175	5.0 - 9.0% of mass 174	5.3 (7.8) 1
176	95.0 - 101.0% of mass 174	66.3 (97.2) 1
177	5.0 - 9.0% of mass 176	4.4 (6.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX047719.D	09/23/2025	09:02
VX0923WBL01	VX0923WBL01	VX047721.D	09/23/2025	10:22
VX0923WBS02	VX0923WBS02	VX047723.D	09/23/2025	11:15
BP-VPB-184-TB-20250917	Q3141-01	VX047725.D	09/23/2025	12:22
BP-VPB-184-GW-60-62	Q3141-03	VX047730.D	09/23/2025	14:07
BP-VPB-184-GW-150-152	Q3141-05	VX047732.D	09/23/2025	14:49
BP-VPB-184-GW-200-202	Q3141-06	VX047733.D	09/23/2025	15:10
BP-VPB-184-GW-220-222	Q3141-07	VX047734.D	09/23/2025	15:31
VSTDCCC050EC	VSTDCCC050	VX047744.D	09/23/2025	19:00

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q3141

Lab File ID: VX047745.D

BFB Injection Date: 09/24/2025

Instrument ID: MSVOA_X

BFB Injection Time: 01:30

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.9
75	30.0 - 60.0% of mass 95	51.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.8 (1.2) 1
174	50.0 - 100.0% of mass 95	68
175	5.0 - 9.0% of mass 174	5 (7.4) 1
176	95.0 - 101.0% of mass 174	65.9 (97) 1
177	5.0 - 9.0% of mass 176	4.3 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX047746.D	09/24/2025	02:08
VX0923WBL02	VX0923WBL02	VX047747.D	09/24/2025	03:11
VX0923WBS03	VX0923WBS03	VX047748.D	09/24/2025	03:53
VX0923WBSD03	VX0923WBSD03	VX047749.D	09/24/2025	04:14
BP-VPB-184-GW-100-102	Q3141-04	VX047760.D	09/24/2025	08:22
BP-VPB-184-EB-20250918	Q3141-02	VX047761.D	09/24/2025	08:43
VSTDCCC050EC	VSTDCCC050	VX047767.D	09/24/2025	10:50

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG NO.: Q3141
Lab File ID: VX047719.D Date Analyzed: 09/23/2025
Instrument ID: MSVOA_X Time Analyzed: 09:02
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	145672	5.53	265022	6.74	235703	10.04
UPPER LIMIT	291344	6.031	530044	7.239	471406	10.537
LOWER LIMIT	72836	5.031	132511	6.239	117852	9.537
EPA SAMPLE NO.						
BP-VPB-184-TB-20250917	131211	5.54	272690	6.75	268706	10.04
BP-VPB-184-GW-60-62	141797	5.54	298818	6.75	301046	10.04
BP-VPB-184-GW-150-152	145260	5.54	301134	6.75	306124	10.04
BP-VPB-184-GW-200-202	141844	5.54	300954	6.75	298268	10.04
BP-VPB-184-GW-220-222	147065	5.54	305643	6.75	308233	10.04
VX0923WBL01	119065	5.53	248954	6.74	254277	10.04
VX0923WBS02	181787	5.53	322664	6.75	289792	10.04

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG NO.: Q3141
 Lab File ID: VX047719.D Date Analyzed: 09/23/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:02
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	115061	12.00€				
UPPER LIMIT	230122	12.50€				
LOWER LIMIT	57530.5	11.50€				
EPA SAMPLE NO.						
BP-VPB-184-TB-20250917	128339	12.01				
BP-VPB-184-GW-60-62	155521	12.01				
BP-VPB-184-GW-150-152	153459	12.01				
BP-VPB-184-GW-200-202	146402	12.01				
BP-VPB-184-GW-220-222	157643	12.01				
VX0923WBL01	132951	12.01				
VX0923WBS02	141661	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG NO.: Q3141
 Lab File ID: VX047746.D Date Analyzed: 09/24/2025
 Instrument ID: MSVOA_X Time Analyzed: 02:08
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	166745	5.53	290663	6.74	252770	10.04
UPPER LIMIT	333490	6.032	581326	7.239	505540	10.537
LOWER LIMIT	83372.5	5.032	145332	6.239	126385	9.537
EPA SAMPLE NO.						
BP-VPB-184-EB-20250918	157285	5.54	293648	6.75	269079	10.04
BP-VPB-184-GW-100-102	165497	5.54	300349	6.75	276860	10.04
VX0923WBL02	197095	5.54	377977	6.75	356070	10.04
VX0923WBS03	156932	5.54	284011	6.75	253957	10.04
VX0923WBSD03	154222	5.54	281988	6.75	254297	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3141</u>
Lab File ID:	<u>VX047746.D</u>	Date Analyzed:	<u>09/24/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>02:08</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	121948	12.00€				
UPPER LIMIT	243896	12.50€				
LOWER LIMIT	60974	11.50€				
EPA SAMPLE NO.						
BP-VPB-184-EB-20250918	126784	12.01				
BP-VPB-184-GW-100-102	136496	12.01				
VX0923WBL02	171478	12.01				
VX0923WBS03	123765	12.01				
VX0923WBSD03	124629	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0923WBL01		SDG No.:	Q3141
Lab Sample ID:	VX0923WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047721.D	1	09/23/25 10:22	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0923WBL01		SDG No.:	Q3141
Lab Sample ID:	VX0923WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047721.D	1	09/23/25 10:22	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.6		81 - 118		101%	SPK: 50
1868-53-7	Dibromofluoromethane	47.7		80 - 119		95%	SPK: 50
2037-26-5	Toluene-d8	48.0		89 - 112		96%	SPK: 50
460-00-4	4-Bromofluorobenzene	56.5		85 - 114		113%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	119000	5.531				
540-36-3	1,4-Difluorobenzene	249000	6.739				
3114-55-4	Chlorobenzene-d5	254000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	133000	12.006				

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0923WBL01		SDG No.:	Q3141
Lab Sample ID:	VX0923WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047721.D	1	09/23/25 10:22	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0923WBL02		SDG No.:	Q3141
Lab Sample ID:	VX0923WBL02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047747.D	1	09/24/25 03:11	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0923WBL02		SDG No.:	Q3141
Lab Sample ID:	VX0923WBL02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047747.D	1	09/24/25 03:11	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.7		81 - 118		103%	SPK: 50
1868-53-7	Dibromofluoromethane	50.8		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	50.2		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.7		85 - 114		103%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	197000	5.544				
540-36-3	1,4-Difluorobenzene	378000	6.751				
3114-55-4	Chlorobenzene-d5	356000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	171000	12.006				

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0923WBL02		SDG No.:	Q3141
Lab Sample ID:	VX0923WBL02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047747.D	1	09/24/25 03:11	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0923WBS02		SDG No.:	Q3141
Lab Sample ID:	VX0923WBS02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047723.D	1	09/23/25 11:15	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.2		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	17.6		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	17.4		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	18.5		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	16.5		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	16.6		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	17.1		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	16.7		0.23	0.75	1.00	ug/L
67-64-1	Acetone	71.7		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	17.2		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	16.3		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	16.6		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	16.4		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	17.0		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	16.5		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	16.7		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	78.7		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	15.9		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	16.2		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	16.8		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	16.6		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	16.4		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	17.3		0.16	0.50	1.00	ug/L
71-43-2	Benzene	16.7		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	16.5		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	17.5		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	16.8		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	16.6		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	86.0		0.68	2.50	5.00	ug/L
108-88-3	Toluene	16.9		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0923WBS02		SDG No.:	Q3141
Lab Sample ID:	VX0923WBS02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047723.D	1	09/23/25 11:15	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	16.3		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	16.8		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	17.2		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	83.0		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	17.2		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	17.4		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	17.1		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	16.7		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	16.7		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	33.9		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	16.9		0.12	0.50	1.00	ug/L
100-42-5	Styrene	16.6		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	16.6		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	16.4		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	16.6		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	16.9		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	16.6		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	16.7		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	15.2		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	16.2		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	16.4		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	45.0		81 - 118		90%	SPK: 50
1868-53-7	Dibromofluoromethane	46.1		80 - 119		92%	SPK: 50
2037-26-5	Toluene-d8	46.4		89 - 112		93%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.8		85 - 114		94%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	182000	5.532				
540-36-3	1,4-Difluorobenzene	323000	6.745				
3114-55-4	Chlorobenzene-d5	290000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	142000	12.006				

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0923WBS02		SDG No.:	Q3141
Lab Sample ID:	VX0923WBS02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047723.D	1	09/23/25 11:15	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0923WBS03		SDG No.:	Q3141
Lab Sample ID:	VX0923WBS03		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047748.D	1	09/24/25 03:53	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	23.6		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	21.7		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	22.4		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	23.5		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	21.0		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	20.9		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	21.0		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	21.2		0.23	0.75	1.00	ug/L
67-64-1	Acetone	87.7		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	20.9		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	20.6		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	22.4		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	20.7		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	21.4		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	21.0		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	20.9		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	110		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.7		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	20.9		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	19.1		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	21.3		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	21.2		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	20.1		0.16	0.50	1.00	ug/L
71-43-2	Benzene	20.9		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.4		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	21.3		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	21.0		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	20.7		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.68	2.50	5.00	ug/L
108-88-3	Toluene	21.1		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0923WBS03		SDG No.:	Q3141
Lab Sample ID:	VX0923WBS03		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047748.D	1	09/24/25 03:53	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	18.8		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.1		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.7		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	21.2		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	22.0		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	21.4		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	20.7		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	20.8		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	42.3		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	20.9		0.12	0.50	1.00	ug/L
100-42-5	Styrene	20.7		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	20.9		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	20.5		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	21.6		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	20.6		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	20.5		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.7		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	20.4		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	19.8		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	20.7		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	49.0		81 - 118		98%	SPK: 50
1868-53-7	Dibromofluoromethane	50.7		80 - 119		101%	SPK: 50
2037-26-5	Toluene-d8	50.6		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.1		85 - 114		102%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	157000	5.544				
540-36-3	1,4-Difluorobenzene	284000	6.751				
3114-55-4	Chlorobenzene-d5	254000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	124000	12.006				

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0923WBS03		SDG No.:	Q3141
Lab Sample ID:	VX0923WBS03		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047748.D	1	09/24/25 03:53	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0923WBSD03		SDG No.:	Q3141
Lab Sample ID:	VX0923WBSD03		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047749.D	1	09/24/25 04:14	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.6		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	18.5		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	18.4		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	19.7		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	17.6		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	17.6		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	17.3		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.0		0.23	0.75	1.00	ug/L
67-64-1	Acetone	75.9		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	18.0		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	17.5		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	19.8		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	17.7		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	17.8		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	17.6		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	17.4		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	90.8		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	16.5		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	17.9		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	21.7		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	17.9		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	18.0		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	16.6		0.16	0.50	1.00	ug/L
71-43-2	Benzene	17.7		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	16.8		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	17.9		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	17.9		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	17.5		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	96.4		0.68	2.50	5.00	ug/L
108-88-3	Toluene	17.7		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0923WBSD03		SDG No.:	Q3141
Lab Sample ID:	VX0923WBSD03		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047749.D	1	09/24/25 04:14	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	16.0		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	16.0		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	18.2		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	94.0		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	18.1		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	19.1		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	18.5		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	17.7		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	17.2		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	35.3		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	17.5		0.12	0.50	1.00	ug/L
100-42-5	Styrene	17.5		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	17.5		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	16.9		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	18.5		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	17.4		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	17.1		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	17.5		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	17.6		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	16.4		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	17.1		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.4		81 - 118		101%	SPK: 50
1868-53-7	Dibromofluoromethane	51.3		80 - 119		103%	SPK: 50
2037-26-5	Toluene-d8	52.2		89 - 112		104%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.1		85 - 114		104%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	154000	5.537				
540-36-3	1,4-Difluorobenzene	282000	6.751				
3114-55-4	Chlorobenzene-d5	254000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	125000	12.006				

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0923WBSD03		SDG No.:	Q3141
Lab Sample ID:	VX0923WBSD03		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047749.D	1	09/24/25 04:14	VX092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3141</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Calibration Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u> ID: <u>0.18</u> (mm)		

LAB FILE ID:	RRF001 = VX047585.D			RRF005 = VX047586.D		RRF020 = VX047587.D		
	RRF050 = VX047588.D			RRF100 = VX047589.D		RRF150 = VX047590.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.447	0.575	0.582	0.501	0.510	0.492	0.518	10
Chloromethane	0.610	0.749	0.739	0.661	0.677	0.647	0.680	7.9
Vinyl Chloride	0.572	0.713	0.716	0.661	0.677	0.658	0.666	7.9
Bromomethane		0.493	0.451	0.424	0.423	0.321	0.422	15
Chloroethane	0.433	0.450	0.469	0.443	0.450	0.426	0.445	3.4
Trichlorofluoromethane	0.865	1.033	1.038	0.995	1.018	0.986	0.989	6.5
1,1,2-Trichlorotrifluoroethane	0.519	0.591	0.603	0.583	0.590	0.583	0.578	5.2
1,1-Dichloroethene	0.526	0.612	0.618	0.604	0.609	0.596	0.594	5.7
Acetone	0.343	0.416	0.400	0.317	0.302	0.299	0.346	14.7
Carbon Disulfide	1.555	1.735	1.758	1.566	1.589	1.555	1.626	5.8
Methyl tert-butyl Ether	1.884	2.238	2.221	2.247	2.212	2.209	2.169	6.5
Methyl Acetate	0.612	0.673	0.714	0.876	0.865	0.881	0.770	15.4
Methylene Chloride	0.718	0.772	0.754	0.730	0.716	0.705	0.732	3.5
trans-1,2-Dichloroethene	0.543	0.703	0.665	0.649	0.643	0.635	0.640	8.3
1,1-Dichloroethane	1.090	1.325	1.299	1.297	1.287	1.278	1.263	6.8
Cyclohexane		1.161	1.092	1.030	0.984	0.993	1.052	7.1
2-Butanone	0.370	0.470	0.473	0.439	0.414	0.424	0.432	9
Carbon Tetrachloride	0.482	0.574	0.555	0.539	0.526	0.518	0.532	6
cis-1,2-Dichloroethene	0.698	0.806	0.803	0.808	0.787	0.797	0.783	5.4
Bromochloromethane	0.457	0.607	0.450	0.577	0.587	0.630	0.551	14.2
Chloroform	1.034	1.364	1.366	1.323	1.282	1.288	1.276	9.7
1,1,1-Trichloroethane	0.901	1.123	1.141	1.127	1.100	1.103	1.082	8.3
Methylcyclohexane	0.490	0.589	0.591	0.565	0.548	0.554	0.556	6.6
Benzene	1.330	1.577	1.566	1.523	1.473	1.459	1.488	6.1
1,2-Dichloroethane	0.513	0.595	0.600	0.582	0.554	0.553	0.566	5.8
Trichloroethene	0.296	0.382	0.382	0.370	0.364	0.363	0.360	8.9
1,2-Dichloropropane	0.326	0.403	0.394	0.397	0.382	0.384	0.381	7.4
Bromodichloromethane	0.461	0.599	0.601	0.607	0.579	0.586	0.572	9.7
4-Methyl-2-Pentanone	0.390	0.493	0.503	0.511	0.479	0.487	0.477	9.3
Toluene	0.841	0.967	0.960	0.940	0.897	0.897	0.917	5.2

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3141</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Calibration Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u> ID: <u>0.18</u> (mm)		

LAB FILE ID:		RRF001 = VX047585.D		RRF005 = VX047586.D		RRF020 = VX047587.D		
		RRF050 = VX047588.D		RRF100 = VX047589.D		RRF150 = VX047590.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.487	0.588	0.602	0.619	0.595	0.611	0.584	8.3
cis-1,3-Dichloropropene	0.513	0.652	0.644	0.653	0.637	0.644	0.624	8.7
1,1,2-Trichloroethane	0.323	0.365	0.371	0.367	0.344	0.351	0.354	5.1
2-Hexanone	0.274	0.374	0.372	0.360	0.333	0.343	0.343	10.8
Dibromochloromethane	0.321	0.414	0.431	0.435	0.421	0.422	0.407	10.5
1,2-Dibromoethane	0.287	0.379	0.384	0.379	0.364	0.367	0.360	10.1
Tetrachloroethene	0.299	0.354	0.342	0.326	0.320	0.314	0.326	6
Chlorobenzene	0.998	1.182	1.177	1.178	1.139	1.142	1.136	6.2
Ethyl Benzene	1.625	2.077	2.049	2.039	1.996	1.973	1.960	8.6
m/p-Xylenes	0.604	0.774	0.765	0.765	0.739	0.730	0.729	8.8
o-Xylene	0.576	0.734	0.734	0.749	0.721	0.719	0.706	9.1
Styrene	1.019	1.288	1.290	1.299	1.257	1.265	1.236	8.7
Bromoform	0.221	0.293	0.301	0.313	0.312	0.317	0.293	12.4
Isopropylbenzene	3.124	3.907	3.986	4.047	3.907	3.839	3.801	8.9
1,1,2,2-Tetrachloroethane	0.965	1.182	1.203	1.225	1.152	1.173	1.150	8.2
1,3-Dichlorobenzene	1.462	1.800	1.829	1.817	1.768	1.758	1.739	8
1,4-Dichlorobenzene	1.570	1.890	1.868	1.832	1.767	1.783	1.785	6.5
1,2-Dichlorobenzene	1.359	1.691	1.757	1.746	1.696	1.692	1.657	9
1,2-Dibromo-3-Chloropropane	0.173	0.218	0.242	0.254	0.250	0.260	0.233	14.1
1,2,4-Trichlorobenzene	0.831	1.099	1.107	1.156	1.140	1.126	1.077	11.3
1,2,3-Trichlorobenzene	0.753	0.986	1.031	1.088	1.078	1.073	1.002	12.7
1,2-Dichloroethane-d4		0.884	0.647	0.770	0.815	0.863	0.796	11.8
Dibromofluoromethane		0.356	0.266	0.331	0.355	0.368	0.335	12.3
Toluene-d8		1.237	0.943	1.147	1.211	1.263	1.160	11.1
4-Bromofluorobenzene		0.478	0.360	0.427	0.452	0.484	0.440	11.4

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3141</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/23/2025</u> <u>09:02</u>
Lab File ID:	<u>VX047719.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.656		26.64	20
Chloromethane	0.680	0.808	0.1	18.82	20
Vinyl Chloride	0.666	0.805		20.87	20
Bromomethane	0.422	0.530		25.59	20
Chloroethane	0.445	0.514		15.51	20
Trichlorofluoromethane	0.989	1.104		11.63	20
1,1,2-Trichlorotrifluoroethane	0.578	0.650		12.46	20
1,1-Dichloroethene	0.594	0.677		13.97	20
Acetone	0.346	0.388		12.14	20
Carbon Disulfide	1.626	1.899		16.79	20
Methyl tert-butyl Ether	2.169	2.452		13.05	20
Methyl Acetate	0.770	0.873		13.38	20
Methylene Chloride	0.732	0.832		13.66	20
trans-1,2-Dichloroethene	0.640	0.741		15.78	20
1,1-Dichloroethane	1.263	1.428	0.1	13.06	20
Cyclohexane	1.052	1.142		8.56	20
2-Butanone	0.432	0.477		10.42	20
Carbon Tetrachloride	0.532	0.561		5.45	20
cis-1,2-Dichloroethene	0.783	0.903		15.33	20
Bromochloromethane	0.551	0.612		11.07	20
Chloroform	1.276	1.444		13.17	20
1,1,1-Trichloroethane	1.082	1.211		11.92	20
Methylcyclohexane	0.556	0.599		7.73	20
Benzene	1.488	1.647		10.69	20
1,2-Dichloroethane	0.566	0.606		7.07	20
Trichloroethene	0.360	0.403		11.94	20
1,2-Dichloropropane	0.381	0.426		11.81	20
Bromodichloromethane	0.572	0.634		10.84	20
4-Methyl-2-Pentanone	0.477	0.525		10.06	20
Toluene	0.917	1.016		10.8	20
t-1,3-Dichloropropene	0.584	0.658		12.67	20
cis-1,3-Dichloropropene	0.624	0.704		12.82	20
1,1,2-Trichloroethane	0.354	0.392		10.73	20
2-Hexanone	0.343	0.374		9.04	20
Dibromochloromethane	0.407	0.464		14.01	20
1,2-Dibromoethane	0.360	0.410		13.89	20
Tetrachloroethene	0.326	0.353		8.28	20
Chlorobenzene	1.136	1.246	0.3	9.68	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3141</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/23/2025</u> <u>09:02</u>
Lab File ID:	<u>VX047719.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	2.150		9.69	20
m/p-Xylenes	0.729	0.807		10.7	20
o-Xylene	0.706	0.790		11.9	20
Styrene	1.236	1.382		11.81	20
Bromoform	0.293	0.332	0.1	13.31	20
Isopropylbenzene	3.801	4.072		7.13	20
1,1,2,2-Tetrachloroethane	1.150	1.253	0.3	8.96	20
1,3-Dichlorobenzene	1.739	1.887		8.51	20
1,4-Dichlorobenzene	1.785	1.893		6.05	20
1,2-Dichlorobenzene	1.657	1.825		10.14	20
1,2-Dibromo-3-Chloropropane	0.233	0.241		3.43	20
1,2,4-Trichlorobenzene	1.077	1.150		6.78	20
1,2,3-Trichlorobenzene	1.002	1.103		10.08	20
1,2-Dichloroethane-d4	0.796	0.816		2.51	20
Dibromofluoromethane	0.335	0.352		5.07	20
Toluene-d8	1.160	1.200		3.45	20
4-Bromofluorobenzene	0.440	0.459		4.32	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3141</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/23/2025 19:00</u>
Lab File ID:	<u>VX047744.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.632		22.01	50
Chloromethane	0.680	0.784	0.1	15.29	50
Vinyl Chloride	0.666	0.776		16.52	50
Bromomethane	0.422	0.496		17.54	50
Chloroethane	0.445	0.505		13.48	50
Trichlorofluoromethane	0.989	1.097		10.92	50
1,1,2-Trichlorotrifluoroethane	0.578	0.639		10.55	50
1,1-Dichloroethene	0.594	0.657		10.61	50
Acetone	0.346	0.295		-14.74	50
Carbon Disulfide	1.626	1.828		12.42	50
Methyl tert-butyl Ether	2.169	2.338		7.79	50
Methyl Acetate	0.770	0.876		13.77	50
Methylene Chloride	0.732	0.797		8.88	50
trans-1,2-Dichloroethene	0.640	0.715		11.72	50
1,1-Dichloroethane	1.263	1.376	0.1	8.95	50
Cyclohexane	1.052	1.100		4.56	50
2-Butanone	0.432	0.442		2.32	50
Carbon Tetrachloride	0.532	0.522		-1.88	50
cis-1,2-Dichloroethene	0.783	0.862		10.09	50
Bromochloromethane	0.551	0.579		5.08	50
Chloroform	1.276	1.391		9.01	50
1,1,1-Trichloroethane	1.082	1.150		6.28	50
Methylcyclohexane	0.556	0.569		2.34	50
Benzene	1.488	1.586		6.59	50
1,2-Dichloroethane	0.566	0.582		2.83	50
Trichloroethene	0.360	0.386		7.22	50
1,2-Dichloropropane	0.381	0.410		7.61	50
Bromodichloromethane	0.572	0.607		6.12	50
4-Methyl-2-Pentanone	0.477	0.521		9.22	50
Toluene	0.917	0.974		6.22	50
t-1,3-Dichloropropene	0.584	0.590		1.03	50
cis-1,3-Dichloropropene	0.624	0.642		2.88	50
1,1,2-Trichloroethane	0.354	0.387		9.32	50
2-Hexanone	0.343	0.358		4.37	50
Dibromochloromethane	0.407	0.441		8.35	50
1,2-Dibromoethane	0.360	0.396		10	50
Tetrachloroethene	0.326	0.337		3.37	50
Chlorobenzene	1.136	1.185	0.3	4.31	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3141</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/23/2025</u> <u>19:00</u>
Lab File ID:	<u>VX047744.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	2.020		3.06	50
m/p-Xylenes	0.729	0.755		3.57	50
o-Xylene	0.706	0.741		4.96	50
Styrene	1.236	1.305		5.58	50
Bromoform	0.293	0.305	0.1	4.1	50
Isopropylbenzene	3.801	3.780		-0.55	50
1,1,2,2-Tetrachloroethane	1.150	1.194	0.3	3.83	50
1,3-Dichlorobenzene	1.739	1.714		-1.44	50
1,4-Dichlorobenzene	1.785	1.729		-3.14	50
1,2-Dichlorobenzene	1.657	1.655		-0.12	50
1,2-Dibromo-3-Chloropropane	0.233	0.228		-2.15	50
1,2,4-Trichlorobenzene	1.077	1.029		-4.46	50
1,2,3-Trichlorobenzene	1.002	0.989		-1.3	50
1,2-Dichloroethane-d4	0.796	0.781		-1.88	50
Dibromofluoromethane	0.335	0.336		0.3	50
Toluene-d8	1.160	1.139		-1.81	50
4-Bromofluorobenzene	0.440	0.434		-1.36	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3141</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/24/2025 02:08</u>
Lab File ID:	<u>VX047746.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.672		29.73	20
Chloromethane	0.680	0.783	0.1	15.15	20
Vinyl Chloride	0.666	0.789		18.47	20
Bromomethane	0.422	0.468		10.9	20
Chloroethane	0.445	0.492		10.56	20
Trichlorofluoromethane	0.989	1.101		11.32	20
1,1,2-Trichlorotrifluoroethane	0.578	0.655		13.32	20
1,1-Dichloroethene	0.594	0.663		11.62	20
Acetone	0.346	0.298		-13.87	20
Carbon Disulfide	1.626	1.796		10.45	20
Methyl tert-butyl Ether	2.169	2.253		3.87	20
Methyl Acetate	0.770	0.868		12.73	20
Methylene Chloride	0.732	0.771		5.33	20
trans-1,2-Dichloroethene	0.640	0.695		8.59	20
1,1-Dichloroethane	1.263	1.349	0.1	6.81	20
Cyclohexane	1.052	1.077		2.38	20
2-Butanone	0.432	0.438		1.39	20
Carbon Tetrachloride	0.532	0.554		4.14	20
cis-1,2-Dichloroethene	0.783	0.815		4.09	20
Bromochloromethane	0.551	0.629		14.16	20
Chloroform	1.276	1.328		4.07	20
1,1,1-Trichloroethane	1.082	1.123		3.79	20
Methylcyclohexane	0.556	0.576		3.6	20
Benzene	1.488	1.584		6.45	20
1,2-Dichloroethane	0.566	0.569		0.53	20
Trichloroethene	0.360	0.387		7.5	20
1,2-Dichloropropane	0.381	0.401		5.25	20
Bromodichloromethane	0.572	0.594		3.85	20
4-Methyl-2-Pentanone	0.477	0.516		8.18	20
Toluene	0.917	0.950		3.6	20
t-1,3-Dichloropropene	0.584	0.561		-3.94	20
cis-1,3-Dichloropropene	0.624	0.614		-1.6	20
1,1,2-Trichloroethane	0.354	0.369		4.24	20
2-Hexanone	0.343	0.357		4.08	20
Dibromochloromethane	0.407	0.431		5.9	20
1,2-Dibromoethane	0.360	0.382		6.11	20
Tetrachloroethene	0.326	0.346		6.14	20
Chlorobenzene	1.136	1.179	0.3	3.79	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3141</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/24/2025</u> <u>02:08</u>
Lab File ID:	<u>VX047746.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	2.043		4.24	20
m/p-Xylenes	0.729	0.773		6.04	20
o-Xylene	0.706	0.741		4.96	20
Styrene	1.236	1.308		5.82	20
Bromoform	0.293	0.317	0.1	8.19	20
Isopropylbenzene	3.801	3.949		3.89	20
1,1,2,2-Tetrachloroethane	1.150	1.228	0.3	6.78	20
1,3-Dichlorobenzene	1.739	1.760		1.21	20
1,4-Dichlorobenzene	1.785	1.788		0.17	20
1,2-Dichlorobenzene	1.657	1.706		2.96	20
1,2-Dibromo-3-Chloropropane	0.233	0.234		0.43	20
1,2,4-Trichlorobenzene	1.077	1.065		-1.11	20
1,2,3-Trichlorobenzene	1.002	1.020		1.8	20
1,2-Dichloroethane-d4	0.796	0.863		8.42	20
Dibromofluoromethane	0.335	0.398		18.81	20
Toluene-d8	1.160	1.325		14.22	20
4-Bromofluorobenzene	0.440	0.493		12.05	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3141</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/24/2025 10:50</u>
Lab File ID:	<u>VX047767.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.593		14.48	50
Chloromethane	0.680	0.735	0.1	8.09	50
Vinyl Chloride	0.666	0.726		9.01	50
Bromomethane	0.422	0.463		9.72	50
Chloroethane	0.445	0.466		4.72	50
Trichlorofluoromethane	0.989	1.013		2.43	50
1,1,2-Trichlorotrifluoroethane	0.578	0.586		1.38	50
1,1-Dichloroethene	0.594	0.631		6.23	50
Acetone	0.346	0.323		-6.65	50
Carbon Disulfide	1.626	1.720		5.78	50
Methyl tert-butyl Ether	2.169	2.352		8.44	50
Methyl Acetate	0.770	0.940		22.08	50
Methylene Chloride	0.732	0.779		6.42	50
trans-1,2-Dichloroethene	0.640	0.685		7.03	50
1,1-Dichloroethane	1.263	1.341	0.1	6.18	50
Cyclohexane	1.052	1.066		1.33	50
2-Butanone	0.432	0.500		15.74	50
Carbon Tetrachloride	0.532	0.515		-3.19	50
cis-1,2-Dichloroethene	0.783	0.835		6.64	50
Bromochloromethane	0.551	0.591		7.26	50
Chloroform	1.276	1.376		7.84	50
1,1,1-Trichloroethane	1.082	1.134		4.81	50
Methylcyclohexane	0.556	0.525		-5.58	50
Benzene	1.488	1.534		3.09	50
1,2-Dichloroethane	0.566	0.562		-0.71	50
Trichloroethene	0.360	0.371		3.06	50
1,2-Dichloropropane	0.381	0.401		5.25	50
Bromodichloromethane	0.572	0.591		3.32	50
4-Methyl-2-Pentanone	0.477	0.565		18.45	50
Toluene	0.917	0.943		2.84	50
t-1,3-Dichloropropene	0.584	0.540		-7.53	50
cis-1,3-Dichloropropene	0.624	0.587		-5.93	50
1,1,2-Trichloroethane	0.354	0.379		7.06	50
2-Hexanone	0.343	0.396		15.45	50
Dibromochloromethane	0.407	0.437		7.37	50
1,2-Dibromoethane	0.360	0.406		12.78	50
Tetrachloroethene	0.326	0.340		4.29	50
Chlorobenzene	1.136	1.145	0.3	0.79	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3141</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/24/2025</u> <u>10:50</u>
Lab File ID:	<u>VX047767.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	1.949		-0.56	50
m/p-Xylenes	0.729	0.741		1.65	50
o-Xylene	0.706	0.724		2.55	50
Styrene	1.236	1.277		3.32	50
Bromoform	0.293	0.316	0.1	7.85	50
Isopropylbenzene	3.801	3.694		-2.82	50
1,1,2,2-Tetrachloroethane	1.150	1.264	0.3	9.91	50
1,3-Dichlorobenzene	1.739	1.732		-0.4	50
1,4-Dichlorobenzene	1.785	1.729		-3.14	50
1,2-Dichlorobenzene	1.657	1.696		2.35	50
1,2-Dibromo-3-Chloropropane	0.233	0.254		9.01	50
1,2,4-Trichlorobenzene	1.077	1.059		-1.67	50
1,2,3-Trichlorobenzene	1.002	1.049		4.69	50
1,2-Dichloroethane-d4	0.796	0.866		8.79	50
Dibromofluoromethane	0.335	0.362		8.06	50
Toluene-d8	1.160	1.237		6.64	50
4-Bromofluorobenzene	0.440	0.475		7.95	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.